



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98966

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.3	4.01	11/01/2018	09:45
2	CAL2	1	Water			21.2	7.00	11/01/2018	09:47
3	CAL3	1	Water			21.5	10.04	11/01/2018	09:49
4	ICV	1	Water			21.4	7.01	11/01/2018	09:51
5	CCV1	1	Water			21.3	2.00	11/01/2018	09:53
6	J5755-04	1	Solid	20	20	21.7	8.03	11/01/2018	09:55
7	J5755-04DUP	1	Solid	20.02	20	21.8	8.04	11/01/2018	09:57
8	J5755-08	1	Solid	20.02	20	21.6	7.75	11/01/2018	09:59
9	J5755-12	1	Solid	20.04	20	21.6	7.89	11/01/2018	10:01
10	J5759-04	1	Solid	20.03	20	21.4	9.66	11/01/2018	10:03
11	J5759-08	1	Solid	20.02	20	21.5	9.52	11/01/2018	10:05
12	J5759-12	1	Solid	20.02	20	21.4	8.99	11/01/2018	10:07
13	J5759-16	1	Solid	20.03	20	21.3	10.77	11/01/2018	10:09
14	J5759-20	1	Solid	20.04	20	21.2	9.43	11/01/2018	10:11
15	J5759-24	1	Solid	20.02	20	21.3	8.61	11/01/2018	10:13
16	CCV2	1	Water			21.3	12.00	11/01/2018	10:15

WORKLIST(Hardcopy Internal Chain)

LB 98966

WorkList Name : CORROSIONITY110118

WorkList ID : 118714

Date : 11/1/2018 9:31:39 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/31/2018	Solid	J5755-04	Corrosivity	Cool 4 deg C	PSEG01	I21	TP-1-S	10/31/2018	9045D
10/31/2018	Solid	J5755-08	Corrosivity	Cool 4 deg C	PSEG01	I21	TP-2A-S	10/31/2018	9045D
10/31/2018	Solid	J5755-12	Corrosivity	Cool 4 deg C	PSEG01	I21	TP-2A-D	10/31/2018	9045D
10/31/2018	Solid	J5759-04	Corrosivity	Cool 4 deg C	PSEG01	I13	TP-1	10/31/2018	9045D
10/31/2018	Solid	J5759-08	Corrosivity	Cool 4 deg C	PSEG01	I13	TP-2	10/31/2018	9045D
10/31/2018	Solid	J5759-12	Corrosivity	Cool 4 deg C	PSEG01	I13	TP-3	10/31/2018	9045D
10/31/2018	Solid	J5759-16	Corrosivity	Cool 4 deg C	PSEG01	I13	TP-4	10/31/2018	9045D
10/31/2018	Solid	J5759-20	Corrosivity	Cool 4 deg C	PSEG01	I13	TP-5	10/31/2018	9045D
10/31/2018	Solid	J5759-24	Corrosivity	Cool 4 deg C	PSEG01	I13	TP-6	10/31/2018	9045D

Date/Time 11/01/18 9:35
 Received by: SP
 Relinquished by: CP

Date/Time 11/01/18 10:25
 Received by: CP
 Relinquished by: SP