

LB98972


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34001



A= 0.0000e+000

B= 2.9258e-004

C= -1.7701e-002

Rho= 0.9999340

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	-4	0.000	-4				
0.05	0.050	0.054	0.004	245	0.0 %	245				
0.2	0.200	0.194	-0.006	724	0.0 %	724				
2.5	2.500	2.476	-0.024	8523	0.0 %	8523				
5.0	5.000	5.045	0.045	17305	0.0 %	17305				
7.5	7.500	7.569	0.069	25932	0.0 %	25932				
10.0	10.000	9.930	-0.070	34001	0.0 %	34001				

LB98972 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 01 Nov 2018 12:46:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-4	-	0.0000	245.1	PPB	01 Nov 2018 12:47:54	S	Std
0.05 - 1	50.05	245	-	0.0500	245.1	PPB	01 Nov 2018 12:50:12	S	Std
0.2 - 1	50.2	724	-	0.2000	245.1	PPB	01 Nov 2018 12:52:30	S	Std
2.5 - 1	52.5	8523	-	2.5000	245.1	PPB	01 Nov 2018 12:54:48	S	Std
5.0 - 1	55	17305	-	5.0000	245.1	PPB	01 Nov 2018 12:57:05	S	Std
7.5 - 1	57.5	25932	-	7.5000	245.1	PPB	01 Nov 2018 12:59:22	S	Std
10.0 - 1	51)	34001	-	10.0000	245.1	PPB	01 Nov 2018 13:04:18	S	Std
ICV51 - 1	ICV51	14322	4.1726	-245.1		PPB	01 Nov 2018 13:07:06	U	SMPL
ICB51 - 1	ICB51	-23	-0.0244	-245.1		PPB	01 Nov 2018 13:09:21	U	SMPL
CCV49 - 1	CCV49	17747	5.1746	-245.1		PPB	01 Nov 2018 13:11:37	U	SMPL
CCB49 - 1	CCB49	-45	-0.0309	-245.1		PPB	01 Nov 2018 13:13:53	U	SMPL
CRA - 1	CRA	211	0.0440	-245.1		PPB	01 Nov 2018 13:16:09	U	SMPL
HighStd - 1	HighStd	33818	9.8766	-245.1		PPB	01 Nov 2018 13:18:25	U	SMPL
ChkStd - 1	ChkStd	24305	7.0933	-245.1		PPB	01 Nov 2018 13:23:42	U	SMPL
PB114364BL - 1	PBW	-101	-0.0473	-245.1		PPB	01 Nov 2018 13:25:58	U	SMPL
PB114364BS - 1	LCSW	13284	3.8689	-245.1		PPB	01 Nov 2018 13:28:16	U	SMPL
CCV50 - 1	CCV50	17691	5.1583	-245.1		PPB	01 Nov 2018 13:41:58	U	SMPL
CCB50 - 1	CCB50	-44	-0.0306	-245.1		PPB	01 Nov 2018 13:44:13	U	SMPL
J5722-01 - 1	102618-SIDI-F-D-B	-146	-0.0604	-245.1		PPB	01 Nov 2018 13:57:51	U	SMPL
J5722-01DUP - 1	102618-SIDI-F-D-BDUP	-50	-0.0323	-245.1		PPB	01 Nov 2018 14:00:08	U	SMPL
J5722-01MS - 1	102618-SIDI-F-D-BMS	14884	4.3370	-245.1		PPB	01 Nov 2018 14:02:26	U	SMPL
J5722-01MSD - 1	102618-SIDI-F-D-BMSD	15180	4.4236	-245.1		PPB	01 Nov 2018 14:04:44	U	SMPL
J5722-02 - 1	102618-SIDI-F-D-M	-156	-0.0633	-245.1		PPB	01 Nov 2018 14:07:00	U	SMPL
CCV51 - 1	CCV51	17514	5.1065	-245.1		PPB	01 Nov 2018 14:09:17	U	SMPL
CCB51 - 1	CCB51	-39	-0.0291	-245.1		PPB	01 Nov 2018 14:11:33	U	SMPL
J5722-03 - 1	102618-SIDI-F-D-S	-115	-0.0513	-245.1		PPB	01 Nov 2018 14:13:50	U	SMPL
J5722-04 - 1	102618-SIDI-F-U-B	-45	-0.0309	-245.1		PPB	01 Nov 2018 14:16:06	U	SMPL
J5722-05 - 1	102618-SIDI-F-U-M	-74	-0.0394	-245.1		PPB	01 Nov 2018 14:18:22	U	SMPL
J5722-06 - 1	102618-SIDI-F-U-S	-68	-0.0376	-245.1		PPB	01 Nov 2018 14:20:39	U	SMPL
J5722-07 - 1	102618-TIPC-F-D-B	-98	-0.0464	-245.1		PPB	01 Nov 2018 14:22:56	U	SMPL
J5722-08 - 1	102618-TIPC-F-D-S	-81	-0.0414	-245.1		PPB	01 Nov 2018 14:25:13	U	SMPL
J5722-09 - 1	102618-TIPC-F-U-B	-43	-0.0303	-245.1		PPB	01 Nov 2018 14:27:30	U	SMPL
J5722-10 - 1	102618-TIPC-F-U-S	-105	-0.0484	-245.1		PPB	01 Nov 2018 14:29:48	U	SMPL
PB114365BL - 1	PBW	10	-0.0148	-245.1		PPB	01 Nov 2018 14:32:05	U	SMPL
PB114365BS - 1	LCSW	13357	3.8902	-245.1		PPB	01 Nov 2018 14:34:22	U	SMPL
CCV52 - 1	CCV52	16951	4.9417	-245.1		PPB	01 Nov 2018 14:36:39	U	SMPL
CCB52 - 1	CCB52	34	-0.0078	-245.1		PPB	01 Nov 2018 14:38:55	U	SMPL
J5730-01 - 1	102918-TIPC-E-D-M	-89	-0.0437	-245.1		PPB	01 Nov 2018 14:41:13	U	SMPL
J5730-01DUP - 1	102918-TIPC-E-D-MDUP	-53	-0.0332	-245.1		PPB	01 Nov 2018 14:43:29	U	SMPL
J5730-01MS - 1	102918-TIPC-E-D-MMS	15006	4.3727	-245.1		PPB	01 Nov 2018 14:45:45	U	SMPL
J5730-01MSD - 1	102918-TIPC-E-D-MMSD	14762	4.3013	-245.1		PPB	01 Nov 2018 14:48:02	U	SMPL
J5730-02 - 1	102918-TIPC-E-U-M	-162	-0.0651	-245.1		PPB	01 Nov 2018 14:50:18	U	SMPL
J5730-03 - 1	102918-SIDI-E-D-B	-96	-0.0458	-245.1		PPB	01 Nov 2018 14:52:36	U	SMPL
J5730-04 - 1	102918-SIDI-E-D-M	-95	-0.0455	-245.1		PPB	01 Nov 2018 14:54:53	U	SMPL
J5730-05 - 1	102918-SIDI-E-D-S	-81	-0.0414	-245.1		PPB	01 Nov 2018 14:57:10	U	SMPL
J5730-06 - 1	102918-SIDI-E-U-B	-75	-0.0396	-245.1		PPB	01 Nov 2018 14:59:28	U	SMPL
J5730-07 - 1	102918-SIDI-E-U-M	-52	-0.0329	-245.1		PPB	01 Nov 2018 15:01:45	U	SMPL
CCV53 - 1	CCV53	16987	4.9523	-245.1		PPB	01 Nov 2018 15:04:01	U	SMPL
CCB53 - 1	CCB53	35	-0.0075	-245.1		PPB	01 Nov 2018 15:06:17	U	SMPL
J5730-08 - 1	102918-SIDI-E-U-S	-48	-0.0317	-245.1		PPB	01 Nov 2018 15:08:34	U	SMPL
PB114366BL - 1	PBW	54	-0.0019	-245.1		PPB	01 Nov 2018 15:10:51	U	SMPL
PB114366BS - 1	LCSW	13461	3.9207	-245.1		PPB	01 Nov 2018 15:13:07	U	SMPL
J5731-01 - 1	103018-TIPC-E-D-M	-140	-0.0587	-245.1		PPB	01 Nov 2018 15:15:24	U	SMPL
J5731-01DUP - 1	103018-TIPC-E-D-MDUP	-85	-0.0426	-245.1		PPB	01 Nov 2018 15:17:40	U	SMPL
J5731-01MS - 1	103018-TIPC-E-D-MMS	14810	4.3153	-245.1		PPB	01 Nov 2018 15:19:57	U	SMPL
J5731-01MSD - 1	103018-TIPC-E-D-MMSD	15308	4.4610	-245.1		PPB	01 Nov 2018 15:22:15	U	SMPL
J5731-02 - 1	103018-TIPC-E-U-M	-63	-0.0361	-245.1		PPB	01 Nov 2018 15:24:32	U	SMPL
J5731-03 - 1	103018-SIDI-E-D-B	-36	-0.0282	-245.1		PPB	01 Nov 2018 15:26:50	U	SMPL
J5731-04 - 1	103018-SIDI-E-D-M	-43	-0.0303	-245.1		PPB	01 Nov 2018 15:29:06	U	SMPL
CCV54 - 1	CCV54	16960	4.9444	-245.1		PPB	01 Nov 2018 15:31:23	U	SMPL
CCB54 - 1	CCB54	58	-0.0007	-245.1		PPB	01 Nov 2018 15:33:39	U	SMPL
J5731-05 - 1	103018-SIDI-E-D-S	0	-0.0177	-245.1		PPB	01 Nov 2018 15:35:57	U	SMPL
J5731-06 - 1	103018-SIDI-E-U-B	-73	-0.0391	-245.1		PPB	01 Nov 2018 15:38:14	U	SMPL
J5731-07 - 1	103018-SIDI-E-U-M	-111	-0.0502	-245.1		PPB	01 Nov 2018 15:40:30	U	SMPL
J5731-08 - 1	103018-SIDI-E-U-S	-50	-0.0323	-245.1		PPB	01 Nov 2018 15:42:47	U	SMPL
J5731-09 - 1	103018-DBRI-E-D-B	-80	-0.0411	-245.1		PPB	01 Nov 2018 15:45:04	U	SMPL
J5731-10 - 1	103018-DBRI-E-D-M	-29	-0.0262	-245.1		PPB	01 Nov 2018 15:47:20	U	SMPL
J5731-11 - 1	103018-DBRI-E-D-S	-46	-0.0312	-245.1		PPB	01 Nov 2018 15:49:38	U	SMPL
J5731-12 - 1	103018-DBRI-E-U-B	-44	-0.0306	-245.1		PPB	01 Nov 2018 15:51:55	U	SMPL
J5731-13 - 1	103018-DBRI-E-U-M	-57	-0.0344	-245.1		PPB	01 Nov 2018 15:54:13	U	SMPL
J5731-14 - 1	103018-DBRI-E-U-S	-85	-0.0426	-245.1		PPB	01 Nov 2018 15:56:30	U	SMPL
CCV55 - 1	CCV55	17607	5.1337	-245.1		PPB	01 Nov 2018 15:58:47	U	SMPL
CCB55 - 1	CCB55	-43	-0.0303	-245.1		PPB	01 Nov 2018 16:01:04	U	SMPL
J5722-01LX5 - 1		4	-0.0165	-245.1		PPB	01 Nov 2018 16:03:21	U	SMPL
J5722-01A - 1		14822	4.3189	-245.1		PPB	01 Nov 2018 16:05:38	U	SMPL
J5730-01LX5 - 1		-37	-0.0285	-245.1		PPB	01 Nov 2018 16:07:55	U	SMPL
J5730-01A - 1		15084	4.3955	-245.1		PPB	01 Nov 2018 16:10:12	U	SMPL
J5731-01LX5 - 1		-55	-0.0338	-245.1		PPB	01 Nov 2018 16:12:29	U	SMPL
J5731-01A - 1		14779	4.3063	-245.1		PPB	01 Nov 2018 16:14:46	U	SMPL
CCV56 - 1	CCV56	17128	4.9935	-245.1		PPB	01 Nov 2018 16:17:04	U	SMPL
CCB56 - 1	CCB56	-9	-0.0203	-245.1		PPB	01 Nov 2018 16:19:21	U	SMPL

LB98972 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 01 Nov 2018 12:46:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
J5706-01 - 1	102518-SIDI-F-D-B	930	0.2544	-245.1		PPB	01 Nov 2018 16:22:33	U	SMPL
J5706-02 - 1	102518-SIDI-F-D-M	939	0.2570	-245.1		PPB	01 Nov 2018 16:24:50	U	SMPL
J5706-03 - 1	102518-SIDI-F-D-S	823	0.2231	-245.1		PPB	01 Nov 2018 16:27:07	U	SMPL
J5706-04 - 1	102518-SIDI-F-U-B	1006	0.2766	-245.1		PPB	01 Nov 2018 16:29:26	U	SMPL
J5706-05 - 1	102518-SIDI-F-U-M	890	0.2427	-245.1		PPB	01 Nov 2018 16:31:42	U	SMPL
J5706-06 - 1	102518-SIDI-F-U-S	884	0.2409	-245.1		PPB	01 Nov 2018 16:36:33	U	SMPL
J5706-07 - 1	102518-TIPC-F-D-B	930	0.2544	-245.1		PPB	01 Nov 2018 16:38:48	U	SMPL
J5706-08 - 1	102518-TIPC-F-D-S	970	0.2661	-245.1		PPB	01 Nov 2018 16:41:04	U	SMPL
J5706-09 - 1	102518-TIPC-F-U-B	1034	0.2848	-245.1		PPB	01 Nov 2018 16:43:21	U	SMPL
J5706-10 - 1	102518-TIPC-F-U-S	937	0.2564	-245.1		PPB	01 Nov 2018 16:45:38	U	SMPL
CCV57 - 1	CCV57	17309	5.0465	-245.1		PPB	01 Nov 2018 16:48:17	U	SMPL
CCB57 - 1	CCB57	7	-0.0157	-245.1		PPB	01 Nov 2018 16:50:34	U	SMPL