


7471B

LB 98985
INSTRUMENT ID: CV2

Linear

 μ Abs.:

50262



A= 0.0000e+000

B= 1.9909e-004

C= -5.0218e-002

Rho= 0.9998993

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.055	-0.055	-23	0.000	-23				
0.2	0.200	0.228	0.028	1398	0.0 %	1398				
2.5	2.500	2.473	-0.027	12672	0.0 %	12672				
5.0	5.000	5.100	0.100	25869	0.0 %	25869				
7.5	7.500	7.498	-0.002	37912	0.0 %	37912				
10.0	10.000	9.956	-0.044	50262	0.0 %	50262				

LB98985 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Nov 2018 17:21:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-23	-	0.00007471B		PPB	01 Nov 2018 17:28:33	S	Std
0.2 - 1	50.2	1398	-	0.20007471B		PPB	01 Nov 2018 17:30:50	S	Std
2.5 - 1	52.5	12672	-	2.50007471B		PPB	01 Nov 2018 17:33:08	S	Std
5.0 - 1	55	25869	-	5.00007471B		PPB	01 Nov 2018 17:35:26	S	Std
7.5 - 1	57.5	37912	-	7.50007471B		PPB	01 Nov 2018 17:37:43	S	Std
10.0 - 1	510	50262	-	10.00007471B		PPB	01 Nov 2018 17:42:36	S	Std
ICV52 - 1	ICV52	20301	3.9915	-7471B		PPB	01 Nov 2018 17:45:13	U	SMPL
ICB52 - 1	ICB52	-13	-0.0528	-7471B		PPB	01 Nov 2018 17:47:29	U	SMPL
CCV58 - 1	CCV58	25757	5.0777	-7471B		PPB	01 Nov 2018 17:49:46	U	SMPL
CCB58 - 1	CCB58	-66	-0.0634	-7471B		PPB	01 Nov 2018 17:52:02	U	SMPL
CRA - 1	CRA	1456	0.2397	-7471B		PPB	01 Nov 2018 17:54:18	U	SMPL
HighStd - 1	HighStd	50957	10.0948	-7471B		PPB	01 Nov 2018 17:56:34	U	SMPL
ChkStd - 1	ChkStd	35751	7.0674	-7471B		PPB	01 Nov 2018 17:58:50	U	SMPL
PB114373BL - 1	PBS	342	0.0179	-7471B		PPB	01 Nov 2018 18:01:09	U	SMPL
PB114373BS - 1	LCSS	18015	3.5364	-7471B		PPB	01 Nov 2018 18:03:27	U	SMPL
J5192-05 - 1	AU-05-103118-A	3419	0.6305	-7471B		PPB	01 Nov 2018 18:05:45	U	SMPL
J5192-05DUP - 1	AU-05-103118-ADUP	3410	0.6287	-7471B		PPB	01 Nov 2018 18:08:02	U	SMPL
J5192-05MS - 1	AU-05-103118-AMS	21403	4.2109	-7471B		PPB	01 Nov 2018 18:10:20	U	SMPL
J5192-05MSD - 1	AU-05-103118-AMSD	21145	4.1595	-7471B		PPB	01 Nov 2018 18:12:38	U	SMPL
J5192-07 - 1	AU-05-103118-B	3210	0.5889	-7471B		PPB	01 Nov 2018 18:14:55	U	SMPL
CCV59 - 1	CCV59	25845	5.0952	-7471B		PPB	01 Nov 2018 18:17:11	U	SMPL
CCB59 - 1	CCB59	-105	-0.0711	-7471B		PPB	01 Nov 2018 18:19:27	U	SMPL
J5200-17 - 1	OR-03-103118-A	1389	0.2263	-7471B		PPB	01 Nov 2018 18:21:45	U	SMPL
J5200-19 - 1	OR-03-103118-B	2387	0.4250	-7471B		PPB	01 Nov 2018 18:24:00	U	SMPL
J5734-02 - 1	B-21(0-10)	4756	0.8966	-7471B		PPB	01 Nov 2018 18:26:16	U	SMPL
J5734-05 - 1	B-22(0-10)	9035	1.7486	-7471B		PPB	01 Nov 2018 18:28:33	U	SMPL
J5755-03 - 1	TP-1-S	35559	7.0292	-7471B		PPB	01 Nov 2018 18:30:49	U	SMPL
J5755-07 - 1	TP-2A-S	703	0.0897	-7471B		PPB	01 Nov 2018 18:33:06	U	SMPL
J5755-11 - 1	TP-2A-D	1090	0.1668	-7471B		PPB	01 Nov 2018 18:35:25	U	SMPL
J5759-01 - 1	TP-1	20218	3.9750	-7471B		PPB	01 Nov 2018 18:37:43	U	SMPL
J5759-05 - 1	TP-2	5163	0.9777	-7471B		PPB	01 Nov 2018 18:40:00	U	SMPL
J5759-09 - 1	TP-3	58296	11.5559	-7471B		PPB	01 Nov 2018 18:42:17	U	SMPL
CCV60 - 1	CCV60	25850	5.0962	-7471B		PPB	01 Nov 2018 18:44:33	U	SMPL
CCB60 - 1	CCB60	-112	-0.0725	-7471B		PPB	01 Nov 2018 18:46:52	U	SMPL
J5759-13 - 1	TP-4	6253	1.1947	-7471B		PPB	01 Nov 2018 18:49:10	U	SMPL
J5759-17 - 1	TP-5	33713	6.6617	-7471B		PPB	01 Nov 2018 18:51:26	U	SMPL
J5759-21 - 1	TP-6	31300	6.1813	-7471B		PPB	01 Nov 2018 18:53:42	U	SMPL
PB114374BL - 1	PBS	368	0.0230	-7471B		PPB	01 Nov 2018 18:56:00	U	SMPL
PB114374BS - 1	LCSS	19155	3.7633	-7471B		PPB	01 Nov 2018 18:58:18	U	SMPL
J5741-03 - 1	TP-1	1465	0.2414	-7471B		PPB	01 Nov 2018 19:00:35	U	SMPL
J5741-07 - 1	TP-2	1033	0.1554	-7471B		PPB	01 Nov 2018 19:02:53	U	SMPL
J5741-11 - 1	TP-3	1610	0.2703	-7471B		PPB	01 Nov 2018 19:05:11	U	SMPL
J5741-15 - 1	TP-4	8056	1.5536	-7471B		PPB	01 Nov 2018 19:07:30	U	SMPL
J5741-19 - 1	TP-5	17270	3.3880	-7471B		PPB	01 Nov 2018 19:09:47	U	SMPL
CCV61 - 1	CCV61	25996	5.1253	-7471B		PPB	01 Nov 2018 19:12:03	U	SMPL
CCB61 - 1	CCB61	38	-0.0427	-7471B		PPB	01 Nov 2018 19:14:20	U	SMPL
J5741-23 - 1	TP-6	1410	0.2305	-7471B		PPB	01 Nov 2018 19:16:38	U	SMPL
J5741-27 - 1	TP-7	1143	0.1773	-7471B		PPB	01 Nov 2018 19:18:54	U	SMPL
J5741-31 - 1	TP-8	3637	0.6739	-7471B		PPB	01 Nov 2018 19:21:11	U	SMPL
J5741-35 - 1	TP-9	2961	0.5393	-7471B		PPB	01 Nov 2018 19:23:27	U	SMPL
J5741-39 - 1	TP-10	2915	0.5301	-7471B		PPB	01 Nov 2018 19:25:44	U	SMPL
J5741-43 - 1	TP-11	3101	0.5672	-7471B		PPB	01 Nov 2018 19:28:01	U	SMPL
J5741-47 - 1	TP-12	3177	0.5823	-7471B		PPB	01 Nov 2018 19:30:18	U	SMPL
J5741-51 - 1	TP-13	2584	0.4642	-7471B		PPB	01 Nov 2018 19:32:35	U	SMPL
J5741-55 - 1	TP-14	2135	0.3748	-7471B		PPB	01 Nov 2018 19:34:53	U	SMPL
CCV62 - 1	CCV62	27119	5.3489	-7471B		PPB	01 Nov 2018 19:37:09	U	SMPL
CCB62 - 1	CCB62	40	-0.0423	-7471B		PPB	01 Nov 2018 19:39:26	U	SMPL
J5741-59 - 1	TP-15	2323	0.4123	-7471B		PPB	01 Nov 2018 19:41:43	U	SMPL
J5741-59DUP - 1	TP-15DUP	2174	0.3826	-7471B		PPB	01 Nov 2018 19:43:59	U	SMPL
J5741-59MS - 1	TP-15MS	20823	4.0954	-7471B		PPB	01 Nov 2018 19:46:16	U	SMPL
J5741-59MSD - 1	TP-15MSD	20603	4.0516	-7471B		PPB	01 Nov 2018 19:48:32	U	SMPL
J5192-05LX5 - 1		629	0.0750	-7471B		PPB	01 Nov 2018 19:50:49	U	SMPL
J5192-05A - 1		22917	4.5123	-7471B		PPB	01 Nov 2018 19:53:06	U	SMPL
J5741-59LX5 - 1		366	0.0226	-7471B		PPB	01 Nov 2018 19:55:24	U	SMPL
J5741-59A - 1		21764	4.2828	-7471B		PPB	01 Nov 2018 19:57:41	U	SMPL
J5759-09DLX2 - 1		26480	5.2217	-7471B		PPB	01 Nov 2018 19:59:58	U	SMPL
CCV63 - 1	CCV63	26700	5.2655	-7471B		PPB	01 Nov 2018 20:02:17	U	SMPL
CCB63 - 1	CCB63	-274	-0.1048	-7471B		PPB	01 Nov 2018 20:04:36	U	SMPL