



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98989

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.3	4.01	11/01/2018	17:35
2	CAL2	1	Water			21.4	7.00	11/01/2018	17:37
3	CAL3	1	Water			21.2	10.04	11/01/2018	17:39
4	ICV	1	Water			21.4	7.01	11/01/2018	17:41
5	CCV1	1	Water			21.6	2.00	11/01/2018	17:43
6	J5793-01	1	Solid	20.03	20	23.3	5.37	11/01/2018	17:45
7	J5793-01DUP	1	Solid	20.04	20	23.4	5.38	11/01/2018	17:47
8	J5793-02	1	Solid	20.02	20	22.9	5.67	11/01/2018	17:49
9	J5793-03	1	Solid	20.02	20	23.2	6.76	11/01/2018	17:51
10	CCV2	1	Water			21.8	12.00	11/01/2018	17:53

WORKLIST(Hardcopy Internal Chain)

LB 98989

WorkList Name : pH110118

WorkList ID : 118739

Date : 11/1/2018 5:06:49 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/01/2018	Solid	J5793-01	pH	Cool 4 deg C	CTIA01	I31	WJ-18-11-01-TSHG-18	11/01/2018	9045D
11/01/2018	Solid	J5793-02	pH	Cool 4 deg C	CTIA01	I31	WJ-18-11-01-TSHG-19	11/01/2018	9045D
11/01/2018	Solid	J5793-03	pH	Cool 4 deg C	CTIA01	I31	WJ-18-11-01-TSHG-20	11/01/2018	9045D

Date/Time 11/01/18 5:25
 Received by: SP
 Relinquished by: CP

Date/Time 11/01/18 5:55
 Received by: CP
 Relinquished by: SP