



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98991

Slope : 99.1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.3	4.01	11/01/2018	17:35
2	CAL2	1	Water			21.4	7.00	11/01/2018	17:37
3	CAL3	1	Water			21.2	10.04	11/01/2018	17:39
4	ICV	1	Water			21.4	7.01	11/01/2018	17:41
5	CCV1	1	Water			21.8	2.00	11/01/2018	18:00
6	J5778-06	1	Water			24.3	6.22	11/01/2018	18:02
7	J5778-06DUP	1	Water			24.4	6.23	11/01/2018	18:04
8	CCV2	1	Water			21.9	12.00	11/01/2018	18:20

WORKLIST(Hardcopy Internal Chain)

1308991

WorkList Name : pH111118-1

WorkList ID : 118742

Date : 11/1/2018 5:07:52 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/01/2018	Water	J5778-06	pH	Cool 4 deg C	PSEG01	I23	OILY-WATER	11/01/2018	9040C

Date/Time 11/01/18 5:20
 Received by: SP
 Relinquished by: CP

Date/Time 11/01/18 6:25
 Received by: CP
 Relinquished by: JSP