



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98994

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.5	4.01	11/02/2018	09:45
2	CAL2	1	Water			22.6	7.00	11/02/2018	09:47
3	CAL3	1	Water			22.8	10.04	11/02/2018	09:51
4	ICV	1	Water			22.4	7.01	11/02/2018	09:53
5	CCV1	1	Water			22.5	2.00	11/02/2018	09:55
6	J5735-23	1	Solid	20.02	20	24.1	8.30	11/02/2018	09:57
7	J5735-23DUP	1	Solid	20.03	20	24.2	8.31	11/02/2018	09:59
8	J5745-21	1	Solid	20.02	20	24.3	5.12	11/02/2018	10:01
9	J5745-22	1	Solid	20.04	20	24.4	5.38	11/02/2018	10:03
10	J5745-23	1	Solid	20.02	20	24.2	5.20	11/02/2018	10:05
11	J5745-25	1	Solid	20.02	20	24.5	5.18	11/02/2018	10:07
12	J5745-26	1	Solid	20.03	20	24.3	5.48	11/02/2018	10:09
13	J5750-10	1	Solid	20.02	20	24.2	6.02	11/02/2018	10:11
14	J5750-11	1	Solid	20.02	20	24.0	5.75	11/02/2018	10:13
15	J5750-12	1	Solid	20.04	20	24.3	5.84	11/02/2018	10:15
16	CCV2	1	Water			22.8	12.00	11/02/2018	10:17
17	J5778-05	1	Solid	20.02	20	24.6	9.16	11/02/2018	10:19
18	J5794-04	1	Solid	20.03	20	24.3	6.15	11/02/2018	10:21
19	J5794-04DUP	1	Solid	20.04	20	24.4	6.16	11/02/2018	10:23
20	J5794-08	1	Solid	20.02	20	24.4	8.87	11/02/2018	10:25

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	J5794-12	1	Solid	20.03	20	24.1	7.88	11/02/2018	10:27
22	CCV3	1	Water			22.7	2.00	11/02/2018	10:29

WORKLIST(Hardcopy Internal Chain)

UB 98994

WorkList Name : CORROSIONITY110218

WorkList ID : 118770

Date : 11/2/2018 9:28:39 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/29/2018	Solid	J5735-23	Corrosivity	Cool 4 deg C	REMI02	I51	12B	10/29/2018	9045D
10/29/2018	Solid	J5745-21	Corrosivity	Cool 4 deg C	REMI02	I52	12H	10/29/2018	9045D
10/29/2018	Solid	J5745-22	Corrosivity	Cool 4 deg C	REMI02	I52	6A	10/29/2018	9045D
10/29/2018	Solid	J5745-23	Corrosivity	Cool 4 deg C	REMI02	I52	6B	10/29/2018	9045D
10/29/2018	Solid	J5745-25	Corrosivity	Cool 4 deg C	REMI02	I52	8A	10/29/2018	9045D
10/29/2018	Solid	J5745-26	Corrosivity	Cool 4 deg C	REMI02	I52	8B	10/29/2018	9045D
10/29/2018	Solid	J5750-10	Corrosivity	Cool 4 deg C	REMI02	I51	7A	10/29/2018	9045D
10/29/2018	Solid	J5750-11	Corrosivity	Cool 4 deg C	REMI02	I51	7B	10/29/2018	9045D
10/29/2018	Solid	J5750-12	Corrosivity	Cool 4 deg C	REMI02	I51	7C	10/29/2018	9045D
11/01/2018	Solid	J5778-05	Corrosivity	Cool 4 deg C	PSEG01	I23	OILY-GRAVEL	11/01/2018	9045D
11/01/2018	Solid	J5794-04	Corrosivity	Cool 4 deg C	PSEG01	I21	TP-12B-S	11/01/2018	9045D
11/01/2018	Solid	J5794-08	Corrosivity	Cool 4 deg C	PSEG01	I21	TP-11-S	11/01/2018	9045D
11/01/2018	Solid	J5794-12	Corrosivity	Cool 4 deg C	PSEG01	I21	TP-10B-S	11/01/2018	9045D

Date/Time 11/02/18 9:30
 Received by: SP
 Relinquished by: CP

Date/Time 11/02/18 10:40
 Received by: CP
 Relinquished by: SP