

7470A


LB98995

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35396



A= 0.0000e+000

B= 2.7999e-004

C= 3.8999e-002

Rho= 0.9999224

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.021	0.021	-64	0.000	-64				
0.2	0.200	0.187	-0.013	528	0.0 %	528				
2.5	2.500	2.437	-0.063	8565	0.0 %	8565				
5.0	5.000	5.056	0.056	17917	0.0 %	17917				
7.5	7.500	7.550	0.050	26825	0.0 %	26825				
10.0	10.000	9.950	-0.050	35396	0.0 %	35396				

LB98995 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Nov 2018 11:50:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	-64	-	0.0000	7470A	PPB	02 Nov 2018 11:54:45	S	Std
0.2 - 1	SD	528	-	0.2000	7470A	PPB	02 Nov 2018 11:56:55	S	Std
2.5 - 1	SD	8565	-	2.5000	7470A	PPB	02 Nov 2018 11:59:03	S	Std
5.0 - 1	SD	17917	-	5.0000	7470A	PPB	02 Nov 2018 12:01:11	S	Std
7.5 - 1	SD	26825	-	7.5000	7470A	PPB	02 Nov 2018 12:03:20	S	Std
10.0 - 1	SD	35396	-	10.0000	7470A	PPB	02 Nov 2018 12:07:52	S	Std
ICV53 - 1	ICV53	14006	3.9606	-	7470A	PPB	02 Nov 2018 12:10:28	U	SMPL
ICB53 - 1	ICB53	-86	0.0149	-	7470A	PPB	02 Nov 2018 12:12:34	U	SMPL
CCV64 - 1	CCV64	17904	5.0520	-	7470A	PPB	02 Nov 2018 12:15:10	U	SMPL
CCB64 - 1	CCB64	-198	-0.0164	-	7470A	PPB	02 Nov 2018 12:17:17	U	SMPL
CRA - 1	CRA	515	0.1832	-	7470A	PPB	02 Nov 2018 12:19:24	U	SMPL
HighStd - 1	HighStd	35925	10.0977	-	7470A	PPB	02 Nov 2018 12:21:31	U	SMPL
ChkStd - 1	ChkStd	24378	6.8646	-	7470A	PPB	02 Nov 2018 12:26:09	U	SMPL
PB114407BL - 1	PBW	-152	-0.0036	-	7470A	PPB	02 Nov 2018 12:28:16	U	SMPL
PB114407BS - 1	LCSW	13476	3.8122	-	7470A	PPB	02 Nov 2018 12:30:24	U	SMPL
J5728-04 - 1	EG-WATER	-140	-0.0002	-	7470A	PPB	02 Nov 2018 12:32:32	U	SMPL
J5728-04DUP - 1	EG-WATERDUP	-94	0.0127	-	7470A	PPB	02 Nov 2018 12:34:40	U	SMPL
J5728-04MS - 1	EG-WATERMS	12938	3.6615	-	7470A	PPB	02 Nov 2018 12:36:49	U	SMPL
J5728-04MSD - 1	EG-WATERMSD	12331	3.4916	-	7470A	PPB	02 Nov 2018 12:38:58	U	SMPL
J5729-04 - 1	WS1	-183	-0.0122	-	7470A	PPB	02 Nov 2018 12:41:06	U	SMPL
CCV65 - 1	CCV65	18145	5.1194	-	7470A	PPB	02 Nov 2018 12:43:13	U	SMPL
CCB65 - 1	CCB65	-135	0.0012	-	7470A	PPB	02 Nov 2018 12:45:20	U	SMPL
J5755-14 - 1	TP-2A	-62	0.0216	-	7470A	PPB	02 Nov 2018 12:47:27	U	SMPL
J5728-04LX5 - 1		-74	0.0183	-	7470A	PPB	02 Nov 2018 12:51:40	U	SMPL
J5728-04A - 1		12924	3.6576	-	7470A	PPB	02 Nov 2018 12:53:47	U	SMPL
PB114407TB - 1	PB114362TB	25	0.0460	-	7470A	PPB	02 Nov 2018 12:56:14	U	SMPL
CCV66 - 1	CCV66	18230	5.1432	-	7470A	PPB	02 Nov 2018 12:58:39	U	SMPL
CCB66 - 1	CCB66	-93	0.0130	-	7470A	PPB	02 Nov 2018 13:00:47	U	SMPL