


7471B

LB98996
INSTRUMENT ID: CV1

Linear

 μ Abs.:

47671



A= 0.0000e+000

B= 2.0910e-004

C= 2.8652e-002

Rho= 0.9999903

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	-57	0.000	-57				
0.2	0.200	0.184	-0.016	741	0.0 %	741				
2.5	2.500	2.484	-0.016	11743	0.0 %	11743				
5.0	5.000	5.027	0.027	23903	0.0 %	23903				
7.5	7.500	7.492	-0.008	35695	0.0 %	35695				
10.0	10.000	9.997	-0.003	47671	0.0 %	47671				

LB98996 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Nov 2018 13:12:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-57	-	0.0000	7471B	PPB	02 Nov 2018 13:13:25	S	Std
0.2 - 1	50.2	741	-	0.2000	7471B	PPB	02 Nov 2018 13:16:18	S	Std
2.5 - 1	52.5	11743	-	2.5000	7471B	PPB	02 Nov 2018 13:18:36	S	Std
5.0 - 1	55	23903	-	5.0000	7471B	PPB	02 Nov 2018 13:20:54	S	Std
7.5 - 1	57.5	35695	-	7.5000	7471B	PPB	02 Nov 2018 13:23:13	S	Std
10.0 - 1	50	47671	-	10.0000	7471B	PPB	02 Nov 2018 13:28:11	S	Std
ICV54 - 1	ICV54	18546	3.9066	-	7471B	PPB	02 Nov 2018 13:30:58	U	SMPL
ICB54 - 1	ICB54	-146	-0.0019	-	7471B	PPB	02 Nov 2018 13:33:15	U	SMPL
CCV67 - 1	CCV67	23857	5.0171	-	7471B	PPB	02 Nov 2018 13:35:33	U	SMPL
CCB67 - 1	CCB67	-148	-0.0023	-	7471B	PPB	02 Nov 2018 13:37:49	U	SMPL
CRA - 1	CRA	819	0.1999	-	7471B	PPB	02 Nov 2018 13:40:06	U	SMPL
HighStd - 1	HighStd	47482	9.9570	-	7471B	PPB	02 Nov 2018 13:42:23	U	SMPL
ChkStd - 1	ChkStd	34533	7.2494	-	7471B	PPB	02 Nov 2018 13:44:39	U	SMPL
PB114430BL - 1	PBS	-275	-0.0288	-	7471B	PPB	02 Nov 2018 13:46:56	U	SMPL
PB114430BS - 1	LCSS	20675	4.3517	-	7471B	PPB	02 Nov 2018 13:54:43	U	SMPL
J5735-05 - 1	12B	2003	0.4475	-	7471B	PPB	02 Nov 2018 13:57:00	U	SMPL
J5735-05DUP - 1	12BDUP	1849	0.4153	-	7471B	PPB	02 Nov 2018 13:59:19	U	SMPL
J5735-05MS - 1	12BMS	19210	4.0454	-	7471B	PPB	02 Nov 2018 14:01:36	U	SMPL
J5735-05MSD - 1	12BMSD	20288	4.2708	-	7471B	PPB	02 Nov 2018 14:03:54	U	SMPL
J5745-03 - 1	12H	156	0.0613	-	7471B	PPB	02 Nov 2018 14:06:12	U	SMPL
CCV68 - 1	CCV68	23988	5.0445	-	7471B	PPB	02 Nov 2018 14:08:29	U	SMPL
CCB68 - 1	CCB68	-206	-0.0144	-	7471B	PPB	02 Nov 2018 14:10:45	U	SMPL
J5745-06 - 1	6A	4113	0.8887	-	7471B	PPB	02 Nov 2018 14:13:01	U	SMPL
J5745-09 - 1	6B	481	0.1292	-	7471B	PPB	02 Nov 2018 14:15:18	U	SMPL
J5745-15 - 1	8A	103	0.0502	-	7471B	PPB	02 Nov 2018 14:17:34	U	SMPL
J5745-18 - 1	8B	565	0.1468	-	7471B	PPB	02 Nov 2018 14:19:50	U	SMPL
J5750-03 - 1	7A	362	0.1043	-	7471B	PPB	02 Nov 2018 14:22:07	U	SMPL
J5750-05 - 1	7B	415	0.1154	-	7471B	PPB	02 Nov 2018 14:24:23	U	SMPL
J5750-08 - 1	7C	83	0.0460	-	7471B	PPB	02 Nov 2018 14:26:41	U	SMPL
J5767-01 - 1	CL-02-110118-A	20017	4.2141	-	7471B	PPB	02 Nov 2018 14:28:59	U	SMPL
J5767-03 - 1	CL-02-110118-B	4870	1.0470	-	7471B	PPB	02 Nov 2018 14:31:17	U	SMPL
J5767-05 - 1	CL-02-110118-C	4499	0.9694	-	7471B	PPB	02 Nov 2018 14:33:35	U	SMPL
CCV69 - 1	CCV69	23420	4.9257	-	7471B	PPB	02 Nov 2018 14:35:51	U	SMPL
CCB69 - 1	CCB69	-164	-0.0056	-	7471B	PPB	02 Nov 2018 14:38:10	U	SMPL
J5767-07 - 1	CL-02-110118-D	3930	0.8504	-	7471B	PPB	02 Nov 2018 14:40:26	U	SMPL
PB114431BL - 1	PBS	-70	0.0140	-	7471B	PPB	02 Nov 2018 14:42:43	U	SMPL
PB114431BS - 1	LCSS	18468	3.8903	-	7471B	PPB	02 Nov 2018 14:44:59	U	SMPL
J5769-01 - 1	WAGARAW-RD-1	5237	1.1237	-	7471B	PPB	02 Nov 2018 14:47:17	U	SMPL
J5769-03 - 1	WAGARAW-RD-2	8226	1.7487	-	7471B	PPB	02 Nov 2018 14:49:34	U	SMPL
J5769-05 - 1	WAGARAW-RD-3	56584	11.8602	-	7471B	PPB	02 Nov 2018 14:51:51	U	SMPL
J5769-07 - 1	LEFT-CORNER-BULDG-85	1657	0.3751	-	7471B	PPB	02 Nov 2018 14:54:08	U	SMPL
J5769-09 - 1	FLOOD-AREA-1	9672	2.0510	-	7471B	PPB	02 Nov 2018 14:56:28	U	SMPL
J5769-11 - 1	FLOOD-AREA-2	3693	0.8008	-	7471B	PPB	02 Nov 2018 14:58:45	U	SMPL
J5769-13 - 1	FLOOD-AREA-3	12016	2.5412	-	7471B	PPB	02 Nov 2018 15:01:02	U	SMPL
CCV70 - 1	CCV70	23727	4.9899	-	7471B	PPB	02 Nov 2018 15:03:18	U	SMPL
CCB70 - 1	CCB70	-132	0.0011	-	7471B	PPB	02 Nov 2018 15:05:35	U	SMPL
J5769-15 - 1	GOFFLE-RD-1	1954	0.4372	-	7471B	PPB	02 Nov 2018 15:07:53	U	SMPL
J5769-17 - 1	GOFFLE-RD-2	10405	2.2043	-	7471B	PPB	02 Nov 2018 15:10:09	U	SMPL
J5769-19 - 1	GOFFLE-RD-3	15271	3.2218	-	7471B	PPB	02 Nov 2018 15:12:26	U	SMPL
J5776-01 - 1	NJ-GREEN	373	0.1066	-	7471B	PPB	02 Nov 2018 15:14:42	U	SMPL
J5776-05 - 1	NJ-WHITE	-69	0.0142	-	7471B	PPB	02 Nov 2018 15:16:59	U	SMPL
J5777-01 - 1	ST-PAULS	80194	16.7970	-	7471B	PPB	02 Nov 2018 15:19:16	U	SMPL
J5778-03 - 1	OILY-GRAVEL	-345	-0.0435	-	7471B	PPB	02 Nov 2018 15:21:34	U	SMPL
J5793-01 - 1	WJ-18-11-01-TSHG-18	1419	0.3254	-	7471B	PPB	02 Nov 2018 15:23:54	U	SMPL
J5793-02 - 1	WJ-18-11-01-TSHG-19	1452	0.3323	-	7471B	PPB	02 Nov 2018 15:26:12	U	SMPL
J5793-03 - 1	WJ-18-11-01-TSHG-20	967	0.2308	-	7471B	PPB	02 Nov 2018 15:28:28	U	SMPL
CCV71 - 1	CCV71	23825	5.0104	-	7471B	PPB	02 Nov 2018 15:30:45	U	SMPL
CCB71 - 1	CCB71	-214	-0.0161	-	7471B	PPB	02 Nov 2018 15:33:02	U	SMPL
J5794-03 - 1	TP-12B-S	486	0.1303	-	7471B	PPB	02 Nov 2018 15:35:19	U	SMPL
J5794-07 - 1	TP-11-S	1539	0.3505	-	7471B	PPB	02 Nov 2018 15:37:35	U	SMPL
J5794-11 - 1	TP-10B-S	1632	0.3699	-	7471B	PPB	02 Nov 2018 15:39:52	U	SMPL
J5794-11DUP - 1	TP-10B-SDUP	1562	0.3553	-	7471B	PPB	02 Nov 2018 15:42:08	U	SMPL
J5794-11MS - 1	TP-10B-SMS	20562	4.3281	-	7471B	PPB	02 Nov 2018 15:44:25	U	SMPL
J5794-11MSD - 1	TP-10B-SMSD	19544	4.1152	-	7471B	PPB	02 Nov 2018 15:46:42	U	SMPL
J5735-05LX5 - 1		202	0.0709	-	7471B	PPB	02 Nov 2018 15:49:00	U	SMPL
J5735-05A - 1		21577	4.5403	-	7471B	PPB	02 Nov 2018 15:51:17	U	SMPL
J5794-11LX5 - 1		152	0.0604	-	7471B	PPB	02 Nov 2018 15:53:35	U	SMPL
J5794-11A - 1		21545	4.5336	-	7471B	PPB	02 Nov 2018 15:55:52	U	SMPL
CCV72 - 1	CCV72	24192	5.0871	-	7471B	PPB	02 Nov 2018 15:58:09	U	SMPL
CCB72 - 1	CCB72	-168	-0.0065	-	7471B	PPB	02 Nov 2018 16:00:26	U	SMPL
J5769-05DLX2 - 1		27199	5.7159	-	7471B	PPB	02 Nov 2018 16:06:52	U	SMPL
J5777-01DLX5 - 1		16389	3.4555	-	7471B	PPB	02 Nov 2018 16:09:08	U	SMPL
CCV73 - 1	CCV73	24009	5.0489	-	7471B	PPB	02 Nov 2018 16:11:26	U	SMPL
CCB73 - 1	CCB73	-154	-0.0035	-	7471B	PPB	02 Nov 2018 16:13:43	U	SMPL