



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98997

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			23.2	4.01	11/02/2018	12:20
2	CAL2	1	Water			23.3	7.00	11/02/2018	12:22
3	CAL3	1	Water			22.9	10.04	11/02/2018	12:24
4	ICV	1	Water			23.5	7.01	11/02/2018	12:26
5	CCV1	1	Water			23.3	2.00	11/02/2018	12:28
6	J5799-01	1	Water			24.3	4.59	11/02/2018	12:30
7	J5799-01DUP	1	Water			23.4	4.60	11/02/2018	12:32
8	CCV2	1	Water			23.4	12.00	11/02/2018	12:34

LB 98997

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH110218

WorkList ID : 118786

Date : 11/2/2018 12:02:14 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/31/2018	Water	J5799-01 -H	pH	Cool 4 deg C	STER14	I41	MINI-CELL	10/31/2018	SM 4500-H B

Date/Time 11/02/18 12:05
Received by: SP
Relinquished by: CP

Date/Time 11/02/18 12:40
Received by: CP
Relinquished by: SP