



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB99006

Ambient Barometric Pressure (mmHg): 734

Barometric Scale ID: 55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	82.5	11/02/2018	09:50
2	J5785-01		1	8	76.0	168.8	170.4	11/02/2018	10:20
3	J5785-01DUP		1	8	76.0	168.8	170.4	11/02/2018	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 99006

WorkList Name : fp11218

WorkList ID : 118768

Date : 11/2/2018 9:20:52 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/07/2018	Water	J5778-07	Flash Point	Cool 4 deg C	PSEG01	I23	OILY-WATER	11/01/2018	1010A
11/07/2018	Solid	J5785-01	Flash Point	Cool 4 deg C	PSEG01	P63	OIL-SP	11/01/2018	1010A

Date/Time 11/02/18 9:25 AM
 Received by: KS
 Relinquished by: CP

Date/Time 11/02/18 10:50 AM
 Received by: CP
 Relinquished by: KS