



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99010

Ambient Barometric Pressure (mmHg): 729

Barometric Scale ID: 55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	82.5	11/02/2018	11:20
2	J5778-07		1	9	>100.0	>212.0	>212.0	11/02/2018	11:50
3	J5778-07DUP		1	9	>100.0	>212.0	>212.0	11/02/2018	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB99010

WorkList Name : fp11218

WorkList ID : 118768

Date : 11/2/2018 9:20:52 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/07/2018	✓ Water	J5778-07	Flash Point	Cool 4 deg C	PSEG01	I23	OILY-WATER	11/01/2018	1010A
11/07/2018	Solid	J5785-01	Flash Point	Cool 4 deg C	PSEG01	P63	OIL-SP	11/01/2018	1010A

Date/Time 11/02/18 9:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 11/02/18 12:25 PM
 Received by: CP
 Relinquished by: KS