



## Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB99011

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH meter 1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			24.3	4.01	11/02/2018	16:10
2	CAL2	1	Water			24.2	7.00	11/02/2018	16:12
3	CAL3	1	Water			24.5	10.04	11/02/2018	16:14
4	ICV	1	Water			24.3	7.01	11/02/2018	16:16
5	CCV1	1	Water			24.3	2.00	11/02/2018	16:18
6	J5813-01	1	Solid	20.02	20	25.4	6.70	11/02/2018	16:20
7	J5813-01DUP	1	Solid	20.02	20	25.5	6.71	11/02/2018	16:22
8	J5813-02	1	Solid	20.03	20	25.5	6.24	11/02/2018	16:23
9	J5813-03	1	Solid	20.02	20	25.5	6.70	11/02/2018	16:25
10	J5813-04	1	Solid	20.04	20	25.3	7.22	11/02/2018	16:27
11	J5813-05	1	Solid	20.04	20	25.4	6.90	11/02/2018	16:29
12	J5813-06	1	Solid	20.04	20	25.5	8.12	11/02/2018	16:31
13	J5813-07	1	Solid	20.02	20	25.4	7.30	11/02/2018	16:33
14	J5813-08	1	Solid	20.03	20	25.4	7.40	11/02/2018	16:35
15	J5813-09	1	Solid	20.02	20	25.5	7.32	11/02/2018	16:37
16	CCV2	1	Water			24.3	12.00	11/02/2018	16:39
17	J5813-10	1	Solid	20.02	20	25.3	7.08	11/02/2018	16:41
18	J5813-11	1	Solid	20.02	20	25.5	7.20	11/02/2018	16:43
19	J5813-12	1	Solid	20.04	20	25.6	7.56	11/02/2018	16:45
20	J5813-12DUP	1	Solid	20.04	20	25.7	7.57	11/02/2018	16:47

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water			24.6	2.00	11/02/2018	16:49

# WORKLIST(Hardcopy Internal Chain)

LB99011

WorkList Name : pH110218-1

WorkList ID : 118802

Date : 11/2/2018 3:59:49 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/02/2018	Solid	J5813-01	pH	Cool 4 deg C	ENTA05	I42	TS-A	11/02/2018	9045D
11/02/2018	Solid	J5813-02	pH	Cool 4 deg C	ENTA05	I42	TS-B	11/02/2018	9045D
11/02/2018	Solid	J5813-03	pH	Cool 4 deg C	ENTA05	I42	TS-C	11/02/2018	9045D
11/02/2018	Solid	J5813-04	pH	Cool 4 deg C	ENTA05	I42	TS-D	11/02/2018	9045D
11/02/2018	Solid	J5813-05	pH	Cool 4 deg C	ENTA05	I42	TS-E	11/02/2018	9045D
11/02/2018	Solid	J5813-06	pH	Cool 4 deg C	ENTA05	I42	TS-F	11/02/2018	9045D
11/02/2018	Solid	J5813-07	pH	Cool 4 deg C	ENTA05	I42	TS-G	11/02/2018	9045D
11/02/2018	Solid	J5813-08	pH	Cool 4 deg C	ENTA05	I42	TS-H	11/02/2018	9045D
11/02/2018	Solid	J5813-09	pH	Cool 4 deg C	ENTA05	I42	TS-I	11/02/2018	9045D
11/02/2018	Solid	J5813-10	pH	Cool 4 deg C	ENTA05	I42	TS-J	11/02/2018	9045D
11/02/2018	Solid	J5813-11	pH	Cool 4 deg C	ENTA05	I42	TS-K	11/02/2018	9045D
11/02/2018	Solid	J5813-12	pH	Cool 4 deg C	ENTA05	I42	TS-L	11/02/2018	9045D

Date/Time 11/02/18 4:02  
 Received by: SP  
 Relinquished by: CP

Date/Time 11/02/18 5:55  
 Received by: CP  
 Relinquished by: CP