


7470A

LB 99013
INSTRUMENT ID: CV1 μ Abs.:

35539



Linear

A= 0.0000e+000

B= 2.8268e-004

C= 3.4642e-002

Rho= 0.9998167

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.023	0.023	-42	0.000	-42				
0.2	0.200	0.195	-0.005	566	0.0 %	566				
2.5	2.500	2.496	-0.004	8706	0.0 %	8706				
5.0	5.000	5.049	0.049	17738	0.0 %	17738				
7.5	7.500	7.357	-0.143	25903	0.0 %	25903				
10.0	10.000	10.081	0.081	35539	0.0 %	35539				

LB99013 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Nov 2018 16:44:36

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-42	-	0.0000	7470A	PPB	02 Nov 2018 16:45:56	S	Std
0.2 - 1	58.2	566	-	0.2000	7470A	PPB	02 Nov 2018 16:48:05	S	Std
2.5 - 1	52.5	8706	-	2.5000	7470A	PPB	02 Nov 2018 16:50:13	S	Std
5.0 - 1	55	17738	-	5.0000	7470A	PPB	02 Nov 2018 16:52:21	S	Std
7.5 - 1	57.5	25903	-	7.5000	7470A	PPB	02 Nov 2018 16:54:30	S	Std
10.0 - 1	510	35539	-	10.0000	7470A	PPB	02 Nov 2018 16:56:38	S	Std
ICV55 - 1	ICV55	14056	4.0081	-	7470A	PPB	02 Nov 2018 16:59:26	U	SMPL
ICB55 - 1	ICB55	-127	-0.0013	-	7470A	PPB	02 Nov 2018 17:01:33	U	SMPL
CCV74 - 1	CCV74	18331	5.2165	-	7470A	PPB	02 Nov 2018 17:03:40	U	SMPL
CCB74 - 1	CCB74	-153	-0.0086	-	7470A	PPB	02 Nov 2018 17:05:47	U	SMPL
CRA - 1	CRA	524	0.1828	-	7470A	PPB	02 Nov 2018 17:07:54	U	SMPL
HighStd - 1	HighStd	35740	10.1378	-	7470A	PPB	02 Nov 2018 17:10:00	U	SMPL
ChkStd - 1	ChkStd	24495	6.9590	-	7470A	PPB	02 Nov 2018 17:15:19	U	SMPL
PB114406BL - 1	PBW	-213	-0.0256	-	7470A	PPB	02 Nov 2018 17:17:26	U	SMPL
PB114406BS - 1	LCSW	12459	3.5566	-	7470A	PPB	02 Nov 2018 17:19:34	U	SMPL
J5740-05 - 1	S-24	543	0.1881	-	7470A	PPB	02 Nov 2018 17:21:41	U	SMPL
J5740-06 - 1	S-25	237	0.1016	-	7470A	PPB	02 Nov 2018 17:23:49	U	SMPL
J5740-09 - 1	S-32	-3	0.0338	-	7470A	PPB	02 Nov 2018 17:25:58	U	SMPL
J5740-09DUP - 1	S-32DUP	-20	0.0290	-	7470A	PPB	02 Nov 2018 17:28:06	U	SMPL
J5740-09MS - 1	S-32MS	12752	3.6394	-	7470A	PPB	02 Nov 2018 17:30:13	U	SMPL
CCV75 - 1	CCV75	17865	5.0848	-	7470A	PPB	02 Nov 2018 17:32:19	U	SMPL
CCB75 - 1	CCB75	-83	0.0112	-	7470A	PPB	02 Nov 2018 17:34:26	U	SMPL
J5740-09MSD - 1	S-32MSD	12511	3.5713	-	7470A	PPB	02 Nov 2018 17:36:33	U	SMPL
PB114408BL - 1	PBW	-31	0.0259	-	7470A	PPB	02 Nov 2018 17:38:41	U	SMPL
PB114408BS - 1	LCSW	12483	3.5634	-	7470A	PPB	02 Nov 2018 17:40:48	U	SMPL
J5192-06 - 1	AU-05-103118-A	-112	0.0030	-	7470A	PPB	02 Nov 2018 17:42:55	U	SMPL
J5192-06DUP - 1	AU-05-103118-ADUP	-57	0.0185	-	7470A	PPB	02 Nov 2018 17:45:02	U	SMPL
J5192-06MS - 1	AU-05-103118-AMS	12428	3.5478	-	7470A	PPB	02 Nov 2018 17:47:10	U	SMPL
J5192-06MSD - 1	AU-05-103118-AMSD	12823	3.6595	-	7470A	PPB	02 Nov 2018 17:49:18	U	SMPL
J5192-08 - 1	AU-05-103118-B	-113	0.0027	-	7470A	PPB	02 Nov 2018 17:51:26	U	SMPL
J5200-18 - 1	OR-03-103118-A	-84	0.0109	-	7470A	PPB	02 Nov 2018 17:53:35	U	SMPL
J5200-20 - 1	OR-03-103118-B	-54	0.0194	-	7470A	PPB	02 Nov 2018 17:55:41	U	SMPL
CCV76 - 1	CCV76	17833	5.0758	-	7470A	PPB	02 Nov 2018 17:57:48	U	SMPL
CCB76 - 1	CCB76	-59	0.0180	-	7470A	PPB	02 Nov 2018 17:59:55	U	SMPL
J5741-06 - 1	TP-1	-71	0.0146	-	7470A	PPB	02 Nov 2018 18:02:02	U	SMPL
J5741-10 - 1	TP-2	-26	0.0273	-	7470A	PPB	02 Nov 2018 18:04:09	U	SMPL
J5741-14 - 1	TP-3	-44	0.0222	-	7470A	PPB	02 Nov 2018 18:06:16	U	SMPL
J5741-18 - 1	TP-4	-19	0.0293	-	7470A	PPB	02 Nov 2018 18:08:23	U	SMPL
J5741-22 - 1	TP-5	1150	0.3597	-	7470A	PPB	02 Nov 2018 18:10:30	U	SMPL
J5741-26 - 1	TP-6	-72	0.0143	-	7470A	PPB	02 Nov 2018 18:12:38	U	SMPL
J5741-30 - 1	TP-7	-23	0.0281	-	7470A	PPB	02 Nov 2018 18:14:46	U	SMPL
J5741-34 - 1	TP-8	-44	0.0222	-	7470A	PPB	02 Nov 2018 18:16:54	U	SMPL
J5741-38 - 1	TP-9	-17	0.0298	-	7470A	PPB	02 Nov 2018 18:19:03	U	SMPL
J5741-42 - 1	TP-10	-60	0.0177	-	7470A	PPB	02 Nov 2018 18:21:10	U	SMPL
CCV77 - 1	CCV77	18039	5.1340	-	7470A	PPB	02 Nov 2018 18:23:17	U	SMPL
CCB77 - 1	CCB77	-123	-0.0001	-	7470A	PPB	02 Nov 2018 18:25:24	U	SMPL
J5741-46 - 1	TP-11	-33	0.0253	-	7470A	PPB	02 Nov 2018 18:27:31	U	SMPL
J5741-50 - 1	TP-12	-51	0.0202	-	7470A	PPB	02 Nov 2018 18:29:38	U	SMPL
J5741-54 - 1	TP-13	-98	0.0069	-	7470A	PPB	02 Nov 2018 18:31:45	U	SMPL
J5741-58 - 1	TP-14	-56	0.0188	-	7470A	PPB	02 Nov 2018 18:33:53	U	SMPL
J5741-62 - 1	TP-15	-34	0.0250	-	7470A	PPB	02 Nov 2018 18:36:00	U	SMPL
PB114409BL - 1	PBW	-56	0.0188	-	7470A	PPB	02 Nov 2018 18:38:08	U	SMPL
PB114409BS - 1	LCSW	12281	3.5063	-	7470A	PPB	02 Nov 2018 18:40:16	U	SMPL
J5729-03 - 1	WS1	98	0.0623	-	7470A	PPB	02 Nov 2018 18:42:24	U	SMPL
J5755-13 - 1	TP-2A	701	0.2328	-	7470A	PPB	02 Nov 2018 18:44:32	U	SMPL
J5755-13DUP - 1	TP-2ADUP	740	0.2438	-	7470A	PPB	02 Nov 2018 18:46:40	U	SMPL
CCV78 - 1	CCV78	17747	5.0514	-	7470A	PPB	02 Nov 2018 18:48:47	U	SMPL
CCB78 - 1	CCB78	-97	0.0072	-	7470A	PPB	02 Nov 2018 18:50:54	U	SMPL
J5755-13MS - 1	TP-2AMS	14780	4.2127	-	7470A	PPB	02 Nov 2018 18:53:01	U	SMPL
J5755-13MSD - 1	TP-2AMSD	14550	4.1477	-	7470A	PPB	02 Nov 2018 18:55:08	U	SMPL
J5769-21 - 1	MW-1	-141	-0.0052	-	7470A	PPB	02 Nov 2018 18:57:15	U	SMPL
J5769-22 - 1	FB-10-31-18	-140	-0.0049	-	7470A	PPB	02 Nov 2018 18:59:23	U	SMPL
PB114406TB - 1	PB114360TB	-99	0.0067	-	7470A	PPB	02 Nov 2018 19:01:30	U	SMPL
PB114408TB - 1	PB114371TB	-63	0.0168	-	7470A	PPB	02 Nov 2018 19:03:38	U	SMPL
CCV79 - 1	CCV79	17907	5.0967	-	7470A	PPB	02 Nov 2018 19:05:46	U	SMPL
CCB79 - 1	CCB79	-104	0.0052	-	7470A	PPB	02 Nov 2018 19:07:54	U	SMPL
J5740-09LX5 - 1		0	0.0346	-	7470A	PPB	02 Nov 2018 19:10:03	U	SMPL
J5740-09A - 1		13627	3.8868	-	7470A	PPB	02 Nov 2018 19:12:10	U	SMPL
J5192-06LX5 - 1		-144	-0.0061	-	7470A	PPB	02 Nov 2018 19:14:17	U	SMPL
J5192-06A - 1		13782	3.9306	-	7470A	PPB	02 Nov 2018 19:16:25	U	SMPL
J5755-13LX5 - 1		-51	0.0202	-	7470A	PPB	02 Nov 2018 19:18:32	U	SMPL
J5755-13A - 1		15846	4.5141	-	7470A	PPB	02 Nov 2018 19:20:39	U	SMPL
CCV80 - 1	CCV80	17950	5.1088	-	7470A	PPB	02 Nov 2018 19:22:47	U	SMPL
CCB80 - 1	CCB80	-100	0.0064	-	7470A	PPB	02 Nov 2018 19:24:54	U	SMPL