



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99019

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			24.3	4.01	11/02/2018	16:10
2	CAL2	1	Water			24.2	7.00	11/02/2018	16:12
3	CAL3	1	Water			24.5	10.04	11/02/2018	16:14
4	ICV	1	Water			24.3	7.01	11/02/2018	16:16
5	CCV1	1	Water			24.3	2.00	11/02/2018	17:34
6	J5804-04	1	Water			24.4	6.02	11/02/2018	17:36
7	J5804-04DUP	1	Water			24.5	6.03	11/02/2018	17:38
8	CCV2	1	Water			24.5	12.00	11/02/2018	17:42

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity110218-2

WorkList ID : 118809

Date : 11/2/2018 5:20:43 PM

UB99019

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/02/2018	Water	J5804-04	Corrosivity	Cool 4 deg C	PSEG01	I21	BS-DRUM-3	11/02/2018	9040C

11/02/18
 Date/Time 11/02/18 5:25
 Received by: SP
 Relinquished by: CP

11/02/18
 Date/Time 11/02/18 6:45
 Received by: CP
 Relinquished by: SP