

7470A


LB99025

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37817



A= 0.0000e+000

B= 2.6335e-004

C= 6.8771e-002

Rho= 0.9998694

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.067	0.067	-5	0.000	-5				
0.2	0.200	0.228	0.028	603	0.0 %	603				
2.5	2.500	2.376	-0.124	8763	0.0 %	8763				
5.0	5.000	5.000	0.000	18727	0.0 %	18727				
7.5	7.500	7.500	0.000	28220	0.0 %	28220				
10.0	10.000	10.028	0.028	37817	0.0 %	37817				

LB99025 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Nov 2018 08:56:28

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-5	-	0.0000	7470A	PPB	05 Nov 2018 09:01:17	S	Std
0.2 - 1	50.2	603	-	0.2000	7470A	PPB	05 Nov 2018 09:03:25	S	Std
2.5 - 1	52.5	8763	-	2.5000	7470A	PPB	05 Nov 2018 09:05:34	S	Std
5.0 - 1	55	18727	-	5.0000	7470A	PPB	05 Nov 2018 09:07:42	S	Std
7.5 - 1	57.5	28220	-	7.5000	7470A	PPB	05 Nov 2018 09:09:50	S	Std
10.0 - 1	510	37817	-	10.0000	7470A	PPB	05 Nov 2018 09:11:57	S	Std
ICV56 - 1	ICV56	15328	4.1053	-	7470A	PPB	05 Nov 2018 09:14:41	U	SMPL
ICB56 - 1	ICB56	-79	0.0480	-	7470A	PPB	05 Nov 2018 09:16:47	U	SMPL
CCV81 - 1	CCV81	19713	5.2601	-	7470A	PPB	05 Nov 2018 09:18:53	U	SMPL
CCB81 - 1	CCB81	-89	0.0453	-	7470A	PPB	05 Nov 2018 09:21:00	U	SMPL
CRA - 1	CRA	592	0.2247	-	7470A	PPB	05 Nov 2018 09:23:07	U	SMPL
HighStd - 1	HighStd	38370	10.1733	-	7470A	PPB	05 Nov 2018 09:25:13	U	SMPL
ChkStd - 1	ChkStd	27058	7.1944	-	7470A	PPB	05 Nov 2018 09:27:20	U	SMPL
PB114457BL - 1	PBW	-12	0.0656	-	7470A	PPB	05 Nov 2018 09:29:28	U	SMPL
PB114457BS - 1	LCSW	13388	3.5944	-	7470A	PPB	05 Nov 2018 09:34:01	U	SMPL
J5734-03 - 1	B-21(0-10)	-26	0.0619	-	7470A	PPB	05 Nov 2018 09:36:09	U	SMPL
J5734-06 - 1	B-22(0-10)	80	0.0898	-	7470A	PPB	05 Nov 2018 09:38:17	U	SMPL
J5755-04 - 1	TP-1-S	177	0.1154	-	7470A	PPB	05 Nov 2018 09:40:25	U	SMPL
J5755-08 - 1	TP-2A-S	-29	0.0611	-	7470A	PPB	05 Nov 2018 09:42:34	U	SMPL
J5755-12 - 1	TP-2A-D	138	0.1051	-	7470A	PPB	05 Nov 2018 09:44:40	U	SMPL
CCV82 - 1	CCV82	18847	5.0320	-	7470A	PPB	05 Nov 2018 09:46:47	U	SMPL
CCB82 - 1	CCB82	-78	0.0482	-	7470A	PPB	05 Nov 2018 09:48:55	U	SMPL
J5759-04 - 1	TP-1	123	0.1012	-	7470A	PPB	05 Nov 2018 09:51:01	U	SMPL
J5759-08 - 1	TP-2	252	0.1351	-	7470A	PPB	05 Nov 2018 09:53:08	U	SMPL
J5759-12 - 1	TP-3	237	0.1312	-	7470A	PPB	05 Nov 2018 09:55:14	U	SMPL
J5759-16 - 1	TP-4	230	0.1293	-	7470A	PPB	05 Nov 2018 09:57:21	U	SMPL
J5759-20 - 1	TP-5	-76	0.0488	-	7470A	PPB	05 Nov 2018 09:59:29	U	SMPL
J5759-24 - 1	TP-6	10	0.0714	-	7470A	PPB	05 Nov 2018 10:01:36	U	SMPL
J5767-02 - 1	CL-02-110118-A	148	0.1077	-	7470A	PPB	05 Nov 2018 10:03:44	U	SMPL
J5767-04 - 1	CL-02-110118-B	225	0.1280	-	7470A	PPB	05 Nov 2018 10:05:53	U	SMPL
J5767-06 - 1	CL-02-110118-C	214	0.1251	-	7470A	PPB	05 Nov 2018 10:08:01	U	SMPL
J5767-08 - 1	CL-02-110118-D	240	0.1320	-	7470A	PPB	05 Nov 2018 10:10:09	U	SMPL
CCV83 - 1	CCV83	18722	4.9991	-	7470A	PPB	05 Nov 2018 10:12:15	U	SMPL
CCB83 - 1	CCB83	-89	0.0453	-	7470A	PPB	05 Nov 2018 10:14:22	U	SMPL
J5776-04 - 1	NJ-GREEN	-22	0.0630	-	7470A	PPB	05 Nov 2018 10:16:29	U	SMPL
J5776-08 - 1	NJ-WHITE	17	0.0732	-	7470A	PPB	05 Nov 2018 10:18:36	U	SMPL
J5776-08DUP - 1	NJ-WHITEDUP	12	0.0719	-	7470A	PPB	05 Nov 2018 10:20:43	U	SMPL
J5776-08MS - 1	NJ-WHITEMS	13992	3.7535	-	7470A	PPB	05 Nov 2018 10:22:50	U	SMPL
J5776-08MSD - 1	NJ-WHITEMSD	14344	3.8462	-	7470A	PPB	05 Nov 2018 10:24:58	U	SMPL
PB114458BL - 1	PBW	-90	0.0451	-	7470A	PPB	05 Nov 2018 10:27:06	U	SMPL
PB114458BS - 1	LCSW	13226	3.5518	-	7470A	PPB	05 Nov 2018 10:29:14	U	SMPL
J5778-06 - 1	OILY-WATER	-31	0.0606	-	7470A	PPB	05 Nov 2018 10:31:22	U	SMPL
J5780-01 - 1	CPS07381	-54	0.0546	-	7470A	PPB	05 Nov 2018 10:33:30	U	SMPL
J5780-02 - 1	CPS07381	-36	0.0593	-	7470A	PPB	05 Nov 2018 10:35:37	U	SMPL
CCV84 - 1	CCV84	18539	4.9509	-	7470A	PPB	05 Nov 2018 10:37:45	U	SMPL
CCB84 - 1	CCB84	-30	0.0609	-	7470A	PPB	05 Nov 2018 10:39:52	U	SMPL
J5780-03 - 1	CPS073082	-15	0.0648	-	7470A	PPB	05 Nov 2018 10:41:59	U	SMPL
J5780-04 - 1	CPS073082	15	0.0727	-	7470A	PPB	05 Nov 2018 10:44:06	U	SMPL
J5780-05 - 1	CPS113081	12	0.0719	-	7470A	PPB	05 Nov 2018 10:46:14	U	SMPL
J5780-06 - 1	CPS113081	-2	0.0682	-	7470A	PPB	05 Nov 2018 10:48:21	U	SMPL
J5780-07 - 1	CPS133081	-23	0.0627	-	7470A	PPB	05 Nov 2018 10:50:28	U	SMPL
J5780-07DUP - 1	CPS133081DUP	-42	0.0577	-	7470A	PPB	05 Nov 2018 10:52:36	U	SMPL
J5780-07MS - 1	CPS133081MS	14874	3.9858	-	7470A	PPB	05 Nov 2018 10:54:44	U	SMPL
J5780-07MSD - 1	CPS133081MSD	14722	3.9457	-	7470A	PPB	05 Nov 2018 10:56:53	U	SMPL
J5780-08 - 1	CPS133081	-91	0.0448	-	7470A	PPB	05 Nov 2018 10:59:01	U	SMPL
PB114360TB - 1		46	0.0809	-	7470A	PPB	05 Nov 2018 11:01:08	U	SMPL
CCV85 - 1	CCV85	18422	4.9201	-	7470A	PPB	05 Nov 2018 11:03:15	U	SMPL
CCB85 - 1	CCB85	-47	0.0564	-	7470A	PPB	05 Nov 2018 11:05:23	U	SMPL
J5776-08LX5 - 1		15	0.0727	-	7470A	PPB	05 Nov 2018 11:07:30	U	SMPL
J5776-08A - 1		14249	3.8212	-	7470A	PPB	05 Nov 2018 11:09:37	U	SMPL
J5780-07LX5 - 1		-23	0.0627	-	7470A	PPB	05 Nov 2018 11:11:44	U	SMPL
J5780-07A - 1		14948	4.0053	-	7470A	PPB	05 Nov 2018 11:13:52	U	SMPL
CCV86 - 1	CCV86	18519	4.9457	-	7470A	PPB	05 Nov 2018 11:15:59	U	SMPL
CCB86 - 1	CCB86	-92	0.0445	-	7470A	PPB	05 Nov 2018 11:18:07	U	SMPL