

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 11/05/2018

TEMP1 IN: 104 11/05/2018 07:15 TEMP OUT: 104 11/05/2018 08:20

TEMP2 IN: 104 11/05/2018 08:40 TEMP OUT: 104 11/05/2018 09:45

TEMP3 IN: 104 11/05/2018 11:30 TEMP OUT: 103 11/05/2018 13:10

TEMP4 IN: 103 11/05/2018 13:35 TEMP OUT: 103 11/05/2018 15:30

Run Number: LB99028

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99028BL	LB99028BL	2.5998	2.5998	100	2.5999	2.5999	2.5999	0.0001	1
2	LB99028BS	LB99028BS	2.4778	2.4778	100	2.5326	2.5326	2.5326	0.0548	548
3	J5804-03	BS-DRUM-3	1.4044	1.4044	250	1.4060	1.4060	1.4060	0.0016	6.4
4	J5804-03DUP	BS-DRUM-3DUP	1.3999	1.3999	250	1.4015	1.4015	1.4015	0.0016	6.4
5	J5806-01	RAV-RAINEY-AFTER	1.3934	1.3934	900	1.4192	1.4192	1.4192	0.0258	28.7
6	J5806-02	RAV-006	1.3997	1.3997	900	1.4014	1.4014	1.4014	0.0017	1.9
7	J5806-03	RAV-01D	1.4404	1.4404	900	1.4443	1.4443	1.4443	0.0039	4.3
8	J5806-04	RAV-01A	1.4013	1.4013	900	1.4044	1.4044	1.4044	0.0031	3.4
9	J5806-05	RAV-RAINEY-BEFORE	1.3985	1.3985	900	1.4023	1.4023	1.4023	0.0038	4.2
10	J5808-01	110118-SIDI-E-D-B	1.4155	1.4155	400	1.4271	1.4271	1.4271	0.0116	29
11	J5808-02	110118-SIDI-E-D-M	1.3998	1.3998	400	1.4077	1.4077	1.4077	0.0079	19.8
12	J5808-03	110118-SIDI-E-D-S	1.4060	1.4060	400	1.4130	1.4130	1.4130	0.0070	17.5
13	J5808-04	110118-DBRI-E-D-B	1.4043	1.4043	400	1.4139	1.4139	1.4139	0.0096	24
14	J5808-05	110118-DBRI-E-D-M	1.3999	1.3999	400	1.4072	1.4072	1.4072	0.0073	18.3
15	J5808-06	110118-DBRI-E-D-S	1.4102	1.4102	400	1.4155	1.4155	1.4155	0.0053	13.3

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 11/05/2018

TEMP1 IN: 104 11/05/2018 07:15 TEMP OUT: 104 11/05/2018 08:20

TEMP2 IN: 104 11/05/2018 08:40 TEMP OUT: 104 11/05/2018 09:45

TEMP3 IN: 104 11/05/2018 11:30 TEMP OUT: 103 11/05/2018 13:10

TEMP4 IN: 103 11/05/2018 13:35 TEMP OUT: 103 11/05/2018 15:30

Run Number: LB99028

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J5808-07	110118-DBRI-E-U-B	1.3981	1.3981	400	1.4185	1.4185	1.4185	0.0204	51
17	J5808-08	110118-DBRI-E-U-M	1.4112	1.4112	400	1.4219	1.4219	1.4219	0.0107	26.8
18	J5808-09	110118-DBRI-E-U-S	1.4033	1.4033	400	1.4085	1.4085	1.4085	0.0052	13
19	J5808-10	110118-TIPC-E-D-B	1.4097	1.4097	400	1.4228	1.4228	1.4228	0.0131	32.8
20	J5808-11	110118-TIPC-E-D-S	1.4031	1.4031	400	1.4164	1.4164	1.4164	0.0133	33.2
21	J5808-12	110118-TIPC-E-U-B	1.4062	1.4062	400	1.4186	1.4186	1.4186	0.0124	31
22	J5808-13	110118-TIPC-E-U-S	1.3973	1.3973	400	1.4085	1.4085	1.4085	0.0112	28
23	J5810-01	110218-DBRI-E-D-B	1.4004	1.4004	200	1.4091	1.4091	1.4091	0.0087	43.5
24	J5810-01DUP	110218-DBRI-E-D-BDUP	1.4034	1.4034	200	1.4118	1.4118	1.4118	0.0084	42
25	J5810-02	110218-DBRI-E-D-M	1.4097	1.4097	400	1.4173	1.4173	1.4173	0.0076	19
26	J5810-03	110218-DBRI-E-D-S	1.4080	1.4080	400	1.4153	1.4153	1.4153	0.0073	18.3
27	J5810-04	110218-SIDI-E-D-B	1.4014	1.4014	400	1.4137	1.4137	1.4137	0.0123	30.8
28	J5810-05	110218-SIDI-E-D-M	1.4032	1.4032	400	1.4127	1.4127	1.4127	0.0095	23.7
29	J5810-06	110218-SIDI-E-D-S	1.4052	1.4052	400	1.4128	1.4128	1.4128	0.0076	19
30	J5810-07	110218-SIDI-E-U-B	1.4052	1.4052	400	1.4158	1.4158	1.4158	0.0106	26.5

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 11/05/2018

TEMP1 IN: 104 11/05/2018 07:15 TEMP OUT: 104 11/05/2018 08:20

TEMP2 IN: 104 11/05/2018 08:40 TEMP OUT: 104 11/05/2018 09:45

TEMP3 IN: 104 11/05/2018 11:30 TEMP OUT: 103 11/05/2018 13:10

TEMP4 IN: 103 11/05/2018 13:35 TEMP OUT: 103 11/05/2018 15:30

Run Number: LB99028

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
31	J5810-08	110218-SIDI-E-U-M	1.4019	1.4019	400	1.4098	1.4098	1.4098	0.0079	19.8
32	J5810-09	110218-SIDI-E-U-S	1.4073	1.4073	400	1.4123	1.4123	1.4123	0.0050	12.5
33	J5810-10	110218-TIPC-E-D-B	1.4010	1.4010	400	1.4111	1.4111	1.4111	0.0101	25.3
34	J5810-11	110218-TIPC-E-D-S	1.4032	1.4032	400	1.4122	1.4122	1.4122	0.0090	22.5
35	J5810-12	110218-TIPC-E-U-B	1.3990	1.3990	400	1.4102	1.4102	1.4102	0.0112	28
36	J5810-13	110218-TIPC-E-U-S	1.4032	1.4032	400	1.4126	1.4126	1.4126	0.0094	23.5
37	J5818-01	EFF-WW	1.4005	1.4005	300	1.4054	1.4054	1.4054	0.0049	16.3
38	J5818-01DUP	EFF-WWDUP	1.4128	1.4128	300	1.4179	1.4179	1.4179	0.0051	17

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tss j5804

WorkList ID : 118837

Date : 11/5/2018 7:45:54 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/07/2018	Water	J5804-03- <i>F</i>	TSS	Cool 4 deg C	PSEG01	I21	BS-DRUM-3	11/02/2018	SM2540 D
11/07/2018	Water	J5806-01- <i>B</i>	TSS	Cool 4 deg C	FOST03	I13	RAV-RAINEY-AFTER	10/31/2018	SM2540 D
11/07/2018	Water	J5806-02- <i>A</i>	TSS	Cool 4 deg C	FOST03	I13	RAV-006	10/31/2018	SM2540 D
11/07/2018	Water	J5806-03- <i>B</i>	TSS	Cool 4 deg C	FOST03	I13	RAV-01D	10/31/2018	SM2540 D
11/07/2018	Water	J5806-04- <i>B</i>	TSS	Cool 4 deg C	FOST03	I13	RAV-01A	10/31/2018	SM2540 D
11/07/2018	Water	J5806-05- <i>B</i>	TSS	Cool 4 deg C	FOST03	I13	RAV-RAINEY-BEFORE	10/31/2018	SM2540 D
11/08/2018	Water	J5808-01- <i>C</i>	TSS	Cool 4 deg C	HDR103	I23	110118-SIDIE-D-B	11/01/2018	SM2540 D
11/08/2018	Water	J5808-02- <i>C</i>	TSS	Cool 4 deg C	HDR103	I23	110118-SIDIE-D-M	11/01/2018	SM2540 D
11/08/2018	Water	J5808-03- <i>C</i>	TSS	Cool 4 deg C	HDR103	I23	110118-SIDIE-D-S	11/01/2018	SM2540 D
11/08/2018	Water	J5808-04- <i>C</i>	TSS	Cool 4 deg C	HDR103	I23	110118-DBRI-E-D-B	11/01/2018	SM2540 D
11/08/2018	Water	J5808-05- <i>C</i>	TSS	Cool 4 deg C	HDR103	I23	110118-DBRI-E-D-M	11/01/2018	SM2540 D
11/08/2018	Water	J5808-06- <i>C</i>	TSS	Cool 4 deg C	HDR103	I23	110118-DBRI-E-D-S	11/01/2018	SM2540 D
11/08/2018	Water	J5808-07- <i>C</i>	TSS	Cool 4 deg C	HDR103	I23	110118-DBRI-E-U-B	11/01/2018	SM2540 D
11/08/2018	Water	J5808-08- <i>C</i>	TSS	Cool 4 deg C	HDR103	I23	110118-DBRI-E-U-M	11/01/2018	SM2540 D
11/08/2018	Water	J5808-09- <i>C</i>	TSS	Cool 4 deg C	HDR103	I23	110118-DBRI-E-U-S	11/01/2018	SM2540 D
11/06/2018	Water	J5808-10- <i>C</i>	TSS	Cool 4 deg C	HDR103	I23	110118-TIPC-E-D-B	11/01/2018	SM2540 D
11/06/2018	Water	J5808-11- <i>C</i>	TSS	Cool 4 deg C	HDR103	I23	110118-TIPC-E-D-S	11/01/2018	SM2540 D
11/06/2018	Water	J5808-12- <i>C</i>	TSS	Cool 4 deg C	HDR103	I23	110118-TIPC-E-U-B	11/01/2018	SM2540 D
11/06/2018	Water	J5808-13- <i>C</i>	TSS	Cool 4 deg C	HDR103	I23	110118-TIPC-E-U-S	11/01/2018	SM2540 D
11/09/2018	Water	J5810-01- <i>C</i>	TSS	Cool 4 deg C	HDR103	I42	110218-DBRI-E-D-B	11/02/2018	SM2540 D
11/09/2018	Water	J5810-02- <i>C</i>	TSS	Cool 4 deg C	HDR103	I42	110218-DBRI-E-D-M	11/02/2018	SM2540 D

Date/Time 11-05-18 7:50 PM

Received by: JG

Relinquished by: JG

Date/Time 11-05-18 11:40 PM

Received by: JG

Relinquished by: JG

1309028

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tss j5804

WorkList ID : 118837

Date : 11/5/2018 7:45:54 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/09/2018	Water	J5810-03- C	TSS	Cool 4 deg C	HDR103	I42	110218-DBRI-E-D-S	11/02/2018	SM2540 D
11/09/2018	Water	J5810-04- C	TSS	Cool 4 deg C	HDR103	I42	110218-SIDI-E-D-B	11/02/2018	SM2540 D
11/09/2018	Water	J5810-05- C	TSS	Cool 4 deg C	HDR103	I42	110218-SIDI-E-D-M	11/02/2018	SM2540 D
11/09/2018	Water	J5810-06- C	TSS	Cool 4 deg C	HDR103	I42	110218-SIDI-E-D-S	11/02/2018	SM2540 D
11/09/2018	Water	J5810-07- C	TSS	Cool 4 deg C	HDR103	I42	110218-SIDI-E-U-B	11/02/2018	SM2540 D
11/09/2018	Water	J5810-08- C	TSS	Cool 4 deg C	HDR103	I42	110218-SIDI-E-U-M	11/02/2018	SM2540 D
11/09/2018	Water	J5810-09- C	TSS	Cool 4 deg C	HDR103	I42	110218-SIDI-E-U-S	11/02/2018	SM2540 D
11/06/2018	Water	J5810-10- C	TSS	Cool 4 deg C	HDR103	I42	110218-TIPC-E-D-B	11/02/2018	SM2540 D
11/06/2018	Water	J5810-11- C	TSS	Cool 4 deg C	HDR103	I42	110218-TIPC-E-D-S	11/02/2018	SM2540 D
11/06/2018	Water	J5810-12- C	TSS	Cool 4 deg C	HDR103	I42	110218-TIPC-E-U-B	11/02/2018	SM2540 D
11/06/2018	Water	J5810-13- C	TSS	Cool 4 deg C	HDR103	I42	110218-TIPC-E-U-S	11/02/2018	SM2540 D
11/09/2018	Water	J5818-01- B	TSS	Cool 4 deg C	ARDM01	I31	EFF-VWV	11/02/2018	SM2540 D

Date/Time 11-05-18 11:50 AM

Received by: JP

Relinquished by: _____

Date/Time 11-05-18 11:40 AM

Received by: JP

Relinquished by: JP