



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99033

Ambient Barometric Pressure (mmHg): 740

Barometric Scale ID: 55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	81.8	11/05/2018	12:45
2	J5804-04		1	8	>100.0	>212.0	>212.0	11/05/2018	13:05
3	J5804-04DUP		1	8	>100.0	>212.0	>212.0	11/05/2018	13:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB99033

WorkList Name : FP11518

WorkList ID : 118855

Date : 11/5/2018 9:30:28 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/07/2018	Water	J5804-04	C Flash Point	Cool 4 deg C	PSEG01	I21	BS-DRUM-3	11/02/2018	1010A

Date/Time 11/5/18 9:35 AM
 Received by: KS
 Relinquished by: CP

Date/Time 11/5/18 13:40
 Received by: CP
 Relinquished by: KS