

LB 99035


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

37476



A= 0.0000e+000

B= 2.6668e-004

C= 3.1237e-002

Rho= 0.9997863

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	-106	0.000	-106				
0.05	0.050	0.057	0.007	97	0.0 %	97				
0.2	0.200	0.179	-0.021	554	0.0 %	554				
2.5	2.500	2.461	-0.039	9112	0.0 %	9112				
5.0	5.000	5.151	0.151	19197	0.0 %	19197				
7.5	7.500	7.374	-0.126	27534	0.0 %	27534				
10.0	10.000	10.025	0.025	37476	0.0 %	37476				

LB99035 INSTRUMENT : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 05 Nov 2018 13:52:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	5.0	-106	-	0.0000	245.1	PPB	05 Nov 2018 13:53:44	S	Std
0.05 - 1	50.05	97	-	0.0500	245.1	PPB	05 Nov 2018 13:56:01	S	Std
0.2 - 1	50.2	554	-	0.2000	245.1	PPB	05 Nov 2018 13:58:19	S	Std
2.5 - 1	52.5	9112	-	2.5000	245.1	PPB	05 Nov 2018 14:00:37	S	Std
5.0 - 1	15	19197	-	5.0000	245.1	PPB	05 Nov 2018 14:02:54	S	Std
7.5 - 1	52.5	27534	-	7.5000	245.1	PPB	05 Nov 2018 14:05:10	S	Std
10.0 - 1	510	37476	-	10.0000	245.1	PPB	05 Nov 2018 14:07:28	S	Std
ICV57 - 1	ICV57	14214	3.8218	-245.1		PPB	05 Nov 2018 14:10:20	U	SMPL
ICB57 - 1	ICB57	-256	-0.0370	-245.1		PPB	05 Nov 2018 14:12:35	U	SMPL
CCV87 - 1	CCV87	19043	5.1096	-245.1		PPB	05 Nov 2018 14:14:51	U	SMPL
CCB87 - 1	CCB87	-269	-0.0405	-245.1		PPB	05 Nov 2018 14:17:07	U	SMPL
CRA - 1	CRA	59	0.0470	-245.1		PPB	05 Nov 2018 14:19:23	U	SMPL
HighStd - 1	HighStd	36412	9.7415	-245.1		PPB	05 Nov 2018 14:21:38	U	SMPL
ChkStd - 1	ChkStd	25707	6.8867	-245.1		PPB	05 Nov 2018 14:23:55	U	SMPL
PB114459BL - 1	PBW	-107	0.0027	-245.1		PPB	05 Nov 2018 14:26:13	U	SMPL
PB114459BS - 1	LCSW	14113	3.7948	-245.1		PPB	05 Nov 2018 14:28:30	U	SMPL
J5789-01 - 1	103118-DBRI-E-D-B	-59	0.0155	-245.1		PPB	05 Nov 2018 14:30:47	U	SMPL
J5789-01DUP - 1	103118-DBRI-E-D-BDUP	36	0.0408	-245.1		PPB	05 Nov 2018 14:33:04	U	SMPL
J5789-01MS - 1	103118-DBRI-E-D-BMS	15781	4.2397	-245.1		PPB	05 Nov 2018 14:35:22	U	SMPL
J5789-01MSD - 1	103118-DBRI-E-D-BMSD	16393	4.4029	-245.1		PPB	05 Nov 2018 14:37:40	U	SMPL
J5789-02 - 1	103118-DBRI-E-D-M	5	0.0326	-245.1		PPB	05 Nov 2018 14:39:56	U	SMPL
CCV88 - 1	CCV88	19299	5.1778	-245.1		PPB	05 Nov 2018 14:42:12	U	SMPL
CCB88 - 1	CCB88	-259	-0.0378	-245.1		PPB	05 Nov 2018 14:44:28	U	SMPL
J5789-03 - 1	103118-DBRI-E-D-S	-31	0.0230	-245.1		PPB	05 Nov 2018 14:46:45	U	SMPL
J5789-04 - 1	103118-SIDI-E-D-B	0	0.0312	-245.1		PPB	05 Nov 2018 14:49:01	U	SMPL
J5789-05 - 1	103118-SIDI-E-D-M	54	0.0456	-245.1		PPB	05 Nov 2018 14:51:17	U	SMPL
J5789-06 - 1	103118-SIDI-E-D-S	40	0.0419	-245.1		PPB	05 Nov 2018 14:53:34	U	SMPL
J5789-07 - 1	103118-DBRI-E-U-B	46	0.0435	-245.1		PPB	05 Nov 2018 14:55:51	U	SMPL
J5789-08 - 1	103118-DBRI-E-U-M	-42	0.0200	-245.1		PPB	05 Nov 2018 14:58:08	U	SMPL
J5789-09 - 1	103118-DBRI-E-U-S	-43	0.0198	-245.1		PPB	05 Nov 2018 15:00:25	U	SMPL
J5789-10 - 1	103118-TIPC-E-D-M	-33	0.0224	-245.1		PPB	05 Nov 2018 15:02:43	U	SMPL
J5789-11 - 1	103118-TIPC-E-U-M	-35	0.0219	-245.1		PPB	05 Nov 2018 15:05:01	U	SMPL
CCV89 - 1	CCV89	19057	5.1133	-245.1		PPB	05 Nov 2018 15:07:17	U	SMPL
CCB89 - 1	CCB89	-237	-0.0320	-245.1		PPB	05 Nov 2018 15:09:33	U	SMPL
J5789-01LX5 - 1		-113	0.0011	-245.1		PPB	05 Nov 2018 15:11:50	U	SMPL
J5789-01A - 1		17017	4.5693	-245.1		PPB	05 Nov 2018 15:14:06	U	SMPL
CCV90 - 1	CCV90	19145	5.1368	-245.1		PPB	05 Nov 2018 15:16:21	U	SMPL
CCB90 - 1	CCB90	-254	-0.0365	-245.1		PPB	05 Nov 2018 15:18:38	U	SMPL