

LB 99042


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53354



A= 0.0000e+000

B= 1.8758e-004

C= 3.2560e-002

Rho= 0.9999662

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.041	0.041	45	0.000	45				
0.2	0.200	0.198	-0.002	883	0.0 %	883				
2.5	2.500	2.474	-0.026	13017	0.0 %	13017				
5.0	5.000	4.977	-0.023	26357	0.0 %	26357				
7.5	7.500	7.469	-0.031	39643	0.0 %	39643				
10.0	10.000	10.041	0.041	53354	0.0 %	53354				

LB99042 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Nov 2018 16:00:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	45	-	0.0000	7471B	PPB	05 Nov 2018 16:03:58	S	Std
0.2 - 1	50.2	883	-	0.2000	7471B	PPB	05 Nov 2018 16:06:16	S	Std
2.5 - 1	52.5	13017	-	2.5000	7471B	PPB	05 Nov 2018 16:08:34	S	Std
5.0 - 1	55	26357	-	5.0000	7471B	PPB	05 Nov 2018 16:10:51	S	Std
7.5 - 1	57.5	39643	-	7.5000	7471B	PPB	05 Nov 2018 16:13:09	S	Std
10.0 - 1	50	53354	-	10.0000	7471B	PPB	05 Nov 2018 16:15:26	S	Std
ICV58 - 1	ICV58	20880	3.9493	-	7471B	PPB	05 Nov 2018 16:21:02	U	SMPL
ICB58 - 1	ICB58	-128	0.0085	-	7471B	PPB	05 Nov 2018 16:23:18	U	SMPL
CCV91 - 1	CCV91	24665	4.6593	-	7471B	PPB	05 Nov 2018 16:25:34	U	SMPL
CCB91 - 1	CCB91	-148	0.0048	-	7471B	PPB	05 Nov 2018 16:27:50	U	SMPL
CRA - 1	CRA	886	0.1988	-	7471B	PPB	05 Nov 2018 16:30:08	U	SMPL
HighStd - 1	HighStd	53871	10.1379	-	7471B	PPB	05 Nov 2018 16:32:23	U	SMPL
ChkStd - 1	ChkStd	37130	6.9975	-	7471B	PPB	05 Nov 2018 16:34:40	U	SMPL
PB114462BL - 1	PBS	-180	-0.0012	-	7471B	PPB	05 Nov 2018 16:36:59	U	SMPL
PB114462BS - 1	LCSS	19241	3.6418	-	7471B	PPB	05 Nov 2018 16:39:17	U	SMPL
J5812-01 - 1	TP1-A	961	0.2128	-	7471B	PPB	05 Nov 2018 16:41:34	U	SMPL
J5812-05 - 1	TP1-B	968	0.2141	-	7471B	PPB	05 Nov 2018 16:43:53	U	SMPL
J5812-09 - 1	TP2-A	410	0.1095	-	7471B	PPB	05 Nov 2018 16:46:11	U	SMPL
J5812-13 - 1	TP2-B	542	0.1342	-	7471B	PPB	05 Nov 2018 16:48:28	U	SMPL
J5812-17 - 1	TP3-A	554	0.1365	-	7471B	PPB	05 Nov 2018 16:50:44	U	SMPL
CCV92 - 1	CCV92	25840	4.8797	-	7471B	PPB	05 Nov 2018 16:53:01	U	SMPL
CCB92 - 1	CCB92	-146	0.0052	-	7471B	PPB	05 Nov 2018 16:55:16	U	SMPL
J5812-21 - 1	TP3-B	698	0.1635	-	7471B	PPB	05 Nov 2018 16:57:34	U	SMPL
J5812-25 - 1	TP4-A	835	0.1892	-	7471B	PPB	05 Nov 2018 16:59:50	U	SMPL
J5812-29 - 1	TP4-B	1018	0.2235	-	7471B	PPB	05 Nov 2018 17:02:06	U	SMPL
J5813-01 - 1	TS-A	6028	1.1633	-	7471B	PPB	05 Nov 2018 17:04:22	U	SMPL
J5813-02 - 1	TS-B	3946	0.7728	-	7471B	PPB	05 Nov 2018 17:06:39	U	SMPL
J5813-03 - 1	TS-C	4672	0.9089	-	7471B	PPB	05 Nov 2018 17:08:56	U	SMPL
J5813-04 - 1	TS-D	5580	1.0793	-	7471B	PPB	05 Nov 2018 17:11:13	U	SMPL
J5813-05 - 1	TS-E	4467	0.8705	-	7471B	PPB	05 Nov 2018 17:13:30	U	SMPL
J5813-06 - 1	TS-F	4739	0.9215	-	7471B	PPB	05 Nov 2018 17:15:48	U	SMPL
J5813-07 - 1	TS-G	5074	0.9844	-	7471B	PPB	05 Nov 2018 17:18:04	U	SMPL
CCV93 - 1	CCV93	24274	4.5860	-	7471B	PPB	05 Nov 2018 17:20:21	U	SMPL
CCB93 - 1	CCB93	-156	0.0033	-	7471B	PPB	05 Nov 2018 17:22:37	U	SMPL
J5813-08 - 1	TS-H	128510	24.1389	-	7471B	PPB	05 Nov 2018 17:24:54	U	SMPL
J5813-09 - 1	TS-I	3324	0.6561	-	7471B	PPB	05 Nov 2018 17:27:10	U	SMPL
J5813-10 - 1	TS-J	5184	1.0050	-	7471B	PPB	05 Nov 2018 17:29:30	U	SMPL
J5813-11 - 1	TS-K	3010	0.5972	-	7471B	PPB	05 Nov 2018 17:31:47	U	SMPL
J5813-12 - 1	TS-L	4847	0.9418	-	7471B	PPB	05 Nov 2018 17:34:04	U	SMPL
J5813-12DUP - 1	TS-LDUP	5240	1.0155	-	7471B	PPB	05 Nov 2018 17:36:21	U	SMPL
J5813-12MS - 1	TS-LMS	25064	4.7341	-	7471B	PPB	05 Nov 2018 17:38:38	U	SMPL
J5813-12MSD - 1	TS-LMSD	24456	4.6201	-	7471B	PPB	05 Nov 2018 17:40:55	U	SMPL
PB114463BL - 1	PBS	-138	0.0067	-	7471B	PPB	05 Nov 2018 17:43:14	U	SMPL
PB114463BS - 1	LCSS	18846	3.5678	-	7471B	PPB	05 Nov 2018 17:45:32	U	SMPL
CCV94 - 1	CCV94	25887	4.8885	-	7471B	PPB	05 Nov 2018 17:47:48	U	SMPL
CCB94 - 1	CCB94	-127	0.0087	-	7471B	PPB	05 Nov 2018 17:50:05	U	SMPL
J5764-01 - 1	JC-03-110218-A	37384	7.0452	-	7471B	PPB	05 Nov 2018 17:52:24	U	SMPL
J5764-03 - 1	JC-03-110218-B	12509	2.3790	-	7471B	PPB	05 Nov 2018 17:54:40	U	SMPL
J5764-05 - 1	JC-03-110218-C	11423	2.1753	-	7471B	PPB	05 Nov 2018 17:56:59	U	SMPL
J5764-07 - 1	JC-03-110218-D	10296	1.9639	-	7471B	PPB	05 Nov 2018 17:59:17	U	SMPL
J5764-09 - 1	JC-03-110218-E	51111	9.6201	-	7471B	PPB	05 Nov 2018 18:01:33	U	SMPL
J5764-11 - 1	JC-03-110218-F	153252	28.7800	-	7471B	PPB	05 Nov 2018 18:03:51	U	SMPL
J5764-13 - 1	JC-03-110218-G	18717	3.5436	-	7471B	PPB	05 Nov 2018 18:06:11	U	SMPL
J5764-15 - 1	JC-03-110218-H	28868	5.4477	-	7471B	PPB	05 Nov 2018 18:08:33	U	SMPL
J5764-17 - 1	JC-03-110218-I	10381	1.9799	-	7471B	PPB	05 Nov 2018 18:10:51	U	SMPL
J5764-17DUP - 1	JC-03-110218-IDUP	10302	1.9650	-	7471B	PPB	05 Nov 2018 18:15:47	U	SMPL
CCV95 - 1	CCV95	25883	4.8878	-	7471B	PPB	05 Nov 2018 18:18:03	U	SMPL
CCB95 - 1	CCB95	-133	0.0076	-	7471B	PPB	05 Nov 2018 18:20:19	U	SMPL
J5764-17MS - 1	JC-03-110218-IMS	31427	5.9277	-	7471B	PPB	05 Nov 2018 18:22:37	U	SMPL
J5764-17MSD - 1	JC-03-110218-IMSD	31372	5.9174	-	7471B	PPB	05 Nov 2018 18:24:53	U	SMPL
J5764-19 - 1	JC-03-110218-J	26049	4.9189	-	7471B	PPB	05 Nov 2018 18:27:11	U	SMPL
J5771-01 - 1	EO-01-110218-A	3252	0.6426	-	7471B	PPB	05 Nov 2018 18:29:29	U	SMPL
J5771-03 - 1	EO-01-110218-B	3740	0.7341	-	7471B	PPB	05 Nov 2018 18:31:48	U	SMPL
J5771-05 - 1	EO-01-110218-C	3558	0.7000	-	7471B	PPB	05 Nov 2018 18:34:05	U	SMPL
J5803-01 - 1	SODUS-TS-001	3929	0.7696	-	7471B	PPB	05 Nov 2018 18:36:22	U	SMPL
J5803-02 - 1	SODUS-TS-002	4233	0.8266	-	7471B	PPB	05 Nov 2018 18:38:40	U	SMPL
J5815-01 - 1	HL-R37719-RT	9250	1.7677	-	7471B	PPB	05 Nov 2018 18:40:58	U	SMPL
J5815-03 - 1	HL-R47183-RT	7165	1.3766	-	7471B	PPB	05 Nov 2018 18:43:15	U	SMPL
CCV96 - 1	CCV96	25728	4.8587	-	7471B	PPB	05 Nov 2018 18:45:32	U	SMPL
CCB96 - 1	CCB96	-94	0.0149	-	7471B	PPB	05 Nov 2018 18:47:49	U	SMPL
J5815-05 - 1	HL-R25951-RT	1123	0.2432	-	7471B	PPB	05 Nov 2018 18:50:08	U	SMPL
J5816-03 - 1	TP-9-S	1093	0.2376	-	7471B	PPB	05 Nov 2018 18:52:24	U	SMPL
J5816-07 - 1	TP-4A-S	2454	0.4929	-	7471B	PPB	05 Nov 2018 18:54:41	U	SMPL
J5816-11 - 1	TP-4A-D	5436	1.0523	-	7471B	PPB	05 Nov 2018 18:56:58	U	SMPL
J5813-12LX5 - 1		952	0.2111	-	7471B	PPB	05 Nov 2018 19:07:52	U	SMPL
J5813-12A - 1		26559	5.0146	-	7471B	PPB	05 Nov 2018 19:10:09	U	SMPL
J5764-17LX5 - 1		1947	0.3978	-	7471B	PPB	05 Nov 2018 19:12:27	U	SMPL
J5764-17A - 1		33481	6.3130	-	7471B	PPB	05 Nov 2018 19:14:46	U	SMPL
J5813-08DLX5 - 1		23883	4.5126	-	7471B	PPB	05 Nov 2018 19:17:03	U	SMPL
J5764-11DLX5 - 1		29028	5.4777	-	7471B	PPB	05 Nov 2018 19:19:22	U	SMPL
CCV97 - 1	CCV97	25723	4.8578	-	7471B	PPB	05 Nov 2018 19:21:40	U	SMPL

LB99042 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Nov 2018 16:00:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB97 - 1	CCB97	-124	0.0093		-7471B	PPB	05 Nov 2018 19:23:59	U	SMPL