



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB99043

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	81.8	11/05/2018	14:30
2	J5807-07		1	8	60.0	140.0	141.2	11/05/2018	15:02
3	J5807-07DUP		1	8	59.0	138.2	139.4	11/05/2018	15:32

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06