



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB99048
Analysis Date: 11/06/2018
BalanceID: WC SC-3
OvenID: EX OVEN-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 11/06/2018
Extraction IN Time: 08:10
Extraction OUT Time: 09:20

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB99048BL	LB99048BL	WATER	2	1000	100	2.4568	2.4568	3.02	2.4568	2.4568	0.0000	0
2	LB99048BS	LB99048BS	WATER	2	1000	100	3.1023	3.1023	3.01	3.1210	3.1210	0.0187	18.7
3	LB99048BSD	LB99048BSD	WATER	2	1000	100	3.0399	3.0399	3.03	3.0589	3.0589	0.0190	19
4	J5825-01	DSN001-SEWER-OUTLET-00	WATER	2	1000	100	3.0069	3.0069	3.04	3.0156	3.0156	0.0087	8.7
5	J5825-02	DSN002-SEWER-OUTLET-00	WATER	2	1000	100	3.1349	3.1349	3.03	3.1474	3.1474	0.0125	12.5



QC Batch# LB99048 Test: Non-Polar Ma Analysis Date: 11/06/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2433
pH Paper 0-14	W2416
Sodium Sulfate	EP1897
1:1 HCL	WP69312
Silica Gel	W2368
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP69313
LCSWD	10.00 ML	WP69314
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0018</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>10:46</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>09:45</u>		
Bal Check Time:	<u>08:40</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>11:25</u>
		Out Time:	<u>10:45</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>12:36</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>11:50</u>		
Bal Check Time:	<u>13:40</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>13:10</u>
		Out Time:	<u>12:35</u>		

WorkList Name : NON POPLAR J5825

WorkList ID : 118891

Date : 11/6/2018 7:43:18 AM

WORKLIST(Hardcopy Internal Chain)

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/19/2018	Water	J5825-01-45	Non-Polar Material	Conc H2SO4 to	TRIS02	J13	DSN001-SEWER-OUTLET-00	11/05/2018	1664A
11/19/2018	Water	J5825-02 -45	Non-Polar Material	Conc H2SO4 to	TRIS02	J13	DSN002-SEWER-OUTLET-00	11/05/2018	1664A

Date/Time 11-06-18 7:50 AM
Received by: JP
Relinquished by: CP

Date/Time 11-06-18 10:00 AM
Received by: CP
Relinquished by: JP