



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99052

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			19.2	4.00	11/05/2018	18:00
2	CAL2	1	Water			19.1	7.02	11/05/2018	18:02
3	CAL3	1	Water			19.1	10.04	11/05/2018	18:04
4	ICV	1	Water			19.4	7.03	11/05/2018	18:06
5	CCV1	1	Water			19.2	2.00	11/05/2018	18:08
6	J5765-02	1	Solid	20.02	20	20.2	7.31	11/05/2018	18:10
7	J5765-02DUP	1	Solid	20.03	20	20.3	7.32	11/05/2018	18:12
8	J5772-02	1	Solid	20.02	20	20.2	7.14	11/05/2018	18:14
9	J5819-09	1	Solid	20.02	20	20.3	8.34	11/05/2018	18:16
10	J5826-04	1	Solid	20.04	20	20.1	7.48	11/05/2018	18:18
11	J5826-08	1	Solid	20.02	20	20.2	7.04	11/05/2018	18:20
12	J5826-12	1	Solid	20.02	20	20.2	7.33	11/05/2018	18:22
13	J5827-10	1	Solid	20.03	20	20.1	7.09	11/05/2018	18:24
14	J5827-20	1	Solid	20.03	20	20.1	9.12	11/05/2018	18:26
15	J5827-30	1	Solid	20.04	20	20.0	8.15	11/05/2018	18:28
16	CCV2	1	Water			19.3	12.00	11/05/2018	18:30
17	J5827-40	1	Solid	20.02	20	19.9	10.85	11/05/2018	18:32
18	J5827-50	1	Solid	20.03	20	20.1	8.68	11/05/2018	18:34
19	J5827-50DUP	1	Solid	20.04	20	20.2	8.69	11/05/2018	18:36
20	CCV3	1	Water			19.4	2.00	11/05/2018	18:38

WORKLIST(Hardcopy Internal Chain)

1299052

WorkList Name : corrosivity110518-1

WorkList ID : 118884

Date : 11/5/2018 4:17:59 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/05/2018	Solid	J5765-02	Corrosivity	Cool 4 deg C	PSEG05	J22	NB-08-110518	11/05/2018	9045D
11/05/2018	Solid	J5772-02	Corrosivity	Cool 4 deg C	PSEG05	J22	PL-05-110518	11/05/2018	9045D
11/02/2018	Solid	J5819-09	Corrosivity	Cool 4 deg C	PARS02	I41	BES48_WC	11/02/2018	9045D
11/05/2018	Solid	J5826-04	Corrosivity	Cool 4 deg C	PSEG01	J13	SB-1	11/05/2018	9045D
11/05/2018	Solid	J5826-08	Corrosivity	Cool 4 deg C	PSEG01	J13	SB-2	11/05/2018	9045D
11/05/2018	Solid	J5826-12	Corrosivity	Cool 4 deg C	PSEG01	J13	SB-S	11/05/2018	9045D
11/05/2018	Solid	J5827-10	Corrosivity	Cool 4 deg C	PSEG01	J33	SP-A	11/05/2018	9045D
11/05/2018	Solid	J5827-20	Corrosivity	Cool 4 deg C	PSEG01	J33	SP-B	11/05/2018	9045D
11/05/2018	Solid	J5827-30	Corrosivity	Cool 4 deg C	PSEG01	J33	SP-C	11/05/2018	9045D
11/05/2018	Solid	J5827-40	Corrosivity	Cool 4 deg C	PSEG01	J33	SP-D	11/05/2018	9045D
11/05/2018	Solid	J5827-50	Corrosivity	Cool 4 deg C	PSEG01	J33	SP-E	11/05/2018	9045D

Date/Time 11/05/18 5:50
 Received by: SP
 Relinquished by: CP

Date/Time 11/05/18 6:40
 Received by: CP
 Relinquished by: SP