



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB99061

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP68911
COD calibration std. 100 ppm	WP68908
COD calibration std. 10 ppm	WP68907
COD calibration std. 150 ppm	WP68910
COD calibration std. 50 ppm	WP68909
COD calibration std. 0 ppm	WP68906
COD Digestion Vials Low Level 0-150Mg/L	W2421
COD ICV-LCS std, 50ppm	WP69362
COD CCV std, 50ppm	WP69361

Temp In (C): 150	Date In: 11/06/2018	Time In: 11:40
Temp Out (C): 150	Date Out: 11/06/2018	Time Out: 13:40

Intercept: -0.7935

Slope: 1.0096

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	10/20/2018	13:05
2	CAL2	10	1	Water	9.000	10/20/2018	13:05
3	CAL3	50	1	Water	49.000	10/20/2018	13:06
4	CAL4	100	1	Water	100.000	10/20/2018	13:06
5	CAL5	150	1	Water	151.000	10/20/2018	13:07



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB99061

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	49.000	49.320	10/20/2018	13:07
2	ICB				1	Water	0.000	0.786	10/20/2018	13:08
3	CCV1	50			1	Water	50.000	50.311	11/06/2018	14:30
4	CCB1				1	Water	0.000	0.786	11/06/2018	14:30
5	LB99061BL				1	Water	0.000	0.786	11/06/2018	14:31
6	LB99061BS	50			1	Water	49.000	49.320	11/06/2018	14:31
7	J5778-06				1	Water	0.000	0.786	11/06/2018	14:32
8	J5778-06DUP				1	Water	0.000	0.786	11/06/2018	14:32
9	J5778-06MS	50			1	Water	50.000	50.311	11/06/2018	14:33
10	J5778-06MSD	50			1	Water	50.000	50.311	11/06/2018	14:33
11	J5804-03				1	Water	9.000	9.700	11/06/2018	14:34
12	J5825-01				100	Water	73.000	73.092	11/06/2018	14:34
13	J5825-02				100	Water	68.000	68.139	11/06/2018	14:35
14	CCV2	50			1	Water	50.000	50.311	11/06/2018	14:35
15	CCB2				1	Water	0.000	0.786	11/06/2018	14:36

1899641

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD110618

WorkList ID : 118915

Date : 11/6/2018 11:06:06 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/07/2018	Water	J5778-06	COD	Conc H2SO4 4c	PSEG01	I23	OILY-WATER	11/01/2018	SM5220 D
11/07/2018	Water	J5804-03	COD	Conc H2SO4 4c	PSEG01	I21	BS-DRUM-3	11/02/2018	SM5220 D
11/19/2018	Water	J5825-01	COD	Conc H2SO4 4c	TRIS02	J13	DSN001-SEWER-OUTLET-00	11/05/2018	SM5220 D
11/19/2018	Water	J5825-02	COD	Conc H2SO4 4c	TRIS02	J13	DSN002-SEWER-OUTLET-00	11/05/2018	SM5220 D

Date/Time 11/06/18 11:10
 Received by: IR
 Relinquished by: IR

Date/Time 11/06/18 12:00
 Received by: IR
 Relinquished by: IR