



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB99062

Slope : 99.1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.6	4.01	11/06/2018	14:50
2	CAL2	1	Water			20.7	7.00	11/06/2018	14:52
3	CAL3	1	Water			20.6	10.04	11/06/2018	14:54
4	ICV	1	Water			20.9	7.01	11/06/2018	14:56
5	CCV1	1	Water			20.5	2.00	11/06/2018	14:58
6	J5836-01	1	Water			19.8	5.14	11/06/2018	15:00
7	J5836-01DUP	1	Water			19.9	5.15	11/06/2018	15:02
8	CCV2	1	Water			20.6	12.00	11/06/2018	15:04

WORKLIST(Hardcopy Internal Chain)

UB 99062

WorkList Name : pH110618-1

WorkList ID : 118926

Date : 11/6/2018 2:48:13 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/05/2018	Water	J5836-01 A	pH	Cool 4 deg C	STER11	J11	POOLWATER	11/05/2018	SM 4500-H B

Date/Time 11/06/18 2:40
 Received by: SP
 Relinquished by: CP

Date/Time 11/06/18 3:10
 Received by: CP
 Relinquished by: SP