

LB 99063


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36468



A= 0.0000e+000

B= 2.7308e-004

C= -1.1797e-002

Rho= 0.9999556

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	25	0.000	25				
0.05	0.050	0.047	-0.003	216	0.0 %	216				
0.2	0.200	0.204	0.004	789	0.0 %	789				
2.5	2.500	2.478	-0.022	9116	0.0 %	9116				
5.0	5.000	5.008	0.008	18381	0.0 %	18381				
7.5	7.500	7.572	0.072	27770	0.0 %	27770				
10.0	10.000	9.947	-0.053	36468	0.0 %	36468				

LB99063 INSTRUMENT ID : CV1

Method: 245.1		Operator: Admin		Date of Analysis: 06 Nov 2018 15:32:55				
Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	25	-	0.0000245.1	PPB	06 Nov 2018 15:36:16	S	Std
0.05 - 1	50.2	216	-	0.0500245.1	PPB	06 Nov 2018 15:38:34	S	Std
0.2 - 1	52.5	789	-	0.2000245.1	PPB	06 Nov 2018 15:40:53	S	Std
2.5 - 1	55	9116	-	2.5000245.1	PPB	06 Nov 2018 15:43:11	S	Std
5.0 - 1	57.5	18381	-	5.0000245.1	PPB	06 Nov 2018 15:45:29	S	Std
7.5 - 1	510	27770	-	7.5000245.1	PPB	06 Nov 2018 15:47:45	S	Std
10.0 - 1		36468	-	10.0000245.1	PPB	06 Nov 2018 15:50:02	S	Std
ICV60 - 1	ICV60	14707	4.0044	-245.1	PPB	06 Nov 2018 15:56:26	U	SMPL
ICB60 - 1	ICB60	-56	-0.0271	-245.1	PPB	06 Nov 2018 15:58:43	U	SMPL
CCV12 - 1	CCV12	18472	5.0326	-245.1	PPB	06 Nov 2018 16:01:00	U	SMPL
CCB12 - 1	CCB12	-24	-0.0184	-245.1	PPB	06 Nov 2018 16:03:16	U	SMPL
CRA - 1	CRA	198	0.0423	-245.1	PPB	06 Nov 2018 16:13:30	U	SMPL
HighStd - 1	HighStd	35983	9.8146	-245.1	PPB	06 Nov 2018 16:15:45	U	SMPL
ChkStd - 1	ChkStd	26218	7.1479	-245.1	PPB	06 Nov 2018 16:18:01	U	SMPL
PB114534BL - 1	PBW	-121	-0.0448	-245.1	PPB	06 Nov 2018 16:20:19	U	SMPL
PB114534BS - 1	LCSW	13703	3.7303	-245.1	PPB	06 Nov 2018 16:22:36	U	SMPL
J5808-01 - 1	110118-SIDI-E-D-B	-81	-0.0339	-245.1	PPB	06 Nov 2018 16:24:54	U	SMPL
J5808-02 - 1	110118-SIDI-E-D-M	-56	-0.0271	-245.1	PPB	06 Nov 2018 16:27:11	U	SMPL
J5808-03 - 1	110118-SIDI-E-D-S	-70	-0.0309	-245.1	PPB	06 Nov 2018 16:29:29	U	SMPL
J5808-04 - 1	110118-DBRI-E-D-B	-48	-0.0249	-245.1	PPB	06 Nov 2018 16:31:47	U	SMPL
J5808-05 - 1	110118-DBRI-E-D-M	-32	-0.0205	-245.1	PPB	06 Nov 2018 16:34:03	U	SMPL
CCV13 - 1	CCV13	18134	4.9403	-245.1	PPB	06 Nov 2018 16:36:18	U	SMPL
CCB13 - 1	CCB13	-67	-0.0301	-245.1	PPB	06 Nov 2018 16:38:35	U	SMPL
J5808-06 - 1	110118-DBRI-E-D-S	-32	-0.0205	-245.1	PPB	06 Nov 2018 16:40:51	U	SMPL
J5808-07 - 1	110118-DBRI-E-U-B	5	-0.0104	-245.1	PPB	06 Nov 2018 16:43:07	U	SMPL
J5808-08 - 1	110118-DBRI-E-U-M	170	0.0346	-245.1	PPB	06 Nov 2018 16:45:23	U	SMPL
J5808-10 - 1	110118-TIPC-E-D-B	212	0.0461	-245.1	PPB	06 Nov 2018 16:49:56	U	SMPL
J5808-11 - 1	110118-TIPC-E-D-S	-19	-0.0170	-245.1	PPB	06 Nov 2018 16:52:13	U	SMPL
J5808-12 - 1	110118-TIPC-E-U-B	-42	-0.0233	-245.1	PPB	06 Nov 2018 16:54:30	U	SMPL
J5808-13 - 1	110118-TIPC-E-U-S	13	-0.0082	-245.1	PPB	06 Nov 2018 16:56:47	U	SMPL
J5818-01 - 1	EFF-WW	15	-0.0077	-245.1	PPB	06 Nov 2018 16:59:05	U	SMPL
J5818-01DUP - 1	EFF-WWDUP	32	-0.0031	-245.1	PPB	06 Nov 2018 17:01:21	U	SMPL
CCV14 - 1	CCV14	18333	4.9946	-245.1	PPB	06 Nov 2018 17:03:38	U	SMPL
CCB14 - 1	CCB14	-45	-0.0241	-245.1	PPB	06 Nov 2018 17:05:54	U	SMPL
J5818-01MS - 1	EFF-WWMS	12046	3.2778	-245.1	PPB	06 Nov 2018 17:08:10	U	SMPL
J5818-01MSD - 1	EFF-WWMSD	12071	3.2846	-245.1	PPB	06 Nov 2018 17:13:32	U	SMPL
PB114535BL - 1	PBW	-27	-0.0192	-245.1	PPB	06 Nov 2018 17:15:48	U	SMPL
PB114535BS - 1	LCSW	13336	3.6300	-245.1	PPB	06 Nov 2018 17:18:04	U	SMPL
J5810-01 - 1	110218-DBRI-E-D-B	-94	-0.0375	-245.1	PPB	06 Nov 2018 17:20:21	U	SMPL
J5810-02 - 1	110218-DBRI-E-D-M	-30	-0.0200	-245.1	PPB	06 Nov 2018 17:22:38	U	SMPL
J5810-03 - 1	110218-DBRI-E-D-S	-61	-0.0285	-245.1	PPB	06 Nov 2018 17:24:55	U	SMPL
J5810-04 - 1	110218-SIDI-E-D-B	5	-0.0104	-245.1	PPB	06 Nov 2018 17:27:13	U	SMPL
J5810-05 - 1	110218-SIDI-E-D-M	-42	-0.0233	-245.1	PPB	06 Nov 2018 17:29:30	U	SMPL
J5810-06 - 1	110218-SIDI-E-D-S	-8	-0.0140	-245.1	PPB	06 Nov 2018 17:31:47	U	SMPL
CCV15 - 1	CCV15	18118	4.9359	-245.1	PPB	06 Nov 2018 17:34:03	U	SMPL
CCB15 - 1	CCB15	-82	-0.0342	-245.1	PPB	06 Nov 2018 17:36:19	U	SMPL
J5810-07 - 1	110218-SIDI-E-U-B	-61	-0.0285	-245.1	PPB	06 Nov 2018 17:38:36	U	SMPL
J5810-08 - 1	110218-SIDI-E-U-M	-8	-0.0140	-245.1	PPB	06 Nov 2018 17:40:52	U	SMPL
J5810-09 - 1	110218-SIDI-E-U-S	-53	-0.0263	-245.1	PPB	06 Nov 2018 17:43:09	U	SMPL
J5810-10 - 1	110218-TIPC-E-D-B	-23	-0.0181	-245.1	PPB	06 Nov 2018 17:45:25	U	SMPL
J5810-11 - 1	110218-TIPC-E-D-S	-71	-0.0312	-245.1	PPB	06 Nov 2018 17:47:42	U	SMPL
J5810-12 - 1	110218-TIPC-E-U-B	-89	-0.0361	-245.1	PPB	06 Nov 2018 17:49:59	U	SMPL
J5810-13 - 1	110218-TIPC-E-U-S	-70	-0.0309	-245.1	PPB	06 Nov 2018 17:52:16	U	SMPL
J5810-13DUP - 1	110218-TIPC-E-U-SDUP	-79	-0.0334	-245.1	PPB	06 Nov 2018 17:54:34	U	SMPL
J5810-13MS - 1	110218-TIPC-E-U-SMS	13426	3.6546	-245.1	PPB	06 Nov 2018 18:00:54	U	SMPL
J5810-13MSD - 1	110218-TIPC-E-U-SMSD	13319	3.6254	-245.1	PPB	06 Nov 2018 18:03:10	U	SMPL
CCV16 - 1	CCV16	18078	4.9250	-245.1	PPB	06 Nov 2018 18:05:27	U	SMPL
CCB16 - 1	CCB16	-132	-0.0478	-245.1	PPB	06 Nov 2018 18:07:44	U	SMPL
J5808-09 - 1	110118-DBRI-E-U-S	140	0.0264	-245.1	PPB	06 Nov 2018 18:12:39	U	SMPL
J5818-01LX5 - 1		-23	-0.0181	-245.1	PPB	06 Nov 2018 18:14:56	U	SMPL
J5818-01A - 1		14155	3.8537	-245.1	PPB	06 Nov 2018 18:17:12	U	SMPL
J5810-13LX5 - 1		-94	-0.0375	-245.1	PPB	06 Nov 2018 18:19:28	U	SMPL
J5810-13A - 1		17350	4.7262	-245.1	PPB	06 Nov 2018 18:21:46	U	SMPL
CCV17 - 1	CCV17	18260	4.9747	-245.1	PPB	06 Nov 2018 18:24:02	U	SMPL
CCB17 - 1	CCB17	-120	-0.0446	-245.1	PPB	06 Nov 2018 18:26:20	U	SMPL