


7471B

LB99068

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.9203e-004

C= 4.4582e-002

Rho= 0.9999355

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	-73	0.000	-73				
0.2	0.200	0.204	0.004	832	0.0 %	832				
2.5	2.500	2.510	0.010	12837	0.0 %	12837				
5.0	5.000	4.942	-0.058	25504	0.0 %	25504				
7.5	7.500	7.452	-0.048	38572	0.0 %	38572				
10.0	10.000	10.062	0.062	52164	0.0 %	52164				

LB99068 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Nov 2018 18:38:44

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	-73	-	0.00007471B	PPB		06 Nov 2018 18:41:42	S	Std
0.2 - 1	50.2	832	-	0.20007471B	PPB		06 Nov 2018 18:44:01	S	Std
2.5 - 1	52.5	12837	-	2.50007471B	PPB		06 Nov 2018 18:46:19	S	Std
5.0 - 1	55	25504	-	5.00007471B	PPB		06 Nov 2018 18:48:37	S	Std
7.5 - 1	57.5	38572	-	7.50007471B	PPB		06 Nov 2018 18:50:54	S	Std
10.0 - 1	51.0	52164	-	10.00007471B	PPB		06 Nov 2018 18:53:12	S	Std
ICV61 - 1	ICV61	20358	3.9539	-7471B	PPB		06 Nov 2018 18:55:55	U	SMPL
ICB61 - 1	ICB61	-123	0.0210	-7471B	PPB		06 Nov 2018 18:58:11	U	SMPL
CCV18 - 1	CCV18	25494	4.9402	-7471B	PPB		06 Nov 2018 19:00:27	U	SMPL
CCB18 - 1	CCB18	-185	0.0091	-7471B	PPB		06 Nov 2018 19:02:43	U	SMPL
CRA - 1	CRA	877	0.2130	-7471B	PPB		06 Nov 2018 19:05:00	U	SMPL
HighStd - 1	HighStd	52334	10.0943	-7471B	PPB		06 Nov 2018 19:07:16	U	SMPL
ChkStd - 1	ChkStd	36208	6.9976	-7471B	PPB		06 Nov 2018 19:09:32	U	SMPL
PB114538BL - 1	PBS	-258	-0.0050	-7471B	PPB		06 Nov 2018 19:11:50	U	SMPL
PB114538BS - 1	LCSS	20565	3.9937	-7471B	PPB		06 Nov 2018 19:16:41	U	SMPL
J5765-01 - 1	NB-08-110518	1708	0.3726	-7471B	PPB		06 Nov 2018 19:18:57	U	SMPL
J5772-01 - 1	PL-05-110518	961	0.2291	-7471B	PPB		06 Nov 2018 19:21:16	U	SMPL
J5819-01 - 1	BES48_SB04_0.75-1.25	1708	0.3726	-7471B	PPB		06 Nov 2018 19:23:34	U	SMPL
J5819-02 - 1	BES48_SB04_2.25-2.75	334	0.1087	-7471B	PPB		06 Nov 2018 19:25:52	U	SMPL
J5819-03 - 1	BES48_SB01_1-1.5	2814	0.5850	-7471B	PPB		06 Nov 2018 19:28:08	U	SMPL
CCV19 - 1	CCV19	25801	4.9992	-7471B	PPB		06 Nov 2018 19:30:24	U	SMPL
CCB19 - 1	CCB19	-199	0.0064	-7471B	PPB		06 Nov 2018 19:32:40	U	SMPL
J5819-04 - 1	BES48_SB01_4-4.5	353	0.1124	-7471B	PPB		06 Nov 2018 19:34:56	U	SMPL
J5819-05 - 1	BES48_SB02_1.5-2	2191	0.4653	-7471B	PPB		06 Nov 2018 19:37:13	U	SMPL
J5819-06 - 1	BES48_SB02_3.5-4	3224	0.6637	-7471B	PPB		06 Nov 2018 19:39:29	U	SMPL
J5819-07 - 1	BES48_SB03_1.5-2	237	0.0901	-7471B	PPB		06 Nov 2018 19:41:45	U	SMPL
J5819-08 - 1	BES48_SB03_3.5-4.0	400	0.1214	-7471B	PPB		06 Nov 2018 19:44:02	U	SMPL
J5826-03 - 1	SB-1	240	0.0907	-7471B	PPB		06 Nov 2018 19:46:19	U	SMPL
J5826-03DUP - 1	SB-1DUP	172	0.0776	-7471B	PPB		06 Nov 2018 19:48:36	U	SMPL
J5826-03MS - 1	SB-1MS	19851	3.8566	-7471B	PPB		06 Nov 2018 19:53:29	U	SMPL
J5826-03MSD - 1	SB-1MSD	19785	3.8439	-7471B	PPB		06 Nov 2018 19:55:46	U	SMPL
J5826-07 - 1	SB-2	267	0.0959	-7471B	PPB		06 Nov 2018 19:58:03	U	SMPL
CCV20 - 1	CCV20	26298	5.0946	-7471B	PPB		06 Nov 2018 20:00:21	U	SMPL
CCB20 - 1	CCB20	-223	0.0018	-7471B	PPB		06 Nov 2018 20:02:37	U	SMPL
J5826-11 - 1	SB-S	495	0.1396	-7471B	PPB		06 Nov 2018 20:04:55	U	SMPL
J5827-01 - 1	SP-A	366	0.1149	-7471B	PPB		06 Nov 2018 20:07:11	U	SMPL
J5827-11 - 1	SP-B	1567	0.3455	-7471B	PPB		06 Nov 2018 20:09:28	U	SMPL
J5827-21 - 1	SP-C	5285	1.0595	-7471B	PPB		06 Nov 2018 20:11:44	U	SMPL
J5827-31 - 1	SP-D	1452	0.3234	-7471B	PPB		06 Nov 2018 20:14:01	U	SMPL
J5827-41 - 1	SP-E	1236	0.2819	-7471B	PPB		06 Nov 2018 20:16:18	U	SMPL
J5826-03LX5 - 1		-83	0.0286	-7471B	PPB		06 Nov 2018 20:18:36	U	SMPL
J5826-03A - 1		22562	4.3772	-7471B	PPB		06 Nov 2018 20:20:53	U	SMPL
CCV21 - 1	CCV21	25970	5.0316	-7471B	PPB		06 Nov 2018 20:23:12	U	SMPL
CCB21 - 1	CCB21	-171	0.0117	-7471B	PPB		06 Nov 2018 20:25:29	U	SMPL