

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 11/07/2018

TEMP1 IN: 103 11/07/2018 07:20 TEMP OUT: 103 11/07/2018 08:20

TEMP2 IN: 103 11/07/2018 09:20 TEMP OUT: 103 11/07/2018 10:00

TEMP3 IN: 103 11/07/2018 11:25 TEMP OUT: 104 11/07/2018 12:50

TEMP4 IN: 104 11/07/2018 13:30 TEMP OUT: 104 11/07/2018 15:15

Run Number: LB99069

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99069BL	LB99069BL	2.6997	2.6997	100	2.6998	2.6998	2.6998	0.0001	1
2	LB99069BS	LB99069BS	2.1444	2.1444	100	2.1991	2.1991	2.1991	0.0547	547
3	J5825-01	DSN001-SEWER-OUTLET-001	1.4015	1.4015	150	1.4393	1.4393	1.4393	0.0378	252
4	J5825-02	DSN002-SEWER-OUTLET-002	1.4048	1.4048	50	1.4178	1.4178	1.4178	0.0130	260
5	J5825-02DUP	DSN002-SEWER-OUTLET-002DUP	1.4241	1.4241	50	1.4368	1.4368	1.4368	0.0127	254
6	J5829-01	110518-SIDI-F-D-B	1.4057	1.4057	200	1.4336	1.4336	1.4336	0.0279	139.5
7	J5829-02	110518-SIDI-F-D-M	1.4015	1.4015	200	1.4205	1.4205	1.4205	0.0190	95
8	J5829-03	110518-SIDI-F-D-S	1.4172	1.4172	300	1.4389	1.4389	1.4389	0.0217	72.3
9	J5829-04	110518-SIDI-F-U-B	1.4055	1.4055	300	1.4293	1.4293	1.4293	0.0238	79.3
10	J5829-05	110518-SIDI-F-U-M	1.4025	1.4025	300	1.4224	1.4224	1.4224	0.0199	66.3
11	J5829-06	110518-SIDI-F-U-S	1.4008	1.4008	300	1.4171	1.4171	1.4171	0.0163	54.3
12	J5829-07	110518-TIPC-E-D-M	1.3958	1.3958	400	1.4113	1.4113	1.4113	0.0155	38.8
13	J5829-08	110518-TIPC-E-U-M	1.3929	1.3929	400	1.4052	1.4052	1.4052	0.0123	30.8
14	J5829-09	110518-DBRI-E-D-B	1.4050	1.4050	400	1.4327	1.4327	1.4327	0.0277	69.3
15	J5829-10	110518-DBRI-E-D-M	1.3968	1.3968	400	1.4245	1.4245	1.4245	0.0277	69.3

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 11/07/2018

TEMP1 IN: 103 11/07/2018 07:20 TEMP OUT: 103 11/07/2018 08:20

TEMP2 IN: 103 11/07/2018 09:20 TEMP OUT: 103 11/07/2018 10:00

TEMP3 IN: 103 11/07/2018 11:25 TEMP OUT: 104 11/07/2018 12:50

TEMP4 IN: 104 11/07/2018 13:30 TEMP OUT: 104 11/07/2018 15:15

Run Number: LB99069

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J5829-11	110518-DBRI-E-D-S	1.4021	1.4021	400	1.4278	1.4278	1.4278	0.0257	64.3
17	J5829-12	110518-DBRI-E-U-B	1.3925	1.3925	400	1.4317	1.4317	1.4317	0.0392	98
18	J5829-13	110518-DBRI-E-U-M	1.4019	1.4019	400	1.4302	1.4302	1.4302	0.0283	70.8
19	J5829-14	110518-DBRI-E-U-S	1.4069	1.4069	400	1.4387	1.4387	1.4387	0.0318	79.5
20	J5832-02	VB16089	1.4094	1.4094	50	1.4270	1.4270	1.4270	0.0176	352

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

2399069

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS J5829

WorkList ID : 118937

Date : 11/7/2018 7:31:28 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/12/2018	Water	J5825-01 -L	TSS	Cool 4 deg C	TRIS02	J13	DSN001-SEWER-OUTLET-00	11/05/2018	SM2540 D
11/12/2018	Water	J5825-02 -L	TSS	Cool 4 deg C	TRIS02	J13	DSN002-SEWER-OUTLET-00	11/05/2018	SM2540 D
11/12/2018	Water	J5829-01 -L	TSS	Cool 4 deg C	HDR103	J12	110518-SIDI-F-D-B	11/05/2018	SM2540 D
11/12/2018	Water	J5829-02 -L	TSS	Cool 4 deg C	HDR103	J12	110518-SIDI-F-D-M	11/05/2018	SM2540 D
11/12/2018	Water	J5829-03 -L	TSS	Cool 4 deg C	HDR103	J12	110518-SIDI-F-D-S	11/05/2018	SM2540 D
11/12/2018	Water	J5829-04 -L	TSS	Cool 4 deg C	HDR103	J12	110518-SIDI-F-U-B	11/05/2018	SM2540 D
11/12/2018	Water	J5829-05 -L	TSS	Cool 4 deg C	HDR103	J12	110518-SIDI-F-U-M	11/05/2018	SM2540 D
11/12/2018	Water	J5829-06 -L	TSS	Cool 4 deg C	HDR103	J12	110518-SIDI-F-U-S	11/05/2018	SM2540 D
11/08/2018	Water	J5829-07 -L	TSS	Cool 4 deg C	HDR103	J12	110518-TIPC-E-D-M	11/05/2018	SM2540 D
11/08/2018	Water	J5829-08 -L	TSS	Cool 4 deg C	HDR103	J12	110518-TIPC-E-U-M	11/05/2018	SM2540 D
11/12/2018	Water	J5829-09 -L	TSS	Cool 4 deg C	HDR103	J12	110518-DBRI-E-D-B	11/05/2018	SM2540 D
11/12/2018	Water	J5829-10 -L	TSS	Cool 4 deg C	HDR103	J12	110518-DBRI-E-D-M	11/05/2018	SM2540 D
11/12/2018	Water	J5829-11 -L	TSS	Cool 4 deg C	HDR103	J12	110518-DBRI-E-D-S	11/05/2018	SM2540 D
11/12/2018	Water	J5829-12 -L	TSS	Cool 4 deg C	HDR103	J12	110518-DBRI-E-U-B	11/05/2018	SM2540 D
11/12/2018	Water	J5829-13 -L	TSS	Cool 4 deg C	HDR103	J12	110518-DBRI-E-U-M	11/05/2018	SM2540 D
11/12/2018	Water	J5829-14 -L	TSS	Cool 4 deg C	HDR103	J12	110518-DBRI-E-U-S	11/05/2018	SM2540 D
11/09/2018	Water	J5832-02 -L	TSS	Cool 4 deg C	PSEG01	I21	VB16089	11/06/2018	SM2540 D

Date/Time 11-07-18 7:32 AM
 Received by: JD
 Relinquished by: JD

Date/Time 11-07-18 11:25 AM
 Received by: CD
 Relinquished by: JD