



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB99070

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.6	4.01	11/06/2018	17:10
2	CAL2	1	Water			20.4	7.01	11/06/2018	17:12
3	CAL3	1	Water			20.3	10.05	11/06/2018	17:14
4	ICV	1	Water			20.4	7.01	11/06/2018	17:16
5	CCV1	1	Water			20.4	2.00	11/06/2018	17:18
6	J5762-02	1	Solid	20.02	20	21.0	10.09	11/06/2018	17:20
7	J5762-02DUP	1	Solid	20.03	20	21.1	10.10	11/06/2018	17:22
8	J5762-04	1	Solid	20.02	20	21.1	5.65	11/06/2018	17:24
9	J5807-02	1	Solid	20.02	20	19.3	7.46	11/06/2018	17:26
10	J5832-01	1	Solid	20.02	20	20.5	5.74	11/06/2018	17:28
11	J5839-04	1	Solid	20.04	20	21.1	6.58	11/06/2018	17:30
12	J5839-08	1	Solid	20.02	20	21.0	6.69	11/06/2018	17:32
13	J5839-12	1	Solid	20.03	20	21.3	8.67	11/06/2018	17:34
14	J5839-16	1	Solid	20.02	20	21.1	7.12	11/06/2018	17:36
15	CCV2	1	Water			20.6	12.00	11/06/2018	17:38

WORKLIST(Hardcopy Internal Chain)

LB 99070

WorkList Name : corrosivity110618

WorkList ID : 118929

Date : 11/6/2018 4:08:26 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/06/2018	Solid	J5762-02	Corrosivity	Cool 4 deg C	PSEG05	I21	BU-01-110518	11/06/2018	9045D
11/06/2018	Solid	J5762-04	Corrosivity	Cool 4 deg C	PSEG05	I21	BU-02-110518	11/06/2018	9045D
11/06/2018	Solid	J5832-01	Corrosivity	Cool 4 deg C	PSEG01	I21	OILY-DEBRIS	11/06/2018	9045D
11/06/2018	Solid	J5839-04	Corrosivity	Cool 4 deg C	PSEG01	I31	SB-4	11/06/2018	9045D
11/06/2018	Solid	J5839-08	Corrosivity	Cool 4 deg C	PSEG01	I31	SB-6	11/06/2018	9045D
11/06/2018	Solid	J5839-12	Corrosivity	Cool 4 deg C	PSEG01	I31	SB-7-S	11/06/2018	9045D
11/06/2018	Solid	J5839-16	Corrosivity	Cool 4 deg C	PSEG01	I31	SB-7-D	11/06/2018	9045D

Date/Time 11/06/18 4:55

Received by: SP

Relinquished by: CP

Date/Time 11/06/18 5:40

Received by: CP

Relinquished by: SP