



Analytical Summary Report

Analysis Method: 9041A

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB99071

Slope : N/A

BalanceID: WC SC-4

pH Meter ID : N/A

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267
3.0-5.5 pH paper refill	W1934
Test Paper, PH, Short Range 6.0/8.0	W2350
Test Paper,PH Short Range 9.0/10.0	W1939
Test Paper, PH, Short Range 6.0/8.0	W2350
0.0-3.0 pH paper refill	W1931
10.0-12.0 pH paper refill	W1940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.0 Control Limits[+/- 0.1].

True Value of CCV2 = 12.0 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.6	4.0	11/06/2018	17:10
2	CAL2	1	Water			20.4	7.0	11/06/2018	17:12
3	CAL3	1	Water			20.3	10.0	11/06/2018	17:14
4	ICV	1	Water			20.4	7.0	11/06/2018	17:16
5	CCV1	1	Water			20.4	2.0	11/06/2018	17:18
6	J5807-24	1	Solid	20.02	20	19.3	7.2	11/06/2018	17:26
7	CCV2	1	Water			20.6	12.0	11/06/2018	17:38