



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB99073

Slope : 99.1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.6	4.01	11/06/2018	17:10
2	CAL2	1	Water			20.4	7.01	11/06/2018	17:12
3	CAL3	1	Water			20.3	10.05	11/06/2018	17:14
4	ICV	1	Water			20.4	7.01	11/06/2018	17:16
5	CCV1	1	Water			20.5	2.00	11/06/2018	17:45
6	J5832-02	1	Water			16.5	7.21	11/06/2018	17:47
7	J5832-02DUP	1	Water			16.6	7.22	11/06/2018	17:49
8	CCV2	1	Water			20.6	12.00	11/06/2018	17:51

WORKLIST(Hardcopy Internal Chain)

VB 99073

WorkList Name : pH110618-2

WorkList ID : 118932

Date : 11/6/2018 4:09:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/06/2018	Water	J5832-02	pH	Cool 4 deg C	PSEG01	I21	VB16089	11/06/2018	9040C

Date/Time 11/06/18 4:55
 Received by: SP
 Relinquished by: CP

Date/Time 11/06/18 6:00
 Received by: CP
 Relinquished by: SP