


LB99083

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.7180e-004

C= 3.1004e-002

Rho= 0.9999035

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.035	0.035	25	0.000	25				
0.2	0.200	0.227	0.027	1141	0.0 %	1141				
2.5	2.500	2.420	-0.080	13907	0.0 %	13907				
5.0	5.000	4.962	-0.038	28700	0.0 %	28700				
7.5	7.500	7.575	0.075	43910	0.0 %	43910				
10.0	10.000	9.981	-0.019	57914	0.0 %	57914				

LB99083 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 07 Nov 2018 16:17:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	25	-	0.0000	7471B	PPB	07 Nov 2018 16:19:46	S	Std
0.2 - 1	50.2	1141	-	0.2000	7471B	PPB	07 Nov 2018 16:22:04	S	Std
2.5 - 1	52.5	13907	-	2.5000	7471B	PPB	07 Nov 2018 16:24:21	S	Std
5.0 - 1	55	28700	-	5.0000	7471B	PPB	07 Nov 2018 16:26:39	S	Std
7.5 - 1	57.5	43910	-	7.5000	7471B	PPB	07 Nov 2018 16:28:56	S	Std
10.0 - 1	57.0	57914	-	10.0000	7471B	PPB	07 Nov 2018 16:31:14	S	Std
ICV63 - 1	ICV63	21717	3.7620	-	7471B	PPB	07 Nov 2018 16:34:07	U	SMPL
ICB63 - 1	ICB63	-119	0.0106	-	7471B	PPB	07 Nov 2018 16:36:23	U	SMPL
CCV29 - 1	CCV29	27557	4.7654	-	7471B	PPB	07 Nov 2018 16:38:40	U	SMPL
CCB29 - 1	CCB29	-124	0.0097	-	7471B	PPB	07 Nov 2018 16:40:56	U	SMPL
CRA - 1	CRA	1100	0.2200	-	7471B	PPB	07 Nov 2018 16:43:12	U	SMPL
HighStd - 1	HighStd	55371	9.5439	-	7471B	PPB	07 Nov 2018 16:45:28	U	SMPL
ChkStd - 1	ChkStd	39804	6.8694	-	7471B	PPB	07 Nov 2018 16:47:45	U	SMPL
PB114589BL - 1	PBS	-178	0.0004	-	7471B	PPB	07 Nov 2018 16:50:03	U	SMPL
PB114589BS - 1	LCSS	20224	3.5055	-	7471B	PPB	07 Nov 2018 16:52:21	U	SMPL
J5762-01 - 1	BU-01-110518	3815	0.6864	-	7471B	PPB	07 Nov 2018 16:54:39	U	SMPL
J5762-03 - 1	BU-02-110518	1135	0.2260	-	7471B	PPB	07 Nov 2018 16:56:57	U	SMPL
J5835-01 - 1	ES-OILY-DEBRI	1546	0.2966	-	7471B	PPB	07 Nov 2018 16:59:14	U	SMPL
J5839-03 - 1	SB-4	512	0.1190	-	7471B	PPB	07 Nov 2018 17:01:32	U	SMPL
J5839-07 - 1	SB-6	1942	0.3646	-	7471B	PPB	07 Nov 2018 17:03:48	U	SMPL
CCV30 - 1	CCV30	27437	4.7448	-	7471B	PPB	07 Nov 2018 17:06:04	U	SMPL
CCB30 - 1	CCB30	-123	0.0099	-	7471B	PPB	07 Nov 2018 17:08:19	U	SMPL
J5839-11 - 1	SB-7-S	3225	0.5851	-	7471B	PPB	07 Nov 2018 17:16:28	U	SMPL
J5839-11DUP - 1	SB-7-SDUP	3246	0.5887	-	7471B	PPB	07 Nov 2018 17:18:43	U	SMPL
J5839-11MS - 1	SB-7-SMS	24579	4.2537	-	7471B	PPB	07 Nov 2018 17:23:46	U	SMPL
J5839-11MSD - 1	SB-7-SMSD	24083	4.1685	-	7471B	PPB	07 Nov 2018 17:26:02	U	SMPL
J5839-15 - 1	SB-7-D	638	0.1406	-	7471B	PPB	07 Nov 2018 17:28:19	U	SMPL
J5839-11LX5 - 1		645	0.1418	-	7471B	PPB	07 Nov 2018 17:30:37	U	SMPL
J5839-11A - 1		24872	4.3041	-	7471B	PPB	07 Nov 2018 17:32:54	U	SMPL
CCV31 - 1	CCV31	27605	4.7736	-	7471B	PPB	07 Nov 2018 17:35:12	U	SMPL
CCB31 - 1	CCB31	-139	0.0071	-	7471B	PPB	07 Nov 2018 17:37:30	U	SMPL