



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99087

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.8	4.01	11/07/2018	17:14
2	CAL2	1	Water			20.6	7.00	11/07/2018	17:16
3	CAL3	1	Water			20.9	10.04	11/07/2018	17:18
4	ICV	1	Water			20.7	7.01	11/07/2018	17:20
5	CCV1	1	Water			20.8	2.00	11/07/2018	17:22
6	J5768-02	1	Solid	20.02	20	21.9	7.79	11/07/2018	17:24
7	J5768-02DUP	1	Solid	20.04	20	22.0	7.80	11/07/2018	17:26
8	J5768-04	1	Solid	20.02	20	21.8	7.67	11/07/2018	17:28
9	J5838-10	1	Solid	20.02	20	22.1	8.33	11/07/2018	17:30
10	J5838-20	1	Solid	20.03	20	21.9	8.08	11/07/2018	17:33
11	J5858-04	1	Solid	20.02	20	21.5	7.41	11/07/2018	17:34
12	J5858-08	1	Solid	20.03	20	21.3	6.51	11/07/2018	17:36
13	J5858-12	1	Solid	20.04	20	21.4	10.92	11/07/2018	17:38
14	J5859-03	1	Solid	20.02	20	21.6	8.33	11/07/2018	17:40
15	J5859-06	1	Solid	20.03	20	21.5	7.94	11/07/2018	17:42
16	CCV2	1	Water			20.8	12.00	11/07/2018	17:52

WORKLIST(Hardcopy Internal Chain)

LB99087

WorkList Name : corrosivity110718

WorkList ID : 118970

Date : 11/7/2018 4:44:47 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/07/2018	Solid	J5768-02	Corrosivity	Cool 4 deg C	PSEG05	I21	OK-02-111718-A	11/07/2018	9045D
11/07/2018	Solid	J5768-04	Corrosivity	Cool 4 deg C	PSEG05	I21	OK-02-111718-B	11/07/2018	9045D
11/07/2018	Solid	J5838-10	Corrosivity	Cool 4 deg C	PSEG01	I31	HC-SOIL-1	11/07/2018	9045D
11/07/2018	Solid	J5838-20	Corrosivity	Cool 4 deg C	PSEG01	I31	HC-SOIL-2	11/07/2018	9045D
11/07/2018	Solid	J5858-04	Corrosivity	Cool 4 deg C	PSEG01	I31	SB-8-S	11/07/2018	9045D
11/07/2018	Solid	J5858-08	Corrosivity	Cool 4 deg C	PSEG01	I31	SB-8-D	11/07/2018	9045D
11/07/2018	Solid	J5858-12	Corrosivity	Cool 4 deg C	PSEG01	I31	SB-9-S	11/07/2018	9045D
11/06/2018	Solid	J5859-03	Corrosivity	Cool 4 deg C	WSP001	J61	B-23(0-10)	11/06/2018	9045D
11/05/2018	Solid	J5859-06	Corrosivity	Cool 4 deg C	WSP001	J61	B-24(0-10)	11/05/2018	9045D

Date/Time 11/07/18 5:05
 Received by: SP
 Relinquished by: CP

Date/Time 11/07/18 6:02
 Received by: CP
 Relinquished by: SP