



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99088

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.8	4.01	11/07/2018	17:14
2	CAL2	1	Water			20.6	7.00	11/07/2018	17:16
3	CAL3	1	Water			20.9	10.04	11/07/2018	17:18
4	ICV	1	Water			20.7	7.01	11/07/2018	17:20
5	CCV1	1	Water			20.8	2.00	11/07/2018	17:22
6	J5844-02	1	Water			19.8	6.52	11/07/2018	17:44
7	J5844-02DUP	1	Water			19.9	6.53	11/07/2018	17:46
8	J5844-05	1	Water			19.9	6.56	11/07/2018	17:48
9	J5858-14	1	Water			18.1	5.87	11/07/2018	17:50
10	CCV2	1	Water			20.8	12.00	11/07/2018	17:52

WORKLIST(Hardcopy Internal Chain)

LB99088

WorkList Name : corrosivity110718-1

WorkList ID : 118971

Date : 11/7/2018 4:44:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/07/2018	Water	J5844-02	Corrosivity	Cool 4 deg C	PSEG01	I31	MH-A-B	11/07/2018	9040C
11/07/2018	Water	J5844-05	Corrosivity	Cool 4 deg C	PSEG01	I31	MH-C	11/07/2018	9040C
11/07/2018	Water	J5858-14	Corrosivity	Cool 4 deg C	PSEG01	I31	SB-9-GW	11/07/2018	9040C

Date/Time 11/07/18 5:05
 Received by: SP
 Relinquished by: SP

Date/Time 11/07/18 6:00
 Received by: CP
 Relinquished by: SP