



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 11/08/2018

TEMP1 IN: 103 11/08/2018 07:15 TEMP OUT: 103 11/08/2018 08:15

TEMP2 IN: 103 11/08/2018 08:35 TEMP OUT: 103 11/08/2018 09:35

TEMP3 IN: 103 11/08/2018 11:10 TEMP OUT: 104 11/08/2018 12:30

TEMP4 IN: 104 11/08/2018 13:00 TEMP OUT: 105 11/08/2018 14:35

Run Number: LB99091

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99091BL	LB99091BL	2.5133	2.5133	100	2.5134	2.5134	2.5134	0.0001	1
2	LB99091BS	LB99091BS	2.7411	2.7411	100	2.7956	2.7956	2.7956	0.0545	545
3	J5844-01	MH-A-B	1.4058	1.4058	500	1.4264	1.4264	1.4264	0.0206	41.2
4	J5844-01DUP	MH-A-BDUP	1.4061	1.4061	500	1.4276	1.4276	1.4276	0.0215	43
5	J5844-04	MH-C	1.4013	1.4013	800	1.4279	1.4279	1.4279	0.0266	33.2
6	J5845-02	OUTFALL001-181106	1.4055	1.4055	900	1.4062	1.4062	1.4062	0.0007	0.8
7	J5851-01	110618-DBRI-F-D-B	1.3911	1.3911	300	1.4151	1.4151	1.4151	0.0240	80
8	J5851-02	110618-DBRI-F-D-M	1.4022	1.4022	300	1.4236	1.4236	1.4236	0.0214	71.3
9	J5851-03	110618-DBRI-F-D-S	1.3803	1.3803	300	1.4009	1.4009	1.4009	0.0206	68.7
10	J5851-04	110618-DBRI-F-U-B	1.4088	1.4088	250	1.4557	1.4557	1.4557	0.0469	187.6
11	J5851-05	110618-DBRI-F-U-M	1.3995	1.3995	300	1.4210	1.4210	1.4210	0.0215	71.7
12	J5851-06	110618-DBRI-F-U-S	1.3972	1.3972	400	1.4172	1.4172	1.4172	0.0200	50
13	J5851-07	110618-JETT-F-D-B	1.4058	1.4058	250	1.4424	1.4424	1.4424	0.0366	146.4
14	J5851-08	110618-JETT-F-D-M	1.4010	1.4010	250	1.4255	1.4255	1.4255	0.0245	98
15	J5851-09	110618-JETT-F-D-S	1.4018	1.4018	250	1.4229	1.4229	1.4229	0.0211	84.4

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: jignesh

Date: 11/08/2018

TEMP1 IN: 103 11/08/2018 07:15 TEMP OUT: 103 11/08/2018 08:15

TEMP2 IN: 103 11/08/2018 08:35 TEMP OUT: 103 11/08/2018 09:35

TEMP3 IN: 103 11/08/2018 11:10 TEMP OUT: 104 11/08/2018 12:30

TEMP4 IN: 104 11/08/2018 13:00 TEMP OUT: 105 11/08/2018 14:35

Run Number: LB99091

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J5851-10	110618-SIDI-F-D-B	1.3941	1.3941	300	1.4238	1.4238	1.4238	0.0297	99
17	J5851-11	110618-SIDI-F-D-M	1.3969	1.3969	250	1.4180	1.4180	1.4180	0.0211	84.4
18	J5851-12	110618-SIDI-F-D-S	1.3993	1.3993	300	1.4242	1.4242	1.4242	0.0249	83
19	J5851-13	110618-SIDI-F-U-B	1.3988	1.3988	400	1.4183	1.4183	1.4183	0.0195	48.8
20	J5851-14	110618-SIDI-F-U-M	1.3919	1.3919	400	1.4103	1.4103	1.4103	0.0184	46
21	J5851-15	110618-SIDI-F-U-S	1.4013	1.4013	400	1.4226	1.4226	1.4226	0.0213	53.3
22	J5851-16	110618-TIPC-E-D-M	1.3961	1.3961	400	1.4103	1.4103	1.4103	0.0142	35.5
23	J5851-17	110618-TIPC-E-U-M	1.3929	1.3929	400	1.4067	1.4067	1.4067	0.0138	34.5
24	J5855-01	COMP	1.4045	1.4045	250	1.4128	1.4128	1.4128	0.0083	33.2
25	J5858-13	SB-9-GW	1.3989	1.3989	250	1.4254	1.4254	1.4254	0.0265	106
26	J5858-13DUP	SB-9-GWDUP	1.3975	1.3975	250	1.4242	1.4242	1.4242	0.0267	106.8

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

$$\text{Weight (g)} = C - B$$

$$\text{Result mg/L} = \frac{D}{A} * 1000 * 1000$$

1399091

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tss j5855

WorkList ID : 118980

Date : 11/8/2018 7:34:48 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/12/2018	Water	J5844-01-F	TSS	Cool 4 deg C	PSEG01	I31	MH-A-B	11/07/2018	SM2540 D
11/12/2018	Water	J5844-04-F	TSS	Cool 4 deg C	PSEG01	I31	MH-C	11/07/2018	SM2540 D
11/13/2018	Water	J5845-02-E	TSS	Cool 4 deg C	SHAW22	F21	OUTFALL001-181106	11/06/2018	SM2540 D
11/13/2018	Water	J5851-01-C	TSS	Cool 4 deg C	HDR103	J23	110618-DBRI-F-D-B	11/06/2018	SM2540 D
11/13/2018	Water	J5851-02-C	TSS	Cool 4 deg C	HDR103	J23	110618-DBRI-F-D-M	11/06/2018	SM2540 D
11/13/2018	Water	J5851-03-C	TSS	Cool 4 deg C	HDR103	J23	110618-DBRI-F-D-S	11/06/2018	SM2540 D
11/13/2018	Water	J5851-04-C	TSS	Cool 4 deg C	HDR103	J23	110618-DBRI-F-U-B	11/06/2018	SM2540 D
11/13/2018	Water	J5851-05-C	TSS	Cool 4 deg C	HDR103	J23	110618-DBRI-F-U-M	11/06/2018	SM2540 D
11/13/2018	Water	J5851-06-C	TSS	Cool 4 deg C	HDR103	J23	110618-DBRI-F-U-S	11/06/2018	SM2540 D
11/13/2018	Water	J5851-07-C	TSS	Cool 4 deg C	HDR103	J23	110618-JETT-F-D-B	11/06/2018	SM2540 D
11/13/2018	Water	J5851-08-C	TSS	Cool 4 deg C	HDR103	J23	110618-JETT-F-D-M	11/06/2018	SM2540 D
11/13/2018	Water	J5851-09-C	TSS	Cool 4 deg C	HDR103	J23	110618-JETT-F-D-S	11/06/2018	SM2540 D
11/13/2018	Water	J5851-10-C	TSS	Cool 4 deg C	HDR103	J23	110618-SIDI-F-D-B	11/06/2018	SM2540 D
11/13/2018	Water	J5851-11-C	TSS	Cool 4 deg C	HDR103	J23	110618-SIDI-F-D-M	11/06/2018	SM2540 D
11/13/2018	Water	J5851-12-C	TSS	Cool 4 deg C	HDR103	J23	110618-SIDI-F-D-S	11/06/2018	SM2540 D
11/13/2018	Water	J5851-13-C	TSS	Cool 4 deg C	HDR103	J23	110618-SIDI-F-U-B	11/06/2018	SM2540 D
11/13/2018	Water	J5851-14-C	TSS	Cool 4 deg C	HDR103	J23	110618-SIDI-F-U-M	11/06/2018	SM2540 D
11/13/2018	Water	J5851-15-C	TSS	Cool 4 deg C	HDR103	J23	110618-SIDI-F-U-S	11/06/2018	SM2540 D
11/09/2018	Water	J5851-16-C	TSS	Cool 4 deg C	HDR103	J23	110618-TIPC-E-D-M	11/06/2018	SM2540 D
11/09/2018	Water	J5851-17-C	TSS	Cool 4 deg C	HDR103	J23	110618-TIPC-E-U-M	11/06/2018	SM2540 D
11/14/2018	Water	J5855-01	TSS	Cool 4 deg C	ARAM01	F31	COMP	11/07/2018	SM2540 D

Date/Time 11-08-18 7:20 PM

Received by: [Signature]

Relinquished by: [Signature]

Date/Time 11-08-18 11:20 AM

Received by: [Signature]

Relinquished by: [Signature]

1399091

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tss J5855

WorkList ID : 118980

Date : 11/8/2018 7:34:48 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/12/2018	Water	J5858-13 -G	TSS	Cool 4 deg C	PSEG01	I31	SB-9-GW	11/07/2018	SM2540 D

Date/Time 11-08-18 7:20 AM
Received by: JG
Relinquished by: JG

Date/Time 11-08-18 11:30 AM
Received by: JG
Relinquished by: JG