



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99097

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.9	11/08/2018	09:45
2	J5844-02		1	7	>100.0	>212.0	>212.0	11/08/2018	10:20
3	J5844-02DUP		1	7	>100.0	>212.0	>212.0	11/08/2018	10:50
4	J5844-05		1	9	>100.0	>212.0	>212.0	11/08/2018	11:20
5	J5858-14		1	9	>100.0	>212.0	>212.0	11/08/2018	11:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 99097

WorkList Name : FP11818

WorkList ID : 118997

Date : 11/8/2018 9:10:45 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/12/2018	Water	J5844-02	Flash Point	Cool 4 deg C	PSEG01	I31	MH-A-B	11/07/2018	1010A
11/12/2018	Water	J5844-05	Flash Point	Cool 4 deg C	PSEG01	I31	MH-C	11/07/2018	1010A
11/12/2018	Water	J5858-14	Flash Point	Cool 4 deg C	PSEG01	I31	SB-9-GW	11/07/2018	1010A

Date/Time 11/8/18 9:15 AM
 Received by: KS
 Relinquished by: CP

Date/Time 11/8/18 11:50 AM
 Received by: CP
 Relinquished by: KS