

=====LB99103=====

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

11/7/2018 13:17

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.835	0.0	0.104	
ICB1	0.122	0.0	0.004	
CCV1	241.087	0.0	0.250	
CCB1	0.034	0.0	0.004	
PB114563BL	0.172	0.0	0.004	
PB114563BS	100.129	0.0	0.106	
J5807-06	1861.207	0.0	1.903	Test limit high
J5807-06DLX10	181.037	0.0	0.188	
CCV2	243.058	0.0	0.252	
CCB2	0.463	0.0	0.004	

N	10
Mean	272.614
SD	566.7326
CV%	207.89

Aquakem v. 7.2AQ1

Results from time period:

Wed Nov 07 13:16:08 2018

Wed Nov 07 13:16:17 2018

A

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.3449	µg/l	11/7/2018 10:19:57	
5.0PPBCN	A	Total CN	P	4.7332	µg/l	11/7/2018 10:19:58	
10PPBCN	A	Total CN	P	9.8978	µg/l	11/7/2018 10:19:59	
50PPBCN	A	Total CN	P	48.4168	µg/l	11/7/2018 10:20:00	
100PPBCN	A	Total CN	P	102.1235	µg/l	11/7/2018 10:20:01	
250PPBCN	A	Total CN	P	250.8708	µg/l	11/7/2018 10:20:02	
500PPBCN	A	Total CN	P	499.303	µg/l	11/7/2018 10:20:03	
ICV1	S	Total CN	P	98.8345	µg/l	11/7/2018 13:16:08	
ICB1	S	Total CN	P	0.1215	µg/l	11/7/2018 13:16:09	
CCV1	S	Total CN	P	241.0865	µg/l	11/7/2018 13:16:10	
CCB1	S	Total CN	P	0.0342	µg/l	11/7/2018 13:16:11	
PB114563BL	S	Total CN	P	0.1725	µg/l	11/7/2018 13:16:12	
PB114563BS	S	Total CN	P	100.1294	µg/l	11/7/2018 13:16:13	
J5807-06	S	Total CN	P	1861.207	µg/l	11/7/2018 13:16:14	
J5807-06DLX10	S	Total CN	P	181.0365	µg/l	11/7/2018 13:16:15	
CCV2	S	Total CN	P	243.0584	µg/l	11/7/2018 13:16:16	
CCB2	S	Total CN	P	0.4634	µg/l	11/7/2018 13:16:17	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : *AK*

11/7/2018 10:20


Test Total CN

Accepted 11/7/2018 10:20

Factor 980
 Bias 0.004

Coeff. of det. 0.999959

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.3449	0.0000	
2	5.0PPBCN	0.008	4.7332	5.0000	
3	10PPBCN	0.014	9.8978	10.0000	
4	50PPBCN	0.053	48.4168	50.0000	
5	100PPBCN	0.108	102.1235	100.0000	
6	250PPBCN	0.260	250.8708	250.0000	
7	500PPBCN	0.513	499.3030	500.0000	