


7471B

LB 99108
INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	-103	0.000	-103				
0.2	0.200	0.186	-0.014	1015	0.0 %	1015				
2.5	2.500	2.488	-0.012	13822	0.0 %	13822				
5.0	5.000	5.054	0.054	28098	0.0 %	28098				
7.5	7.500	7.547	0.047	41972	0.0 %	41972				
10.0	10.000	9.940	-0.060	55285	0.0 %	55285				

LB99108 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Nov 2018 15:49:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-103	-	0.0000	7471B	PPB	08 Nov 2018 15:51:39	S	Std
0.2 - 1	50.2	1015	-	0.2000	7471B	PPB	08 Nov 2018 15:53:57	S	Std
2.5 - 1	52.5	13822	-	2.5000	7471B	PPB	08 Nov 2018 15:56:14	S	Std
5.0 - 1	55	28098	-	5.0000	7471B	PPB	08 Nov 2018 15:58:32	S	Std
7.5 - 1	57.5	41972	-	7.5000	7471B	PPB	08 Nov 2018 16:00:49	S	Std
10.0 - 1	510	55285	-	10.0000	7471B	PPB	08 Nov 2018 16:03:07	S	Std
ICV65 - 1	ICV65	20946	3.7683	-	7471B	PPB	08 Nov 2018 16:05:52	U	SMPL
ICB65 - 1	ICB65	-220	-0.0360	-	7471B	PPB	08 Nov 2018 16:08:08	U	SMPL
CCV37 - 1	CCV37	26793	4.8191	-	7471B	PPB	08 Nov 2018 16:15:15	U	SMPL
CCB37 - 1	CCB37	-221	-0.0361	-	7471B	PPB	08 Nov 2018 16:17:31	U	SMPL
CRA - 1	CRA	982	0.1801	-	7471B	PPB	08 Nov 2018 16:19:48	U	SMPL
HighStd - 1	HighStd	52197	9.3851	-	7471B	PPB	08 Nov 2018 16:22:04	U	SMPL
ChkStd - 1	ChkStd	37724	6.7838	-	7471B	PPB	08 Nov 2018 16:24:19	U	SMPL
PB114621BL - 1	PBS	-264	-0.0439	-	7471B	PPB	08 Nov 2018 16:26:38	U	SMPL
PB114621BS - 1	LCSS	20383	3.6671	-	7471B	PPB	08 Nov 2018 16:32:34	U	SMPL
J5768-01 - 1	OK-02-110718-A	2538	0.4597	-	7471B	PPB	08 Nov 2018 16:34:51	U	SMPL
J5768-03 - 1	OK-02-110718-B	939	0.1724	-	7471B	PPB	08 Nov 2018 16:37:09	U	SMPL
J5838-01 - 1	HC-SOIL-1	1272	0.2322	-	7471B	PPB	08 Nov 2018 16:39:27	U	SMPL
J5838-11 - 1	HC-SOIL-2	1049	0.1921	-	7471B	PPB	08 Nov 2018 16:41:45	U	SMPL
J5858-03 - 1	SB-8-S	1660	0.3019	-	7471B	PPB	08 Nov 2018 16:44:01	U	SMPL
CCV38 - 1	CCV38	26614	4.7870	-	7471B	PPB	08 Nov 2018 16:46:17	U	SMPL
CCB38 - 1	CCB38	-233	-0.0383	-	7471B	PPB	08 Nov 2018 16:48:33	U	SMPL
J5858-07 - 1	SB-8-D	362	0.0687	-	7471B	PPB	08 Nov 2018 16:50:50	U	SMPL
J5858-11 - 1	SB-9-S	3256	0.5888	-	7471B	PPB	08 Nov 2018 16:53:06	U	SMPL
J5859-02 - 1	B-23(0-10)	11587	2.0861	-	7471B	PPB	08 Nov 2018 16:55:22	U	SMPL
J5859-05 - 1	B-24(0-10)	13217	2.3791	-	7471B	PPB	08 Nov 2018 16:57:39	U	SMPL
J5859-05DUP - 1	B-24(0-10)DUP	11668	2.1007	-	7471B	PPB	08 Nov 2018 16:59:55	U	SMPL
J5859-05MS - 1	B-24(0-10)MS	33558	6.0350	-	7471B	PPB	08 Nov 2018 17:04:55	U	SMPL
J5859-05MSD - 1	B-24(0-10)MSD	33168	5.9649	-	7471B	PPB	08 Nov 2018 17:07:12	U	SMPL
J5859-05LX5 - 1		2209	0.4006	-	7471B	PPB	08 Nov 2018 17:09:32	U	SMPL
J5859-05A - 1		34843	6.2660	-	7471B	PPB	08 Nov 2018 17:11:51	U	SMPL
CCV39 - 1	CCV39	26748	4.8111	-	7471B	PPB	08 Nov 2018 17:14:07	U	SMPL
CCB39 - 1	CCB39	-243	-0.0401	-	7471B	PPB	08 Nov 2018 17:16:25	U	SMPL