



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99113

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.9	4.01	11/08/2018	18:00
2	CAL2	1	Water			20.8	7.00	11/08/2018	18:02
3	CAL3	1	Water			20.9	10.04	11/08/2018	18:04
4	ICV	1	Water			20.7	7.01	11/08/2018	18:06
5	CCV1	1	Water			20.8	2.00	11/08/2018	18:08
6	J5863-03	1	Solid	20.02	20	20.7	6.28	11/08/2018	18:10
7	J5863-03DUP	1	Solid	20.04	20	20.8	6.29	11/08/2018	18:12
8	J5876-02	1	Solid	20.02	20	21.3	8.32	11/08/2018	18:14
9	J5877-01	1	Solid	20.02	20	20.3	7.67	11/08/2018	18:16
10	J5884-04	1	Solid	20.03	20	21.3	7.04	11/08/2018	18:18
11	J5884-08	1	Solid	20.02	20	20.9	6.94	11/08/2018	18:20
12	J5884-12	1	Solid	20.02	20	21.1	5.86	11/08/2018	18:22
13	J5884-16	1	Solid	20.04	20	20.8	6.45	11/08/2018	18:24
14	CCV2	1	Water			20.9	12.00	11/08/2018	18:26

WORKLIST(Hardcopy Internal Chain)

LB99113

WorkList Name : corrosivity110818

WorkList ID : 119024

Date : 11/8/2018 5:13:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/08/2018	Solid	J5863-03	Corrosivity	Cool 4 deg C	PSEG01	I31	RT-3263	11/08/2018	9045D
11/08/2018	Solid	J5876-02	Corrosivity	Cool 4 deg C	PSEG01	I33	48382-47191-47202	11/08/2018	9045D
11/06/2018	Solid	J5877-01	Corrosivity	Cool 4 deg C	USAR03	F32	SSI-1106-S-TCLP-1	11/06/2018	9045D
11/08/2018	Solid	J5884-04	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-3-S	11/08/2018	9045D
11/08/2018	Solid	J5884-08	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-7-S	11/08/2018	9045D
11/08/2018	Solid	J5884-12	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-8A-S	11/08/2018	9045D
11/08/2018	Solid	J5884-16	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-8A-D	11/08/2018	9045D

Date/Time 11/08/18 5:50

Received by: SP

Relinquished by: CP

Date/Time 11/08/18 6:45

Received by: CP

Relinquished by: SP