



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB99115

Slope : 99.1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.9	4.01	11/08/2018	18:00
2	CAL2	1	Water			20.8	7.00	11/08/2018	18:02
3	CAL3	1	Water			20.9	10.04	11/08/2018	18:04
4	ICV	1	Water			20.7	7.01	11/08/2018	18:06
5	CCV1	1	Water			20.7	2.00	11/08/2018	18:30
6	J5269-01	1	Water			18.9	6.30	11/08/2018	18:32
7	J5269-01DUP	1	Water			19.0	6.31	11/08/2018	18:34
8	J5861-02	1	Water			15.2	5.12	11/08/2018	18:36
9	J5879-01	1	Water			17.1	7.14	11/08/2018	18:38
10	J5884-17	1	Water			13.2	6.58	11/08/2018	18:40
11	CCV2	1	Water			20.8	12.00	11/08/2018	18:42

WORKLIST(Hardcopy Internal Chain)

LB 99115

WorkList Name : pH110818

WorkList ID : 119026

Date : 11/8/2018 5:13:21 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/08/2018	Water	J5269-01	pH	Cool 4 deg C	PSEG01	I31	SB-12-GW	11/08/2018	9040C
11/08/2018	Water	J5861-02	pH	Cool 4 deg C	PSEG01	I22	CS-OILY-WATER	11/08/2018	9040C
11/08/2018	Water	J5879-01	pH	Cool 4 deg C	PSEG01	I31	4626-2221	11/08/2018	9040C
11/08/2018	Water	J5884-17	pH	Cool 4 deg C	PSEG01	I31	TP-7	11/08/2018	9040C

Date/Time 11/08/18 5:50
 Received by: SP
 Relinquished by: CP

Date/Time 11/08/18 6:45
 Received by: SP
 Relinquished by: SP