

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 11/09/2018

TEMP1 IN: 103 11/09/2018 07:30 TEMP OUT: 104 11/09/2018 08:45

TEMP2 IN: 104 11/09/2018 10:00 TEMP OUT: 104 11/09/2018 11:05

TEMP3 IN: 105 11/09/2018 12:15 TEMP OUT: 105 11/09/2018 13:45

TEMP4 IN: 105 11/09/2018 14:25 TEMP OUT: 105 11/09/2018 16:00

Run Number: LB99120

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB99120BL	LB99120BL	1.3992	1.3992	100	1.3992	1.3992	1.3992	0.0000	0
2	LB99120BS	LB99120BS	1.3938	1.3938	100	1.4440	1.4440	1.4440	0.0502	502
3	J5881-01	110718-TIPC-F-D-B	1.4008	1.4008	200	1.4191	1.4191	1.4191	0.0183	91.5
4	J5881-01DUP	110718-TIPC-F-D-BDUP	1.4052	1.4052	200	1.4231	1.4231	1.4231	0.0179	89.5
5	J5881-02	110718-TIPC-F-D-S	1.4001	1.4001	200	1.4115	1.4115	1.4115	0.0114	57
6	J5881-03	110718-TIPC-F-U-B	1.4039	1.4039	200	1.4179	1.4179	1.4179	0.0140	70
7	J5881-04	110718-TIPC-F-U-S	1.4007	1.4007	200	1.4112	1.4112	1.4112	0.0105	52.5
8	J5881-05	110718-JETT-E-D-B	1.3991	1.3991	300	1.4138	1.4138	1.4138	0.0147	49
9	J5881-06	110718-JETT-E-D-M	1.4036	1.4036	300	1.4166	1.4166	1.4166	0.0130	43.3
10	J5881-07	110718-JETT-E-D-S	1.3953	1.3953	400	1.4069	1.4069	1.4069	0.0116	29
11	J5881-08	110718-JETT-E-U-B	1.4005	1.4005	300	1.4134	1.4134	1.4134	0.0129	43
12	J5881-09	110718-JETT-E-U-M	1.4019	1.4019	300	1.4087	1.4087	1.4087	0.0068	22.7
13	J5881-10	110718-JETT-E-U-S	1.3931	1.3931	400	1.4019	1.4019	1.4019	0.0088	22
14	J5881-11	110718-DBRI-E-D-B	1.4063	1.4063	300	1.4221	1.4221	1.4221	0.0158	52.7
15	J5881-12	110718-DBRI-E-D-M	1.4024	1.4024	300	1.4134	1.4134	1.4134	0.0110	36.7

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 11/09/2018

TEMP1 IN: 103 11/09/2018 07:30 TEMP OUT: 104 11/09/2018 08:45

TEMP2 IN: 104 11/09/2018 10:00 TEMP OUT: 104 11/09/2018 11:05

TEMP3 IN: 105 11/09/2018 12:15 TEMP OUT: 105 11/09/2018 13:45

TEMP4 IN: 105 11/09/2018 14:25 TEMP OUT: 105 11/09/2018 16:00

Run Number: LB99120

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J5881-13	110718-DBRI-E-D-S	1.3947	1.3947	400	1.4094	1.4094	1.4094	0.0147	36.8
17	J5881-14	110718-DBRI-E-U-B	1.3961	1.3961	300	1.4154	1.4154	1.4154	0.0193	64.3
18	J5881-15	110718-DBRI-E-U-M	1.3879	1.3879	300	1.4003	1.4003	1.4003	0.0124	41.3
19	J5881-16	110718-DBRI-E-U-S	1.3974	1.3974	300	1.4057	1.4057	1.4057	0.0083	27.7
20	J5881-17	110718-SIDI-E-D-B	1.4061	1.4061	300	1.4269	1.4269	1.4269	0.0208	69.3
21	J5881-18	110718-SIDI-E-D-S	1.3928	1.3928	300	1.4103	1.4103	1.4103	0.0175	58.3
22	J5881-19	110718-SIDI-E-U-B	1.3965	1.3965	200	1.4152	1.4152	1.4152	0.0187	93.5
23	J5881-20	110718-SIDI-E-U-S	1.4065	1.4065	200	1.4226	1.4226	1.4226	0.0161	80.5
24	J5881-20DUP	110718-SIDI-E-U-SDUP	1.4061	1.4061	200	1.4220	1.4220	1.4220	0.0159	79.5

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS110918-02

WorkList ID : 119061

Date : 11/9/2018 10:28:16 AM

11/9/2018

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/12/2018	Water	J5881-01	C TSS	Cool 4 deg C	HDR103	J31	110718-TIPC-F-D-B	11/07/2018	SM2540 D
11/12/2018	Water	J5881-02	C TSS	Cool 4 deg C	HDR103	J31	110718-TIPC-F-D-S	11/07/2018	SM2540 D
11/12/2018	Water	J5881-03	C TSS	Cool 4 deg C	HDR103	J31	110718-TIPC-F-U-B	11/07/2018	SM2540 D
11/12/2018	Water	J5881-04	C TSS	Cool 4 deg C	HDR103	J31	110718-TIPC-F-U-S	11/07/2018	SM2540 D
11/14/2018	Water	J5881-05	C TSS	Cool 4 deg C	HDR103	J31	110718-JETT-E-D-B	11/07/2018	SM2540 D
11/14/2018	Water	J5881-06	C TSS	Cool 4 deg C	HDR103	J31	110718-JETT-E-D-M	11/07/2018	SM2540 D
11/14/2018	Water	J5881-07	C TSS	Cool 4 deg C	HDR103	J31	110718-JETT-E-D-S	11/07/2018	SM2540 D
11/14/2018	Water	J5881-08	C TSS	Cool 4 deg C	HDR103	J31	110718-JETT-E-U-B	11/07/2018	SM2540 D
11/14/2018	Water	J5881-09	C TSS	Cool 4 deg C	HDR103	J31	110718-JETT-E-U-M	11/07/2018	SM2540 D
11/14/2018	Water	J5881-10	C TSS	Cool 4 deg C	HDR103	J31	110718-JETT-E-U-S	11/07/2018	SM2540 D
11/14/2018	Water	J5881-11	C TSS	Cool 4 deg C	HDR103	J31	110718-DBRI-E-D-B	11/07/2018	SM2540 D
11/14/2018	Water	J5881-12	C TSS	Cool 4 deg C	HDR103	J31	110718-DBRI-E-D-M	11/07/2018	SM2540 D
11/14/2018	Water	J5881-13	C TSS	Cool 4 deg C	HDR103	J31	110718-DBRI-E-D-S	11/07/2018	SM2540 D
11/14/2018	Water	J5881-14	C TSS	Cool 4 deg C	HDR103	J31	110718-DBRI-E-U-B	11/07/2018	SM2540 D
11/14/2018	Water	J5881-15	C TSS	Cool 4 deg C	HDR103	J31	110718-DBRI-E-U-M	11/07/2018	SM2540 D
11/14/2018	Water	J5881-16	C TSS	Cool 4 deg C	HDR103	J31	110718-DBRI-E-U-S	11/07/2018	SM2540 D
11/14/2018	Water	J5881-17	C TSS	Cool 4 deg C	HDR103	J31	110718-SIDI-E-D-B	11/07/2018	SM2540 D
11/14/2018	Water	J5881-18	C TSS	Cool 4 deg C	HDR103	J31	110718-SIDI-E-D-S	11/07/2018	SM2540 D
11/14/2018	Water	J5881-19	C TSS	Cool 4 deg C	HDR103	J31	110718-SIDI-E-U-B	11/07/2018	SM2540 D
11/14/2018	Water	J5881-20	C TSS	Cool 4 deg C	HDR103	J31	110718-SIDI-E-U-S	11/07/2018	SM2540 D

FP-7042602

Date/Time 11/09/18, 10:35
 Received by: AS
 Relinquished by: CJP

Date/Time 11/09/18, 12:30
 Received by: CJP
 Relinquished by: AS