



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99122

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			20.2	4.01	11/09/2018	12:00
2	CAL2	1	Water			20.3	7.00	11/09/2018	12:02
3	CAL3	1	Water			20.4	10.04	11/09/2018	12:04
4	ICV	1	Water			20.3	7.01	11/09/2018	12:06
5	CCV1	1	Water			20.4	2.00	11/09/2018	12:08
6	J5269-10	1	Solid	20.02	20	19.3	6.31	11/09/2018	12:10
7	J5269-10DUP	1	Solid	20.03	20	19.4	6.32	11/09/2018	12:12
8	J5269-14	1	Solid	20.02	20	19.3	5.40	11/09/2018	12:14
9	J5269-18	1	Solid	20.03	20	19.3	6.88	11/09/2018	12:16
10	J5269-22	1	Solid	20.03	20	19.4	6.22	11/09/2018	12:18
11	J5887-04	1	Solid	20.02	20	19.4	8.98	11/09/2018	12:20
12	J5887-08	1	Solid	20.04	20	19.1	8.47	11/09/2018	12:22
13	J5887-12	1	Solid	20.03	20	19.3	8.56	11/09/2018	12:24
14	J5887-16	1	Solid	20.02	20	19.2	8.95	11/09/2018	12:26
15	CCV2	1	Water			20.5	12.00	11/09/2018	12:28

WORKLIST(Hardcopy Internal Chain)

UB 99122

WorkList Name : CORROSIONITY110918

WorkList ID : 119070

Date : 11/9/2018 11:49:33 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/08/2018	Solid	J5269-10	Corrosivity		PSEG01	I31	SB-12-D	11/08/2018	9045D
11/08/2018	Solid	J5269-14	Corrosivity		PSEG01	I31	SB-13-S	11/08/2018	9045D
11/08/2018	Solid	J5269-18	Corrosivity		PSEG01	I31	SB-13-D	11/08/2018	9045D
11/08/2018	Solid	J5269-22	Corrosivity		PSEG01	I31	SB-14	11/08/2018	9045D
11/08/2018	Solid	J5887-04	Corrosivity	Cool 4 deg C	PSEG01	I31	W-1-S	11/08/2018	9045D
11/08/2018	Solid	J5887-08	Corrosivity	Cool 4 deg C	PSEG01	I31	W-1-D	11/08/2018	9045D
11/08/2018	Solid	J5887-12	Corrosivity	Cool 4 deg C	PSEG01	I31	W-2	11/08/2018	9045D
11/08/2018	Solid	J5887-16	Corrosivity	Cool 4 deg C	PSEG01	I31	W-3-S	11/08/2018	9045D

Date/Time 11/09/18 11:50

Received by: SP

Relinquished by: CP

Date/Time 11/09/18 12:35

Received by: CP

Relinquished by: SP