



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99125

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.9	11/09/2018	09:20
2	J5269-02		1	7	>100.0	>212.0	>212.0	11/09/2018	09:50
3	J5269-02DUP		1	7	>100.0	>212.0	>212.0	11/09/2018	10:15
4	J5861-03		1	9	>100.0	>212.0	>212.0	11/09/2018	11:40
5	J5879-02		1	6	>100.0	>212.0	>212.0	11/09/2018	12:05
6	J5884-18		1	7	>100.0	>212.0	>212.0	11/09/2018	12:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 99125

WorkList Name : fp 11918

WorkList ID : 119052

Date : 11/9/2018 8:46:29 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/13/2018	Water	J5269-02	Flash Point	Cool 4 deg C	PSEG01	I31	SB-12-GW	11/08/2018	1010A
11/13/2018	Water	J5861-03	Flash Point	Cool 4 deg C	PSEG01	I22	CS-OILY-WATER	11/08/2018	1010A
11/13/2018	Water	J5879-02	Flash Point	Cool 4 deg C	PSEG01	I31	4626-2221	11/08/2018	1010A
11/13/2018	Water	J5884-18	Flash Point	Cool 4 deg C	PSEG01	I31	TP-7	11/08/2018	1010A

Date/Time 11/9/18 8:50 AM
 Received by: KS
 Relinquished by: CP

Date/Time 11/9/18 12:45 PM
 Received by: CP
 Relinquished by: KS