

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB99130

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP69433
calibration std. phosphate 0.5 ppm	WP69432
calibration std. phosphate 0.3 ppm	WP69431
calibration std. 0 ppm	WP69428
phosphate ICV std.	WP69435
phosphate CCV std.	WP69434
calibration std. phosphate 0.1 ppm	WP69430
calibration std. phosphate 0.05 ppm	WP69429
Combined reagent	WP69438
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
Sodium hydroxide, 1N	WP66924

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	11/09/2018	11:42
2	CAL2	0.05	1	50	50	0.033	11/09/2018	11:42
3	CAL3	0.10	1	50	50	0.065	11/09/2018	11:43
4	CAL4	0.30	1	50	50	0.186	11/09/2018	11:43
5	CAL5	0.50	1	50	50	0.330	11/09/2018	11:44
6	CAL6	1.00	1	50	50	0.649	11/09/2018	11:45



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB99130

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.324	0.500	11/09/2018	11:45
2	ICB		1	50	50	0.000	0.001	11/09/2018	11:46
3	CCV1	0.50	1	50	50	0.329	0.507	11/09/2018	11:47
4	CCB1		1	50	50	0.000	0.001	11/09/2018	11:47
5	PB114697BL		1	50	50	0.000	0.001	11/09/2018	11:48
6	PB114697BS	0.50	1	50	50	0.321	0.495	11/09/2018	11:49
7	J5732-01		1	50	50	0.018	0.029	11/09/2018	11:49
8	J5732-01DUP		1	50	50	0.016	0.026	11/09/2018	11:50
9	J5732-01MS	0.50	1	50	50	0.320	0.493	11/09/2018	11:51
10	J5732-01MSD	0.50	1	50	50	0.310	0.478	11/09/2018	11:52
11	J5732-02		1	50	50	0.021	0.034	11/09/2018	11:52
12	CCV2	0.50	1	50	50	0.328	0.506	11/09/2018	11:53
13	CCB2		1	50	50	0.000	0.001	11/09/2018	11:53