

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB99132

ANALYST: sweta
 SUPERVISOR REVIEW BY: ketankumar
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP68780
IODINE SOLUTION .025N 1L	W2410
Starch Solution, 4L	W2277

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB114691BL		1	50	50	2.00	0.00	1.86	1.86	0.14	0.00	0.00	11/09/2018	14:25
2	J5269-02		1	50	50	2.00	1.86	3.50	1.64	0.36	0.22	1.76	11/09/2018	14:27
3	J5269-02DUP		1	50	50	2.00	3.50	5.16	1.66	0.34	0.20	1.60	11/09/2018	14:29
4	J5861-03		1	50	50	2.00	5.16	6.58	1.42	0.58	0.44	3.52	11/09/2018	14:31
5	J5879-02		1	50	50	2.00	6.58	8.10	1.52	0.48	0.34	2.72	11/09/2018	14:33
6	J5884-18		1	50	50	2.00	8.10	9.74	1.64	0.36	0.22	1.76	11/09/2018	14:35

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume