


7470A

LB99133
INSTRUMENT ID: CV1

 μ Abs.:

35934



Linear

A= 0.0000e+000

B= 2.7823e-004

C= 7.3571e-003

Rho= 0.9999620

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.008	-0.008	-55	0.000	-55				
0.2	0.200	0.180	-0.020	619	0.0 %	619				
2.5	2.500	2.565	0.065	9192	0.0 %	9192				
5.0	5.000	4.962	-0.038	17806	0.0 %	17806				
7.5	7.500	7.497	-0.003	26918	0.0 %	26918				
10.0	10.000	10.005	0.005	35934	0.0 %	35934				

LB99133 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Nov 2018 15:25:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-55	-	0.00007470A	PPB		09 Nov 2018 15:26:48	S	Std
0.2 - 1	50.2	619	-	0.20007470A	PPB		09 Nov 2018 15:28:57	S	Std
2.5 - 1	52.5	9192	-	2.50007470A	PPB		09 Nov 2018 15:31:05	S	Std
5.0 - 1	55	17806	-	5.00007470A	PPB		09 Nov 2018 15:33:14	S	Std
7.5 - 1	57.5	26918	-	7.50007470A	PPB		09 Nov 2018 15:35:22	S	Std
10.0 - 1	510	35934	-	10.00007470A	PPB		09 Nov 2018 15:37:29	S	Std
ICV67 - 1	ICV67	14111	3.9335	-7470A	PPB		09 Nov 2018 15:40:04	U	SMPL
ICB67 - 1	ICB67	-137	-0.0308	-7470A	PPB		09 Nov 2018 15:42:10	U	SMPL
CCV49 - 1	CCV49	18586	5.1785	-7470A	PPB		09 Nov 2018 15:44:17	U	SMPL
CCB49 - 1	CCB49	-134	-0.0299	-7470A	PPB		09 Nov 2018 15:46:24	U	SMPL
CRA - 1	CRA	615	0.1785	-7470A	PPB		09 Nov 2018 15:48:31	U	SMPL
HighStd - 1	HighStd	36345	10.1196	-7470A	PPB		09 Nov 2018 15:50:38	U	SMPL
ChkStd - 1	ChkStd	25363	7.0641	-7470A	PPB		09 Nov 2018 15:52:45	U	SMPL
PB114644BL - 1	PBW	-221	-0.0541	-7470A	PPB		09 Nov 2018 15:54:53	U	SMPL
PB114644BS - 1	LCSW	14856	4.1407	-7470A	PPB		09 Nov 2018 15:57:00	U	SMPL
J5768-02 - 1	OK-02-110718-A	-164	-0.0383	-7470A	PPB		09 Nov 2018 15:59:08	U	SMPL
J5768-02DUP - 1	OK-02-110718-ADUP	-70	-0.0121	-7470A	PPB		09 Nov 2018 16:01:16	U	SMPL
J5768-02MS - 1	OK-02-110718-AMS	11862	3.3077	-7470A	PPB		09 Nov 2018 16:03:24	U	SMPL
J5768-02MSD - 1	OK-02-110718-AMSD	12333	3.4388	-7470A	PPB		09 Nov 2018 16:05:33	U	SMPL
J5768-04 - 1	OK-02-110718-B	-167	-0.0391	-7470A	PPB		09 Nov 2018 16:07:40	U	SMPL
CCV50 - 1	CCV50	18750	5.2242	-7470A	PPB		09 Nov 2018 16:09:48	U	SMPL
CCB50 - 1	CCB50	-146	-0.0333	-7470A	PPB		09 Nov 2018 16:11:54	U	SMPL
J5838-10 - 1	HC-SOIL-1	-85	-0.0163	-7470A	PPB		09 Nov 2018 16:14:01	U	SMPL
J5838-20 - 1	HC-SOIL-2	-96	-0.0194	-7470A	PPB		09 Nov 2018 16:16:07	U	SMPL
J5858-04 - 1	SB-8-S	-86	-0.0166	-7470A	PPB		09 Nov 2018 16:18:14	U	SMPL
J5858-08 - 1	SB-8-D	-83	-0.0157	-7470A	PPB		09 Nov 2018 16:20:21	U	SMPL
J5858-12 - 1	SB-9-S	-95	-0.0191	-7470A	PPB		09 Nov 2018 16:22:28	U	SMPL
J5859-03 - 1	B-23(0-10)	-72	-0.0127	-7470A	PPB		09 Nov 2018 16:24:36	U	SMPL
J5859-06 - 1	B-24(0-10)	-88	-0.0171	-7470A	PPB		09 Nov 2018 16:26:44	U	SMPL
PB114645BL - 1	PBW	-83	-0.0157	-7470A	PPB		09 Nov 2018 16:28:52	U	SMPL
PB114645BS - 1	LCSW	12779	3.5629	-7470A	PPB		09 Nov 2018 16:31:01	U	SMPL
J5844-02 - 1	MH-A-B	-128	-0.0283	-7470A	PPB		09 Nov 2018 16:33:08	U	SMPL
CCV51 - 1	CCV51	18891	5.2634	-7470A	PPB		09 Nov 2018 16:35:14	U	SMPL
CCB51 - 1	CCB51	-129	-0.0285	-7470A	PPB		09 Nov 2018 16:37:21	U	SMPL
J5844-02DUP - 1	MH-A-BDUP	-100	-0.0205	-7470A	PPB		09 Nov 2018 16:39:28	U	SMPL
J5844-02MS - 1	MH-A-BMS	12310	3.4324	-7470A	PPB		09 Nov 2018 16:41:35	U	SMPL
J5844-02MSD - 1	MH-A-BMSD	12555	3.5005	-7470A	PPB		09 Nov 2018 16:43:42	U	SMPL
J5844-05 - 1	MH-C	-154	-0.0355	-7470A	PPB		09 Nov 2018 16:45:50	U	SMPL
J5858-14 - 1	SB-9-GW	-98	-0.0199	-7470A	PPB		09 Nov 2018 16:47:57	U	SMPL
PB114646BL - 1	PBW	-92	-0.0182	-7470A	PPB		09 Nov 2018 16:50:05	U	SMPL
PB114646BS - 1	LCSW	12697	3.5400	-7470A	PPB		09 Nov 2018 16:52:14	U	SMPL
J5844-01 - 1	MH-A-B	-163	-0.0380	-7470A	PPB		09 Nov 2018 16:54:22	U	SMPL
J5844-04 - 1	MH-C	-157	-0.0363	-7470A	PPB		09 Nov 2018 16:56:30	U	SMPL
CCV52 - 1	CCV52	19007	5.2957	-7470A	PPB		09 Nov 2018 16:58:38	U	SMPL
CCB52 - 1	CCB52	-143	-0.0324	-7470A	PPB		09 Nov 2018 17:00:45	U	SMPL
5858-13 - 1	SB-9-GW	-92	-0.0182	-7470A	PPB		09 Nov 2018 17:02:53	U	SMPL
J5858-13DUP - 1	SB-9-GWDUP	-91	-0.0180	-7470A	PPB		09 Nov 2018 17:05:00	U	SMPL
J5858-13MS - 1	SB-9-GWMS	15217	4.2412	-7470A	PPB		09 Nov 2018 17:09:34	U	SMPL
J5858-13MSD - 1	SB-9-GWMSD	15518	4.3249	-7470A	PPB		09 Nov 2018 17:14:11	U	SMPL
PB114644TB - 1	PB114609TB	-132	-0.0294	-7470A	PPB		09 Nov 2018 17:16:17	U	SMPL
PB114645TB - 1	PB114611TB	-86	-0.0166	-7470A	PPB		09 Nov 2018 17:18:25	U	SMPL
CCV53 - 1	CCV53	18839	5.2489	-7470A	PPB		09 Nov 2018 17:20:32	U	SMPL
CCB53 - 1	CCB53	-118	-0.0255	-7470A	PPB		09 Nov 2018 17:22:41	U	SMPL
J5768-02LX5 - 1		-93	-0.0185	-7470A	PPB		09 Nov 2018 17:24:49	U	SMPL
J5768-02A - 1		13806	3.8486	-7470A	PPB		09 Nov 2018 17:26:58	U	SMPL
J5844-02LX5 - 1		-154	-0.0355	-7470A	PPB		09 Nov 2018 17:29:05	U	SMPL
J5844-02A - 1		13789	3.8439	-7470A	PPB		09 Nov 2018 17:31:13	U	SMPL
J5858-13LX5 - 1		-152	-0.0349	-7470A	PPB		09 Nov 2018 17:33:20	U	SMPL
J5858-13A - 1		14731	4.1060	-7470A	PPB		09 Nov 2018 17:35:27	U	SMPL
CCV54 - 1	CCV54	18920	5.2715	-7470A	PPB		09 Nov 2018 17:37:34	U	SMPL
CCB54 - 1	CCB54	-164	-0.0383	-7470A	PPB		09 Nov 2018 17:39:41	U	SMPL