

LB99135


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

54512



A= 0.0000e+000

B= 1.8331e-004

C= 8.9100e-003

Rho= 0.9999534

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.022	0.022	71	0.000	71				
0.2	0.200	0.196	-0.004	1019	0.0 %	1019				
2.5	2.500	2.512	0.012	13655	0.0 %	13655				
5.0	5.000	4.927	-0.073	26831	0.0 %	26831				
7.5	7.500	7.541	0.041	41090	0.0 %	41090				
10.0	10.000	10.002	0.002	54512	0.0 %	54512				

LB99135 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Nov 2018 17:51:29

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	71	-	0.00007471B	PPB	09 Nov 2018 17:53:02	S	Std
0.2 - 1	50.2	1019	-	0.20007471B	PPB	09 Nov 2018 17:55:20	S	Std
2.5 - 1	52.5	13655	-	2.50007471B	PPB	09 Nov 2018 17:57:38	S	Std
5.0 - 1	55	26831	-	5.00007471B	PPB	09 Nov 2018 17:59:55	S	Std
7.5 - 1	57.5	41090	-	7.50007471B	PPB	09 Nov 2018 18:02:12	S	Std
10.0 - 1	510	54512	-	10.00007471B	PPB	09 Nov 2018 18:04:31	S	Std
ICV68 - 1	ICV68	21245	3.9034	-7471B	PPB	09 Nov 2018 18:07:12	U	SMPL
ICB68 - 1	ICB68	11	0.0109	-7471B	PPB	09 Nov 2018 18:09:28	U	SMPL
CCV55 - 1	CCV55	27176	4.9906	-7471B	PPB	09 Nov 2018 18:11:45	U	SMPL
CCB55 - 1	CCB55	-7	0.0076	-7471B	PPB	09 Nov 2018 18:14:01	U	SMPL
CRA - 1	CRA	1008	0.1937	-7471B	PPB	09 Nov 2018 18:16:18	U	SMPL
HighStd - 1	HighStd	54825	10.0591	-7471B	PPB	09 Nov 2018 18:18:34	U	SMPL
ChkStd - 1	ChkStd	37668	6.9140	-7471B	PPB	09 Nov 2018 18:20:49	U	SMPL
PB114660BL - 1	PBS	-137	-0.0162	-7471B	PPB	09 Nov 2018 18:23:08	U	SMPL
PB114660BS - 1	LCSS	18843	3.4631	-7471B	PPB	09 Nov 2018 18:25:27	U	SMPL
J5860-01 - 1	EP2-SB-101(13-15)	128	0.0324	-7471B	PPB	09 Nov 2018 18:27:44	U	SMPL
J5860-02 - 1	EP2-SB-102(13-15)	187	0.0432	-7471B	PPB	09 Nov 2018 18:30:01	U	SMPL
J5860-03 - 1	EP2-SB-103(14-16)	145	0.0355	-7471B	PPB	09 Nov 2018 18:32:19	U	SMPL
J5860-04 - 1	EP2-SB-104(15-17)	226	0.0503	-7471B	PPB	09 Nov 2018 18:34:37	U	SMPL
J5860-04DUP - 1	EP2-SB-104(15-17)DUP	221	0.0494	-7471B	PPB	09 Nov 2018 18:36:53	U	SMPL
CCV56 - 1	CCV56	27241	5.0025	-7471B	PPB	09 Nov 2018 18:39:09	U	SMPL
CCB56 - 1	CCB56	-29	0.0036	-7471B	PPB	09 Nov 2018 18:41:24	U	SMPL
J5860-05 - 1	J5860-04MS	19574	3.5971	-7471B	PPB	09 Nov 2018 18:43:42	U	SMPL
J5860-06 - 1	J5860-04MSD	19332	3.5527	-7471B	PPB	09 Nov 2018 18:45:57	U	SMPL
J5860-07 - 1	EP2-SB-107(14-16)	449	0.0912	-7471B	PPB	09 Nov 2018 18:48:14	U	SMPL
J5860-08 - 1	EP2-SB-112(14-16)	139	0.0344	-7471B	PPB	09 Nov 2018 18:50:31	U	SMPL
J5860-09 - 1	EP2-FD-01-1118	580	0.1152	-7471B	PPB	09 Nov 2018 18:52:48	U	SMPL
J5860-14 - 1	EP2-SB-105(0-4)	28633	5.2577	-7471B	PPB	09 Nov 2018 18:55:05	U	SMPL
J5860-15 - 1	EP2-SB-105(8-12)	701	0.1374	-7471B	PPB	09 Nov 2018 18:57:22	U	SMPL
J5860-16 - 1	EP2-SB-106(2-4)	43418	7.9680	-7471B	PPB	09 Nov 2018 18:59:41	U	SMPL
J5860-17 - 1	EP2-SB-108(14-16)	-69	-0.0037	-7471B	PPB	09 Nov 2018 19:01:59	U	SMPL
J5860-18 - 1	EP2-SB-109(14-16)	138	0.0342	-7471B	PPB	09 Nov 2018 19:04:17	U	SMPL
CCV57 - 1	CCV57	27062	4.9697	-7471B	PPB	09 Nov 2018 19:06:33	U	SMPL
CCB57 - 1	CCB57	-53	-0.0008	-7471B	PPB	09 Nov 2018 19:08:50	U	SMPL
J5860-19 - 1	EP2-SB-109(17-18)	175	0.0410	-7471B	PPB	09 Nov 2018 19:11:08	U	SMPL
J5860-20 - 1	EP2-SB-109(22-24)	137	0.0340	-7471B	PPB	09 Nov 2018 19:13:24	U	SMPL
J5860-21 - 1	EP2-SB-110(14-16)	173	0.0406	-7471B	PPB	09 Nov 2018 19:15:40	U	SMPL
J5860-22 - 1	EP2-SB-110(17-19)	210	0.0474	-7471B	PPB	09 Nov 2018 19:17:57	U	SMPL
J5860-23 - 1	EP2-SB-110(24-28)	303	0.0645	-7471B	PPB	09 Nov 2018 19:20:13	U	SMPL
J5860-24 - 1	EP2-SB-111(14-16)	213	0.0480	-7471B	PPB	09 Nov 2018 19:22:30	U	SMPL
J5860-25 - 1	EP2-SB-111(17-19)	185	0.0428	-7471B	PPB	09 Nov 2018 19:24:47	U	SMPL
J5860-26 - 1	EP2-SB-111(22-24)	169	0.0399	-7471B	PPB	09 Nov 2018 19:27:05	U	SMPL
J5860-27 - 1	EP2-SB-106(8.5-10.5)	268	0.0580	-7471B	PPB	09 Nov 2018 19:29:23	U	SMPL
PB114661BL - 1	PBS	26	0.0137	-7471B	PPB	09 Nov 2018 19:31:39	U	SMPL
CCV58 - 1	CCV58	27373	5.0267	-7471B	PPB	09 Nov 2018 19:33:56	U	SMPL
CCB58 - 1	CCB58	-49	-0.0001	-7471B	PPB	09 Nov 2018 19:36:12	U	SMPL
PB114661BS - 1	LCSS	19852	3.6480	-7471B	PPB	09 Nov 2018 19:38:30	U	SMPL
J5269-09 - 1	SB-12-D	99	0.0271	-7471B	PPB	09 Nov 2018 19:40:47	U	SMPL
J5269-13 - 1	SB-13-S	2109	0.3955	-7471B	PPB	09 Nov 2018 19:43:04	U	SMPL
J5269-17 - 1	SB-13-D	194	0.0445	-7471B	PPB	09 Nov 2018 19:45:21	U	SMPL
J5269-21 - 1	SB-14	999	0.1920	-7471B	PPB	09 Nov 2018 19:47:37	U	SMPL
J5862-01 - 1	OILY-DEBRIS	1656	0.3125	-7471B	PPB	09 Nov 2018 19:49:55	U	SMPL
J5863-01 - 1	RT-3263	815	0.1583	-7471B	PPB	09 Nov 2018 19:52:12	U	SMPL
J5863-01DUP - 1	RT-3263DUP	812	0.1578	-7471B	PPB	09 Nov 2018 19:54:29	U	SMPL
J5863-01MS - 1	RT-3263MS	20508	3.7683	-7471B	PPB	09 Nov 2018 19:56:47	U	SMPL
J5863-01MSD - 1	RT-3263MSD	20130	3.6990	-7471B	PPB	09 Nov 2018 19:59:03	U	SMPL
CCV59 - 1	CCV59	27198	4.9947	-7471B	PPB	09 Nov 2018 20:01:20	U	SMPL
CCB59 - 1	CCB59	-57	-0.0015	-7471B	PPB	09 Nov 2018 20:03:38	U	SMPL
J5864-01 - 1	OLDMANS-1	710	0.1391	-7471B	PPB	09 Nov 2018 20:05:56	U	SMPL
J5864-02 - 1	OLDMANS-2	962	0.1853	-7471B	PPB	09 Nov 2018 20:08:13	U	SMPL
J5876-01 - 1	48382-47191-47202	1131	0.2162	-7471B	PPB	09 Nov 2018 20:10:29	U	SMPL
J5877-03 - 1	SSSI-1107-S-DH-11	2541	0.4747	-7471B	PPB	09 Nov 2018 20:12:45	U	SMPL
J5884-03 - 1	TP-3-S	1692	0.3191	-7471B	PPB	09 Nov 2018 20:15:02	U	SMPL
J5884-07 - 1	TP-7-S	1719	0.3240	-7471B	PPB	09 Nov 2018 20:17:19	U	SMPL
J5884-11 - 1	TP-8A-S	2879	0.5367	-7471B	PPB	09 Nov 2018 20:19:36	U	SMPL
J5884-15 - 1	TP-8A-D	4031	0.7478	-7471B	PPB	09 Nov 2018 20:21:53	U	SMPL
J5900-01 - 1	WJ-18-11-09-TSHG-21	1924	0.3616	-7471B	PPB	09 Nov 2018 20:24:11	U	SMPL
J5900-02 - 1	WJ-18-11-09-TSHG-22	1660	0.3132	-7471B	PPB	09 Nov 2018 20:26:28	U	SMPL
CCV60 - 1	CCV60	27305	5.0143	-7471B	PPB	09 Nov 2018 20:28:45	U	SMPL
CCB60 - 1	CCB60	-72	-0.0043	-7471B	PPB	09 Nov 2018 20:31:01	U	SMPL
J5900-03 - 1	WJ-18-11-09-TSHG-23	1529	0.2892	-7471B	PPB	09 Nov 2018 20:33:19	U	SMPL
J5860-04LX5 - 1		77	0.0230	-7471B	PPB	09 Nov 2018 20:35:36	U	SMPL
J5860-04A - 1		20834	3.8281	-7471B	PPB	09 Nov 2018 20:37:52	U	SMPL
J5863-01LX5 - 1		106	0.0283	-7471B	PPB	09 Nov 2018 20:40:09	U	SMPL
J5863-01A - 1		22060	4.0528	-7471B	PPB	09 Nov 2018 20:42:27	U	SMPL
CCV61 - 1	CCV61	27123	4.9809	-7471B	PPB	09 Nov 2018 20:44:44	U	SMPL
CCB61 - 1	CCB61	-37	0.0021	-7471B	PPB	09 Nov 2018 20:47:02	U	SMPL