



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB99138

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP68911
COD calibration std. 100 ppm	WP68908
COD calibration std. 10 ppm	WP68907
COD calibration std. 150 ppm	WP68910
COD calibration std. 50 ppm	WP68909
COD calibration std. 0 ppm	WP68906
COD Digestion Vials Low Level 0-150Mg/L	W2421
COD ICV-LCS std, 50ppm	WP69459
COD CCV std, 50ppm	WP69460

Temp In (C): 150	Date In: 11/09/2018	Time In: 11:15
Temp Out (C): 150	Date Out: 11/09/2018	Time Out: 13:15

Intercept: -0.7935

Slope: 1.0096

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	10/20/2018	13:05
2	CAL2	10	1	Water	9.000	10/20/2018	13:05
3	CAL3	50	1	Water	49.000	10/20/2018	13:06
4	CAL4	100	1	Water	100.000	10/20/2018	13:06
5	CAL5	150	1	Water	151.000	10/20/2018	13:07



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB99138

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	49.000	49.320	10/20/2018	13:07
2	ICB				1	Water	0.000	0.786	10/20/2018	13:08
3	CCV1	50			1	Water	51.000	51.301	11/09/2018	14:00
4	CCB1				1	Water	0.000	0.786	11/09/2018	14:00
5	LB99138BL				1	Water	0.000	0.786	11/09/2018	14:01
6	LB99138BS	50			1	Water	50.000	50.311	11/09/2018	14:01
7	J5844-01				1	Water	40.000	40.406	11/09/2018	14:02
8	J5844-01DUP				1	Water	41.000	41.396	11/09/2018	14:02
9	J5844-01MS	50			1	Water	90.000	89.930	11/09/2018	14:03
10	J5844-01MSD	50			1	Water	91.000	90.921	11/09/2018	14:03
11	J5844-04				1	Water	33.000	33.472	11/09/2018	14:04
12	J5858-13				1	Water	6.000	6.729	11/09/2018	14:04
13	CCV2	50			1	Water	51.000	51.301	11/09/2018	14:05
14	CCB2				1	Water	0.000	0.786	11/09/2018	14:05