



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99153

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	11/09/2018	17:20
2	CAL2	1	Water			21.4	7.00	11/09/2018	17:22
3	CAL3	1	Water			21.2	10.04	11/09/2018	17:24
4	ICV	1	Water			21.0	7.01	11/09/2018	17:26
5	CCV1	1	Water			21.4	2.00	11/09/2018	17:28
6	J5766-02	1	Solid	20.02	20	19.5	7.70	11/09/2018	17:30
7	J5766-02DUP	1	Solid	20.04	20	19.6	7.71	11/09/2018	17:32
8	J5892-02	1	Solid	20.04	20	20.0	8.55	11/09/2018	17:34
9	J5901-08	1	Solid	20.02	20	15.6	6.60	11/09/2018	17:36
10	J5907-04	1	Solid	20.03	20	19.9	7.69	11/09/2018	17:38
11	J5907-08	1	Solid	20.02	20	19.7	8.36	11/09/2018	17:40
12	J5907-12	1	Solid	20.04	20	19.5	8.43	11/09/2018	17:42
13	J5907-16	1	Solid	20.03	20	19.9	8.51	11/09/2018	17:44
14	CCV2	1	Water			21.2	12.00	11/09/2018	17:46

WORKLIST(Hardcopy Internal Chain)

UB99153

WorkList Name : corrosivity110918-1

WorkList ID : 119083

Date : 11/9/2018 5:41:05 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/09/2018	Solid	J5766-02	Corrosivity	Cool 4 deg C	PSEG05	I32	HD-01-110918	11/09/2018	9045D
11/09/2018	Solid	J5892-02	Corrosivity	Cool 4 deg C	PSEG01	I31	PN-STONE	11/09/2018	9045D
11/08/2018	Solid	J5901-08	Corrosivity	Cool 4 deg C	BRIN05	K21	SS-2	11/08/2018	9045D
11/09/2018	Solid	J5907-04	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-2-S	11/09/2018	9045D
11/09/2018	Solid	J5907-08	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-3-S	11/09/2018	9045D
11/09/2018	Solid	J5907-12	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-3-D	11/09/2018	9045D
11/09/2018	Solid	J5907-16	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-4-S	11/09/2018	9045D

Date/Time 11/09/18 5:10
 Received by: SP
 Relinquished by: CP

Date/Time 11/09/18 6:00
 Received by: CP
 Relinquished by: CP