



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB99154

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2390
buffer solution pH 7 yellow	W2358
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.2	4.01	11/09/2018	17:20
2	CAL2	1	Water			21.4	7.00	11/09/2018	17:22
3	CAL3	1	Water			21.2	10.04	11/09/2018	17:24
4	ICV	1	Water			21.0	7.01	11/09/2018	17:26
5	CCV1	1	Water			21.1	2.00	11/09/2018	17:48
6	J5907-18	1	Water			17.0	6.80	11/09/2018	17:50
7	J5907-18DUP	1	Water			17.1	6.81	11/09/2018	17:52
8	CCV2	1	Water			21.3	12.00	11/09/2018	17:54

WORKLIST(Hardcopy Internal Chain)

UB99154

WorkList Name : corrosivity110918-2

WorkList ID : 119084

Date : 11/9/2018 5:41:14 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/09/2018	Water	J5907-18	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-3	11/09/2018	9040C

Date/Time 11/09/18 5:10
 Received by: SP
 Relinquished by: CP

Date/Time 11/09/18 8:00
 Received by: CP
 Relinquished by: SP