



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: sweta

Run Number: LB99159

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 55971

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.9	11/12/2018	09:40
2	J5907-18		1	9	>100.0	>212.0	>212.0	11/12/2018	10:05
3	J5907-18DUP		1	9	>100.0	>212.0	>212.0	11/12/2018	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LP 99159

WorkList Name : fp111218

WorkList ID : 119133

Date : 11/12/2018 9:06:45 AM

Due Date	Matrix	Sample	Test	Flash Point	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/14/2018	Water	J5907-18		Flash Point	Cool 4 deg C	PSEG01	I31	TP-3	11/09/2018	1010A

Date/Time 11/12/18 9:10 AM
 Received by: KS
 Relinquished by: CP

Date/Time 11/12/18 10:45 AM
 Received by: CP
 Relinquished by: KS